



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

THE MODAL SHOP, INC. AN AMPHENOL COMPANY  
10310 Aerohub Blvd  
Cincinnati, OH 45215  
Lisa Moore Phone: 513 351-9919

CALIBRATION

Valid To: April 30, 2026

Certificate Number: 2649.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,4</sup>:

I. Acoustical Quantities

Parameter/Equipment	Frequency	CMC <sup>2</sup> (±)	Comments
Condenser Microphone –	(10 to 100 000) Hz	0.25 dB	Microphone (Electrostatic Actuator)
With Pre-Amp	(10 to 100 000) Hz	0.25 dB	Microphone (Electrostatic Actuator)

II. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2, 3, 5</sup> (±)	Comments
DC Voltage – Measure	100 mV to 100 V	0.12 %	DMM, oscilloscope function generator, calibrated capacitor, DAQ
AC Voltage Ratio – Measure	100 mV to 10 V	0.12 %	DMM, oscilloscope function generator, calibrated capacitor, DAQ

### III. Mechanical

Parameter/Equipment	Range	CMC <sup>2,3</sup> (±)	Comments
Impulse Force Hammer	(1 to 5000) lbf	2.5 %	Data acquisition card accelerometer w/ calibrated masses
Force Sensor/Load Cell	(1 to 3.5) lbf	2.0 %	Data acquisition card w/ calibrated masses

Parameter/Equipment	Frequency	CMC <sup>2</sup> (±)	Comments
Vibration Laser Primary – Sensitivity Magnitude			
Low Frequency	$(0.5 \leq f \leq 10)$ Hz	0.30 %	Data acquisition card w/ air bearing low frequency shaker
Mid to High Frequency	5 Hz $(5 < f < 100)$ Hz $(100 \leq f \leq 160)$ Hz $(160 < f \leq 1000)$ Hz $(1000 < f \leq 5000)$ Hz $(5000 < f \leq 15\,000)$ Hz $(15\,000 < f \leq 20\,000)$ Hz	1.0 % 0.50 % 0.20 % 0.50 % 0.70 % 1.5 % 2.0 %	Data acquisition card w/ primary laser reference air bearing shaker  <i>f</i> represents the calibration frequency
Vibration Laser Primary – Sensitivity Phase			
Low Frequency	$(0.5 \leq f \leq 10)$ Hz	0.5°	Data acquisition card w/ air bearing low frequency shaker
Mid to High Frequency	$(5 \leq f < 5000)$ Hz $(5000 \leq f \leq 20\,000)$ Hz	0.7° 2.1°	Data acquisition card w/ primary laser reference air bearing shaker

Parameter/Equipment	Frequency	CMC <sup>2,3</sup> (±)	Comments
Vibration General Purpose – Sensitivity Magnitude			Data acquisition card
Mid to High Frequency	$(5 \leq f \leq 10)$ Hz $(10 < f < 100)$ Hz 100 Hz $(100 < f \leq 920)$ Hz $(920 < f \leq 5000)$ Hz $(5000 < f \leq 10\,000)$ Hz $(10\,000 < f \leq 15\,000)$ Hz $(15\,000 < f \leq 20\,000)$ Hz	1.7 % 1.2 % 0.75 % 1.0 % 1.4 % 1.9 % 2.2 % 2.8 %	w/ 396C10/C11 air bearing shaker
Low Frequency	$(0.5 \leq f \leq 1)$ Hz $(1 < f \leq 10)$ Hz	1.1 % 0.8 %	w/ low frequency long stroke shaker w/ optical reference
Mid to High Frequency – Using Secondary Standard	$(10 \leq f < 100)$ Hz 100 Hz $(100 < f \leq 920)$ Hz $(920 < f \leq 2000)$ Hz	3.0 % 1.3 % 1.5 % 3.0 %	w/ MB CAL 50
Vibration General Purpose –Sensitivity Phase			Data acquisition card
Mid to High Frequency	$(5 \leq f < 100)$ Hz $(100 \leq f < 5000)$ Hz $(5000 \leq f \leq 20\,000)$ Hz	1.5° 1.3° 2.8°	w/ 396C10/C11 air bearing shaker
Low Frequency	$(0.5 \leq f \leq 10)$ Hz	1.0°	w/ low frequency long stroke shaker w/ optical reference
Calibration of Portable Vibration Calibrator	$(5 \leq f \leq 10)$ Hz $(10 < f \leq 30)$ Hz $(30 < f < 100)$ Hz 100 Hz $(100 < f < 2000)$ Hz $(2000 \leq f \leq 10\,000)$ Hz	4.0 % 3.0 % 1.5 % 1.0 % 1.5 % 4.0 %	DMM, signal conditioner, standard accelerometer

Parameter/Equipment	Range	CMC <sup>2,3</sup> (±)	Comments
Shock Accelerometer Calibration	(20 to 10 000) g	2.2 %	Data acquisition card shock reference shock exciter

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> In the statement of CMC, percentages are to be read as percent of reading unless otherwise noted.

<sup>4</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.

<sup>5</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



## Accredited Laboratory

A2LA has accredited

**THE MODAL SHOP INC, an Amphenol Company**

*Cincinnati, OH*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of R205 – *Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 4<sup>th</sup> day of April 2024.

A blue ink signature of Mr. Trace McInturff, written over a horizontal line.

Mr. Trace McInturff Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 2649.01  
Valid to April 30, 2026

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*