



## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

CROSS TECHNOLOGIES, INC  
222 Cypress Creek Drive  
Calvert City, KY 42029  
Connie Foster Phone: 800 327 7727

### CALIBRATION

Valid To: May 31, 2026

Certificate Number: 1741.14

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1, 9</sup>:

#### I. Chemical

Parameter/Equipment	Range	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
pH Meters <sup>3</sup>	4 pH 7 pH 10 pH	0.04 pH 0.04 pH 0.05 pH	Standard pH solutions
Conductivity Meters <sup>3</sup>	0.77 $\mu$ S/cm 9.76 $\mu$ S/cm 99.4 $\mu$ S/cm 996 $\mu$ S/cm 10 000 $\mu$ S/cm	0.37 $\mu$ S/cm 0.38 $\mu$ S/cm 2.2 $\mu$ S/cm 5.4 $\mu$ S/cm 47 $\mu$ S/cm	Standard conductivity solutions
Refractometers <sup>3</sup>	(5, 15, 40, 70) % Brix	0.13 %	Sucrose solutions

#### II. Dimensional

Parameter/Equipment	Range	CMC <sup>2</sup> ( $\pm$ )	Comments
Pin Gage <sup>3</sup> – Class ZZ	Up to 1.0 in	85 $\mu$ in	Micrometer

Parameter/Equipment	Range	CMC <sup>2,5</sup> ( $\pm$ )	Comments
Calipers <sup>3</sup>	Up to 60 in	(5.2 + 9.8L) $\mu$ in + 0.6R	Gage blocks
Micrometers <sup>3</sup>	Up to 60 in	(4.6 + 9.8L) $\mu$ in + 0.6R	Gage blocks
Linear Indicators <sup>3</sup> – Dial & Test	Up to 4 in	(3 + 9.4L) $\mu$ in + 0.6R	Gage blocks
Height Gages <sup>3</sup>	Up to 20 in (20 to 48) in	(4.6 + 10L) $\mu$ in + 0.6R (6.8 + 9.8L) $\mu$ in + 0.6R	Gage blocks w/surface plate
Steel Rules <sup>3</sup>	Up to 72 in	(3 + 10L) $\mu$ in + 0.6R	Gage blocks
Tape Measures <sup>3</sup>	Up to 25 ft	(3 + 10L) $\mu$ in + 0.6R	Gage blocks
Angle Indicators & Protractors <sup>3</sup>	30°, 45°, 60°, 75°, 90°	0.03°	Angle block set
Feeler/Thickness Gages <sup>3</sup>	Up to 1 in	85 $\mu$ in	Micrometer
Coating Thickness Gages <sup>3</sup> – Film, Ultrasonic	Up to 60 mils	0.1 mils	Coating thickness standards
Optical Comparators <sup>3</sup>			
Linear Accuracy	Up to 18 in	150 $\mu$ in	Glass scale
Magnification	10x to 250x	0.014 in	Mag checker
Angle	(0 to 90)°	0.1°	Angle block set
Surface Plates <sup>3</sup> – Grades AA, A, & B			
Repeatability Only	0.002 in	40 $\mu$ in	Repeat-o-meter

### III. Dimensional Inspection/Calibration<sup>8</sup>

Parameter/Equipment	Range	CMC <sup>2, 7</sup> ( $\pm$ )	Comments
1D Length <sup>3</sup> – Measure	Up to 3 in	0.000 14 in	Micrometers
	Up to 6 in	0.0013 in	Caliper

### IV. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2, 4, 6</sup> ( $\pm$ )	Comments
DC Voltage – Measure <sup>3</sup>	(0 to 100) mV	7.8 $\mu$ V/V + 0.2 $\mu$ V	Fluke 8588A
	(0.1 to 1.0) V	4.4 $\mu$ V/V + 0.3 $\mu$ V	
DC Voltage – Generate <sup>3</sup>	(1 to 10) V	4.4 $\mu$ V/V + 0.5 $\mu$ V	Vitrek 4700 w/ HVL-100
	(10 to 100) V	6.8 $\mu$ V/V + 30 $\mu$ V	
DC Current – Measure <sup>3</sup>	(100 to 1000) V	7 $\mu$ V/V + 0.5 mV	Fluke 5522A
	(1 to 10) kV	0.05 % + 0.03 V	
DC Current – Generate <sup>3</sup>	(10 to 100) kV	0.07 % + 0.3 V	Empro shunt w/ fluke 8588A
	(0 to 330) mV	17 $\mu$ V/V + 1 $\mu$ V	
DC Current – Measure <sup>3</sup>	330 mV to 3.3 V	9 $\mu$ V/V + 2 $\mu$ V	Fluke 8588A
	(3.3 to 33) V	10 $\mu$ V/V + 20 $\mu$ V	
DC Current – Generate <sup>3</sup>	(33 to 330) V	15 $\mu$ V/V + 150 $\mu$ V	Fluke 374
	(100 to 1020) V	15 $\mu$ V/V + 1.5 mV	
DC Current – Measure <sup>3</sup>	(0 to 10) $\mu$ A	28 $\mu$ A/A + 0.4 nA	Fluke 8588A
	(10 to 100) $\mu$ A	10 $\mu$ A/A + 0.4 nA	
DC Current – Generate <sup>3</sup>	(0.1 to 1.0) mA	10 $\mu$ A/A + 4 nA	Empro shunt w/ fluke 8588A
	(1 to 10) mA	15 $\mu$ A/A + 40 nA	
DC Current – Measure <sup>3</sup>	(10 to 100) mA	58 $\mu$ A/A + 1 $\mu$ A	Fluke 374
	(0.1 to 1.0) A	0.014 % + 0.1mA	
DC Current – Generate <sup>3</sup>	(1 to 10) A	0.024 % + 0.4 mA	Fluke 374
	(1 to 30) A	0.056 % + 4.4 mA	
DC Current – Measure <sup>3</sup>	(1 to 1000) A	0.32 %	Fluke 374
	Up to 600 A	2.8 % + 0.5 A	

Parameter/Equipment	Range	CMC <sup>2, 4, 6</sup> ( $\pm$ )	Comments
DC Current – Generate <sup>3</sup>	(0 to 330) $\mu$ A (0 to 3.3) mA (0 to 33) mA (0 to 330) mA (0 to 1.1) A (1.1 to 3) A (0 to 11) A (11 to 21) A	0.012 % + 0.02 $\mu$ A 0.008 % + 0.05 $\mu$ A 0.01 % + 0.25 $\mu$ A 0.011 % + 2.5 $\mu$ A 0.016 % + 40 $\mu$ A 0.028 % + 40 $\mu$ A 0.04 % + 500 $\mu$ A 0.08 % + 750 $\mu$ A	Fluke 5522A
DC Clamp-On Meters <sup>3</sup> –	Up to 150 A (150 to 1000) A	0.51 % + 0.038 A 0.51 % + 0.51 A	Fluke 5522A w/55xxA coil
DC Power – Generate <sup>3</sup>  33 mV to 1020 V:  (0.33 to 329.99) mA (0.33 to 2.9999) A (3 to 20.5) A	(0.01 to 330) W (0.33 to 3.3) kW (3.3 to 20.5) kW	0.024 % 0.023 % 0.057 %	Fluke 5522A
Resistance – Measure <sup>3</sup>	(0 to 1) $\Omega$ (1 to 10) $\Omega$ (10 to 100) $\Omega$ (0.1 to 1) k $\Omega$ (1 to 10) k $\Omega$ (10 to 100) k $\Omega$ (0.1 to 1) M $\Omega$ (1 to 10) M $\Omega$ (10 to 100) M $\Omega$ (0.1 to 1) G $\Omega$ (1 to 10) G $\Omega$	19 $\mu$ $\Omega$ / $\Omega$ + 4.0 $\mu$ $\Omega$ 11 $\mu$ $\Omega$ / $\Omega$ + 14 $\mu$ $\Omega$ 9.6 $\mu$ $\Omega$ / $\Omega$ + 50 $\mu$ $\Omega$ 9.4 $\mu$ $\Omega$ / $\Omega$ + 0.5 m $\Omega$ 9.4 $\mu$ $\Omega$ / $\Omega$ + 5.0 m $\Omega$ 9.6 $\mu$ $\Omega$ / $\Omega$ + 50 m $\Omega$ 11 $\mu$ $\Omega$ / $\Omega$ + 1 $\Omega$ 24 $\mu$ $\Omega$ / $\Omega$ + 100 $\Omega$ 0.013 % + 10 k $\Omega$ 0.14 % + 1 M $\Omega$ 0.14 % + 10 M $\Omega$	Fluke 8588A

Parameter/Equipment	Range	CMC <sup>2, 4, 6</sup> ( $\pm$ )	Comments
Resistance – Generate <sup>3</sup>	Up to 11 $\Omega$ (11 to 33) $\Omega$ (33 to 110) $\Omega$ 110 $\Omega$ to 1.1 k $\Omega$ (1.1 to 11) k $\Omega$ (11 to 110) k $\Omega$ 110 k $\Omega$ to 1.1 M $\Omega$ (1.1 to 3.3) M $\Omega$ (3.3 to 11) M $\Omega$ (11 to 33) M $\Omega$ (33 to 110) M $\Omega$ (110 to 330) M $\Omega$ (330 to 1100) M $\Omega$	32 $\mu\Omega/\Omega + 0.001 \Omega$ 36 $\mu\Omega/\Omega + 0.0015 \Omega$ 22 $\mu\Omega/\Omega + 0.0014 \Omega$ 22 $\mu\Omega/\Omega + 0.002 \Omega$ 22 $\mu\Omega/\Omega + 0.02 \Omega$ 22 $\mu\Omega/\Omega + 0.2 \Omega$ 25 $\mu\Omega/\Omega + 2 \Omega$ 47 $\mu\Omega/\Omega + 30 \Omega$ 0.01 % + 50 $\Omega$ 0.02 % + 2.5 k $\Omega$ 0.039 % + 3 k $\Omega$ 0.24 % + 100 k $\Omega$ 1.2 % + 500 k $\Omega$	Fluke 5522A
Decade Steps	1 m $\Omega$ to 100 $\Omega$  1 k $\Omega$ to 10 M $\Omega$  10 M $\Omega$ to 100 G $\Omega$	0.026 % + 1 m $\Omega$  0.13 %  1.3 %	IET decade resistor 1433-19-K  IET HRRS decade resistors
Insulation Resistance – Generate <sup>3</sup>			
Fixed Points	0.1 $\Omega$ , 1 $\Omega$ , 100 $\Omega$ , 1 k $\Omega$ , 0 k $\Omega$ , 100 k $\Omega$ , 1 M $\Omega$ , 10 M $\Omega$ , 100 M $\Omega$ , 1 G $\Omega$ 2 G $\Omega$ , 5 G $\Omega$ , 10 G $\Omega$ , 100 G $\Omega$	1.2 %	Standard resistors
Decade Steps	100 M $\Omega$ to 100 G $\Omega$	1.3 %	Decade resistors
Capacitance – Generate <sup>3</sup>	(220.0 to 399.9) pF (0.4 to 1.0999) nF (1.1 to 3.2999) nF (11 to 110) nF (110 to 330) nF (0.33 to 1.1) $\mu$ F (1.1 to 3.3) $\mu$ F (3.3 to 11) $\mu$ F (11 to 33) $\mu$ F (33 to 110) $\mu$ F	0.39 % + 10 pF 0.39 % + 0.01 nF 0.39 % + 0.01 nF 0.2 % + 0.01 nF 0.2 % + 0.1 nF 0.2 % + 0.3 nF 0.2 % + 1 nF 0.2 % + 3 nF 0.2 % + 10 nF 0.32 % + 30 nF	Fluke 5522A

Parameter/Equipment	Range	CMC <sup>2, 4, 6</sup> (±)	Comments
Capacitance – Generate <sup>3</sup> (cont)	(110 to 330) µF 330 µF to 1.1 mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF 330 µF to 1.1 mF	0.36 % + 100 nF 0.36 % + 300 nF 0.36 % + 1 µF 0.37 % + 3 µF 0.37 % + 10 µF 0.6 % + 30 µF 0.87 % + 100 µF	Fluke 5522A
Decade Steps	10 pF to 10nF 10 nF to 10 µF 10 µF to 100 µF	1.5 % + 5 pF 1.3 % 6 %	Time Electronics 1071

Parameter/Equipment	Frequency	CMC <sup>2, 4</sup> (±)	Comments
Capacitance – Measure <sup>3</sup>			
(0.1 to 1) nF (1 to 10) nF (10 to 100) nF (0.1 to 1) µF (1 to 10) µF (10 to 100) µF (0.1 to 1) mF (1 to 10) mF (10 to 100) mF	(50 to 60) Hz (50 to 60) Hz	0.23 % + 0.1 nF 0.13 % + 0.2 nF 0.1 % + 0.01 nF 0.1 % + 0.1 nF 0.1 % + 0.1 nF 0.11 % + 0.01 µF 0.11 % + 0.1 µF 0.12 % + 1 µF 0.12 % + 0.1 mF	Fluke 8588A
AC Voltage – Generate <sup>3</sup>			
(1 to 33) mV	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.063 % + 6 µV 0.014 % + 6 µV 0.017 % + 6 µV 0.078 % + 6 µV 0.28 % + 12 µV 0.63 % + 50 µV	Fluke 5522A
(33 to 330) mV	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.024 % + 8 µV 0.012 % + 8 µV 0.013 % + 8 µV 0.028 % + 8 µV 0.063 % + 32 µV 0.16 % + 70 µV	

Parameter/Equipment	Frequency	CMC <sup>2, 4</sup> (±)	Comments
AC Voltage – Generate <sup>3</sup> (cont)			
330 mV to 3.3 V	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.024 % + 50 µV 0.012 % + 60 µV 0.015 % + 60 µV 0.024 % + 50 µV 0.055 % + 130 µV 0.19 % + 600 µV	Fluke 5522A
(3.3 to 33) V	(10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.024 % + 650 µV 0.012 % + 600 µV 0.019 % + 600 µV 0.028 % + 600 µV 0.07 % + 1.6 mV	
(33 to 330) V	45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.015 % + 2 mV 0.016 % + 6 mV 0.02 % + 6 mV 0.024 % + 6 mV 0.16 % + 50 mV	
(330 to 1020) V	45 Hz to 10 kHz	0.024 % + 10 mV	
AC Voltage – Measure <sup>3</sup>			
(0 to 10) mV	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	0.034 % + 1.1 µV 0.041 % + 1.1 µV 0.042 % + 1.1 µV 0.031 % + 1.1 µV 1.1 % + 4 µV 2.1 % + 4 µV	Fluke 8588A
(10 to 100) mV	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	0.01 % + 0.5 µV 0.014 % + 0.4 µV 0.024 % + 1 µV 0.054 % + 5 µV 0.22 % + 30 µV 1.2 % + 0.1 mV	

Parameter/Range	Frequency	CMC <sup>2, 4</sup> (±)	Comments
AC Voltage – Measure <sup>3</sup> (cont)			
(0.1 to 1) V	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	0.008 % + 5 µV 0.013 % + 5 µV 0.024 % + 10 µV 0.054 % + 50 µV 0.22 % + 0.3 mV 1.1 % + 1 mV	Fluke 8588A
(1 to 10) V	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	0.008 % + 50 µV 0.013 % + 50 µV 0.024 % + 0.1 mV 0.054 % + 0.5 mV 0.22 % + 3 mV 1.1 % + 10 mV	
(10 to 100) V	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	0.016 % + 0.5mV 0.017 % + 0.5mV 0.027 % + 1 mV 0.061 % + 5 mV 0.38 % + 50 mV 1.2 % + 0.5 V	
(100 to 1000) V	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	0.012 % + 25 mV 0.012 % + 25 mV 0.025 % + 25 mV 0.06 % + 0.1 V	
Up to 10 kV	Up to 60 Hz	0.15 % + 0.1 V	Vitrek 4700 w/HVL-100
(10 to 70) kV	60 Hz	0.15 % + 0.6 V	

Parameter/Range	Frequency	CMC <sup>2, 4</sup> (±)	Comments
AC Clamp-On Meters <sup>3</sup> –			
Up to 150 A: Toroidal	(45 to 65) Hz (65 to 440) Hz	0.29 % + 0.026 A 0.8 % + 0.028 A	Fluke 5522A w/ 55xxA coil
Non-Toroidal	(45 to 65) Hz (65 to 440) Hz	0.57 % + 0.25 A 1 % + 0.25 A	
(150 to 1025) A: Toroidal	(45 to 65) Hz (65 to 440) Hz	0.3 % + 0.27 A 0.8 % + 0.27 A	
Non-Toroidal	(45 to 65) Hz (65 to 440) Hz	0.57 % + 0.94 A 1 % + 0.94 A	
AC Current – Generate <sup>3</sup>			
Up to 0.33 mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.16 % + 0.1 µA 0.12 % + 0.1 µA 0.097 % + 0.1 µA 0.24 % + 0.15 µA 0.63 % + 0.2 µA 1.3 % + 0.4 µA	Fluke 5522A
(0.33 to 3.3) mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.16 % + 0.15 µA 0.097 % + 0.15 µA 0.078 % + 0.15 µA 0.16 % + 0.2 µA 0.39 % + 0.3 µA 0.78 % + 0.6 µA	
(3.3 to 33) mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.14 % + 2 µA 0.07 % + 2 µA 0.032 % + 2 µA 0.063 % + 2 µA 0.16 % + 3 µA 0.32 % + 4 µA	
(33 to 330) mA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.14 % + 20 µA 0.07 % + 20 µA 0.032 % + 20 µA 0.078 % + 50 µA 0.16 % + 100 µA 0.32 % + 200 µA	

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Current – Generate <sup>3</sup> (cont)			
(0.33 to 1.1) A	(10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.14 % + 100 µA 0.039 % + 100 µA 0.47 % + 1 mA 2 % + 5 mA	Fluke 5522A
(1.1 to 3.0) A	(10 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.14 % + 100 µA 0.047 % + 100 µA 0.47 % + 1 mA 2 % + 5 mA	
(3.0 to 11) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.047 % + 2mA 0.078 % + 2 mA 3.6 % + 2 mA	
(11 to 20.5) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.094 % + 5 mA 0.12 % + 5 mA 2.4 % + 5 mA	
AC Current – Measure <sup>3</sup>			
(0 to 10) µA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	0.21 % + 2.5 nA 0.21 % + 2.5 nA 0.21 % + 2.5 nA	Fluke 8588A
(10 to 100) µA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	0.029 % + 5 nA 0.054 % + 5 nA 0.075 % + 5 nA 0.41 % + 10 nA	
(0.1 to 1) mA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	0.029 % + 50 nA 0.054 % + 50 nA 0.075 % + 50 nA 0.41 % + 0.1 µA	
(1 to 10) mA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (30 to 100) kHz	0.029 % + 0.5 µA 0.054 % + 0.5 µA 0.075 % + 0.5 µA 0.41 % + 1 µA	
(10 to 100) mA	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	0.029 % + 5 µA 0.053 % + 5 µA 0.075 % + 5 µA	

Parameter/Range	Frequency	CMC <sup>2, 4, 6</sup> (±)	Comments
AC Current – Measure <sup>3</sup> (cont)			
(0.1 to 1) A	1 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	0.031 % + 0.1 mA 0.056 % + 0.1 mA 0.08 % + 0.1 mA	Fluke 8588A
(1 to 10) A	10 Hz to 2 kHz (2 to 10) kHz	0.085 % + 0.5 mA 0.085 % + 0.5 mA	
(10 to 30) A	10 Hz to 2 kHz (2 to 10) kHz	0.085 % + 12 mA 0.13 % + 12 mA	
(1 to 100) A	@ 60 Hz	0.32 %	Empro shunt w/DMM
Up to 600 A	45 to 100 Hz	2.8 % + 0.5 A	Fluke 374

Parameter/Equipment	Range	CMC <sup>2, 4, 6</sup> (±)	Comments
AC Power – Generate <sup>3</sup> (45 to 65) Hz; PF=1			
(33 to 330) mV Range:			
(3.3 to 8.99) mA	110 µW to 3 mW	0.11 %	Fluke 5522A
(9 to 32.99) mA	(3 to 11) mW	0.08 %	
(33 to 89.99) mA	(1.1 to 30) mW	0.11 %	
(90 to 329.99) mA	(3 to 110) mW	0.08 %	
(0.33 to 0.8999) A	(11 to 300) mW	0.11 %	
(0.9 to 2.1999) A	(30 to 730) mW	0.087 %	
(2.2 to 4.4999) A	73 mW to 1.5 W	0.11 %	
(4.5 to 20.5) A	150 mW to 6.8 W	0.087 %	
330 mV to 1020 V Range:			
(3.3 to 8.99) mA	1.1 mW to 9 W	0.095 %	
(9 to 32.99) mA	3 mW to 33 W	0.064 %	
(33 to 89.99) mA	11 mW to 90 W	0.095 %	
(90 to 329.99) mA	30 mW to 330 W	0.064 %	
(0.33 to 0.8999) A	110 mW to 900 W	0.087 %	
(0.9 to 2.1999) A	300 mW to 2200 W	0.072 %	
(2.2 to 4.4999) A	730 mW to 4500 W	0.095 %	
(4.5 to 20.5) A	(1.5 to 20.9) kW	0.08 %	

Parameter/Equipment	Range	CMC <sup>2</sup> ( $\pm$ )	Comments
Thermocouple Simulation <sup>3</sup> –			
Type B	(600 to 800) °C (800 to 1550) °C (1550 to 1820) °C	0.44 °C 0.36 °C 0.3 °C	Fluke 7526A
Type C	(0 to 1000) °C (1000 to 1800) °C (1800 to 2000) °C (2000 to 2316) °C	0.23 °C 0.31 °C 0.34 °C 0.44 °C	
Type E	(-250 to -200) °C (-200 to -100) °C (-100 to 0) °C (0 to 600) °C (600 to 1000) °C	0.32 °C 0.18 °C 0.15 °C 0.14 °C 0.16 °C	
Type J	(-210 to -100) °C (-100 to 800) °C (800 to 1200) °C	0.2 °C 0.15 °C 0.16 °C	
Type K	(-250 to -200) °C (-200 to -100) °C (-100 to 500) °C (500 to 800) °C (800 to 1372) °C	0.56 °C 0.22 °C 0.16 °C 0.16 °C 0.19 °C	
Type N	(-250 to -200) °C (-200 to -100) °C (-100 to 0) °C (0 to 100) °C (100 to 800) °C (800 to 1300) °C	0.89 °C 0.31 °C 0.19 °C 0.17 °C 0.18 °C 0.19 °C	
Type R	(-50 to -25) °C (-25 to 0) °C (0 to 100) °C (100 to 400) °C (400 to 600) °C (600 to 1000) °C (1000 to 1600) °C (1600 to 1767) °C	0.67 °C 0.55 °C 0.48 °C 0.36 °C 0.29 °C 0.28 °C 0.26 °C 0.31 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> ( $\pm$ )	Comments
Thermocouple Simulation <sup>3</sup> – (cont)			
Type S	(-50 to -25) °C (-25 to 0) °C (0 to 100) °C (100 to 400) °C (400 to 600) °C (600 to 1000) °C (1000 to 1600) °C (1600 to 1767) °C	0.63 °C 0.54 °C 0.48 °C 0.37 °C 0.31 °C 0.29 °C 0.3 °C 0.34 °C	Fluke 7526A
Type T	(-250 to -200) °C (-200 to 0) °C (0 to 200) °C (200 to 400) °C	0.44 °C 0.23 °C 0.17 °C 0.17 °C	
Electrical Simulation of RTDs <sup>3</sup>			
Pt 385, 100 Ω	(-200 to 800) °C	0.07 °C	Fluke 7526A
Pt 3926, 100 Ω	(-200 to 630) °C	0.07 °C	
Pt 3916, 100 Ω	(-200 to 630) °C	0.07 °C	
Pt 385, 200 Ω	(-200 to 400) °C (400 to 630) °C	0.48 °C 0.61 °C	
Pt 385, 500 Ω	(-200 to 630) °C	0.21 °C	
Pt 385, 1000 Ω	(-200 to 630) °C	0.12 °C	
Ni 120, 120 Ω	(-80 to 260) °C	0.05 °C	
Cu 427, 10 Ω	(-100 to 260) °C	0.46 °C	
YSI 400	(15 to 50) °C	0.04 °C	

Parameter/Equipment	Range	CMC <sup>2, 4</sup> ( $\pm$ )	Comments
RTD – Measure <sup>3</sup>			
Pt 385, 100 $\Omega$	(-200 to 800) °C	0.06 °C	Fluke 7526A
Pt 3926, 100 $\Omega$	(-200 to 630) °C	0.06 °C	
Pt 3916, 100 $\Omega$	(-200 to 630) °C	0.06 °C	
Pt 385, 200 $\Omega$	(-200 to 400) °C (400 to 630) °C	0.1 °C 0.12 °C	
Pt 385, 500 $\Omega$	(-200 to 630) °C	0.07 °C	
Pt 385, 1000 $\Omega$	(-200 to 630) °C	0.06 °C	
Ni 120, 120 $\Omega$	(-80 to 260) °C	0.04 °C	
Cu 427, 10 $\Omega$	(-100 to 260) °C	0.14 °C	
YSI 400	(15 to 50) °C	0.02 °C	
Oscilloscopes <sup>3</sup> –			
Square Wave Signal:			
50 $\Omega$ Load @ 1 kHz 1 M $\Omega$ Load @ 1 kHz	1 mV to 6.6 V <sub>pk - pk</sub> 1 mV to 130 V <sub>pk - pk</sub>	0.26 % + 40 $\mu$ V 0.12 % + 40 $\mu$ V	Fluke 5522A w/ SC1100
DC Volt Amplitude:			
50 $\Omega$ Load 1 M $\Omega$ Load	(0 to 6.6) V (0 to 130) V	0.26 % + 40 $\mu$ V 0.05 % + 40 $\mu$ V	
Level Sine Wave:			
Frequency	5 kHz to 1100 MHz	2.8 $\mu$ Hz/Hz	
Amplitude	50 kHz Reference 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz (300 to 1100) MHz	2 % + 300 $\mu$ V 3.5 % + 300 $\mu$ V 4 % + 300 $\mu$ V 6 % + 300 $\mu$ V 7 % + 300 $\mu$ V	
Flatness (Bandwidth)	0 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz (300 to 1100) MHz	1.5 % + 100 $\mu$ V 2 % + 100 $\mu$ V 4 % + 100 $\mu$ V 5 % + 100 $\mu$ V	

Parameter/Equipment	Range	CMC <sup>2</sup> ( $\pm$ )	Comments
Oscilloscopes <sup>3</sup> – (cont)			
Time Markers:			
Into a 50 $\Omega$ load	5 s to 50 ms 20 ms to 2 ns	(26 + 1000 $t$ ) $\mu$ s/s 2.9 $\mu$ s/s	$t$ = time in seconds
Rise Time:			
1 kHz to 2 MHz (2 to 10) MHz	$\leq$ 300 ps $\leq$ 350 ps	110 ps 110 ps	

## V. Fluid Quantities

Parameter/Equipment	Range	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
Fume Hoods – Air Velocity Only <sup>3</sup>	(25 to 200) ft/min	4.5 %	Anemometer
Viscosity Meters <sup>3</sup>	Up to 16 000 cP 25 000 cP 32 000 cP	1.5 cP 90 cP 150 cP	Standard viscosity solution w/ bath
Flow Meters <sup>3</sup> – Totalizers	Up to 50 gallons	0.7 %	Gravimetric method

## VI. Mechanical

Parameter/Equipment	Range	CMC <sup>2, 5, 6, 7</sup> ( $\pm$ )	Comments
Force – Measure & Measuring Equipment <sup>3</sup>	Up to 1000 lbf Up to 10 000 lbf	0.017 % + 0.6R 0.14 %	Class F weights Load cells w/ indicator
Scales & Balances <sup>3</sup>	(1 to 20 000) g (> 20 to 5000) kg Up to 1000 lb (1000 to 120 000) lb  (1 to 500) mg Up to 5 g Up to 10 g Up to 30 g Up to 50 g Up to 100 g Up to 200 g Up to 300 g Up to 500 g Up to 1000 g Above 1000 g	0.017 % + 0.6R 0.017 % + 0.6R 0.017 % + 0.6R 0.017 % per 20 000 lb + 0.6R  0.013 mg + 0.6R 0.043 mg + 0.6R 0.062 mg + 0.6R 0.092 mg + 0.6R 0.17 mg + 0.6R 0.31 mg + 0.6R 0.63 mg + 0.6R 0.93 mg + 0.6R 1.5 mg + 0.6R 3.1 mg + 0.6R 3.1 mg per 1000 g + 0.6R	Class F weights (applied load)  Class 1 weights (applied load)
Torque – Torque Wrenches & Tools – Measure <sup>3</sup>	5 ozf·in to 600 lbf·ft	0.65 %	CDI Suretest 5000-ST
Rotary Torque Tools – Measure <sup>3</sup>	(0.02 to 2) N·m (0.5 to 5) N·m (1 to 10) N·m (2 to 20) N·m (0.75 to 75) N·m (18 to 180) N·m (50 to 500) N·m	0.026 N·m 0.065 N·m 0.13 N·m 0.26 N·m 0.98 N·m 2.4 N·m 6.5 N·m	Crane – torque star w/rotary transducers
Torque – Measuring Equipment <sup>3</sup>	Up to 1000 lbf·ft	0.13 %	Class F weights & arms
Vacuum Gages <sup>3</sup>	(0.01 to 28.5) in·Hg	0.056 in·Hg	Fluke 2700G- BG200K

Parameter/Equipment	Range	CMC <sup>2, 6, 7</sup> (±)	Comments
Barometric Pressure <sup>3</sup>	(0 to 30) psia	0.037 psia	Druck DPI 705
Pressure – Measuring, Measuring Equipment <sup>3</sup> –			
Pneumatic	(-15 to 15) kPa (20 to 700) kPa ATM to 7000 kPa Up to 30 psig Up to 100 psig Up to 500 psig	Greater of 0.01 % RDG or 0.9 pa Greater of 0.013 % RDG or 0.061 kPa Greater of 0.013 % RDG or 0.61 kPa 0.022 psig 0.035 psig 0.13 psig	Fluke 8270A w/PM600 modules      Fluke 2700G Module  Heise PPM-2  Fluke 2700G Module
Hydraulic	Up to 3000 psig Up to 5000 psig Up to 10 000 psig Up to 15 000 psig	0.8 psig 1.3 psig 7 psig 11 psig	Fluke 2700G modules
Indirect Verification of Microindentation Hardness Testers <sup>3</sup> (Vickers)	(100 to 240) HV (240 to 600) HV	2.9 HV 11 HV	ASTM E92
Speed <sup>3</sup> – Optic/Non-Contact: RPM Totalizer/Rate Meters	(6 to 200 000) rpm	0.017 %	Monarch PLT200
Contact: RPM Totalizer/Rate Meters	(0.5 to 20 000) rpm	0.22 %	Monarch PLT200 w/ 10 cm wheel
Speed/RPM/Rate Simulation	(6 to 200 000) rpm	0.003 %	Agilent 33120A

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Indirect Verification of Rockwell Hardness Testers <sup>3</sup>	HRC: Low Medium High  HRBW: Low Medium High  HRA: Low Medium High  HREW Low Medium High  HR15N Low Medium High  HR30N Low Medium High  HR45N Low Medium High  HR15TW Low Medium High  HR30TW Low Medium High  HR45TW Low Medium High	0.92 HRC 0.92 HRC 0.92 HRC  1.6 HRBW 1.2 HRBW 1 HRBW  0.82 HRA 0.81 HRA 0.81 HRA  0.81 HREW 0.81 HREW 0.81 HREW  0.81 HR15N 0.81 HR15N 0.81 HR15N  0.82 HR30N 0.82 HR30N 0.82 HR30N  0.82 HR45N 0.81 HR45N 0.82 HR45N  0.81 HR15TW 0.81 HR15TW 0.82 HR15TW  0.82 HR30TW 0.81 HR30TW 0.81 HR30TW  0.81 HR45TW 0.81 HR45TW 0.81 HR45TW	Indirect verification per ASTM E18

## VII. Optical Quantities

Parameter/Equipment	Range	CMC <sup>2, 6, 7</sup> ( $\pm$ )	Comments
Light Booths <sup>3</sup> – Illuminance – Measure Correlated Color Temperature (CCT) – Measure	(10 to 5000) Lux (1700 to 6500) K	2.7 % 100 K	Illuminance spectrophotometer

## VIII. Thermodynamics

Parameter/Equipment	Range	CMC <sup>2, 7</sup> ( $\pm$ )	Comments
Relative Humidity – Measure <sup>3</sup>	(10 to 90) % RH	1.2 % RH	Rotronic HC2A-SH
Relative Humidity – Measuring Equipment <sup>3</sup>	(10 to 90) % RH	1.3 % RH	Rotronic HC2A-SH
Temperature – Measure <sup>3</sup>	(-200 to 420) °C (420 to 660) °C (600 to 1200) °C	0.05 °C 0.086 °C 0.3 °C	Fluke 1524 w/ 5609 PRT probe Fluke 1524 w/ S type TC probe
Temperature – Measuring Equipment <sup>3</sup>	(-75 to 0) °C (-25 to 125) °C (50 to 420) °C (420 to 660) °C (200 to 700) °C (700 to 950) °C (950 to 1200) °C (-25 to 125) °C (50 to 420) °C (420 to 660) °C	0.31 °C 0.076 °C 0.11 °C 0.17 °C 1.9 °C 2.2 °C 2.9 °C 0.3 °C 0.42 °C 0.6 °C	Freezer w/ ref probe Oil bath w/ ref probe Dry well w/ ref probe Furnace w/ ref TC probe Fluke 7103 Fluke 9144

Parameter/Equipment	Range	CMC <sup>2, 7</sup> ( $\pm$ )	Comments
Infrared Thermometry <sup>3</sup> – Measuring Equipment	(35 to 100) °C (100 to 200) °C (200 to 350) °C (350 to 500) °C	1 °C 1.2 °C 1.7 °C 2.3 °C	Fluke 4181 $\lambda = (8 \text{ to } 14) \mu\text{m}$ , where $\epsilon = 0.95$

## IX. Time & Frequency

Parameter/Equipment	Range	CMC <sup>2, 6, 7</sup> ( $\pm$ )	Comments
Timers & Stopwatches <sup>3</sup>	(1 to 3600) s	0.2 s	Stopwatch
Frequency – Generate <sup>3</sup>	0.01 Hz to 2 MHz Up to 15 MHz	30 $\mu\text{Hz}/\text{Hz} + 5 \mu\text{Hz}$ 0.003 %	Fluke 5500A HP 33120A
Frequency – Measure <sup>3</sup>	0.1 Hz to 10 MHz (10 to 225) MHz 225 MHz to 3 GHz	67 $\mu\text{Hz}/\text{Hz}$ 0.45 $\mu\text{Hz}/\text{Hz}$ 0.26 $\mu\text{Hz}/\text{Hz}$	HP53181A opt 010

<sup>1</sup> This laboratory offers commercial calibration and field calibration services, where noted.

<sup>2</sup> Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. Calibration and Measurement Capabilities represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

<sup>5</sup> In the statement of the CMC,  $L$  is the numerical value of the nominal length of the device measured in inches,  $R$  is the numerical value of the resolution of the device in microinches.

<sup>6</sup> In the statement of CMC, percentages are percentage of reading, unless otherwise indicated.

<sup>7</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

<sup>8</sup> This laboratory meets *R205 – Specific Requirements: Calibration Laboratory Accreditation Program* for the types of dimensional tests listed above and is considered equivalent to that of a calibration.

<sup>9</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



# Accredited Laboratory

A2LA has accredited

**CROSS TECHNOLOGIES, INC.**

*Calvert City, KY*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets the requirements of R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system  
(refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 11<sup>th</sup> day of March 2024.

A blue ink signature of the name "Mr. Trace McInturff".

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1741.14  
Valid to May 31, 2026

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*