



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017,  
ANSI/NCSL Z540-1-1994 & ANSI/NCSL Z540.3-2006

TEKTRONIX, INC.  
1185 Campbell Ave. Suite 3  
San Jose, CA 95126  
Michael Rivera Phone: 619 448 0334

CALIBRATION

Valid To: March 31, 2023

Certificate Number: 2357.05

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory at the location listed above performing the following calibrations<sup>1, 6</sup>:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
DC Voltage – Generate <sup>3</sup>	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	9.1 μV/V + 0.62 μV 6.3 μV/V + 0.93 μV 6.2 μV/V + 3.1 μV 6.3 μV/V + 6.2 μV 7.0 μV/V + 78 μV 8.6 μV/V + 0.47 mV	Fluke 5720A
DC Voltage – Measure <sup>3</sup>	Up to 100 mV (0.1 to 1.2) V (1.2 to 10) V (10 to 100) V (100 to 1000) V	12 μV/V + 230 nV 7.3 μV/V + 0.23 μV 4.0 μV/V + 0.39 μV 6.4 μV/V + 23 μV 7.9 μV/V + 78 μV	HP 3458A, opt 002
DC Current – Generate <sup>3</sup>	Up to 220 μA 220 μA to 2.2 mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A (2.2 to 11) A  (1.1 to 2.99999) A (0 to 10.9999) A (11 to 20.5) A	41 μA/A + 5.4 nA 33 μA/A + 6.2 nA 33 μA/A + 39 nA 41 μA/A + 0.62 μA 71 μA/A + 12 μA 0.28 mA/A + 0.37 mA  0.30 mA/A + 31 μA 0.39 mA/A + 0.39 mA 0.78 mA/A + 0.58 mA	Fluke 5720A /5725A  Fluke 5520A w/ SC1100

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
DC Current – Measure <sup>3</sup>	Up to 100 nA (0.1 to 1) μA (1 to 10) μA (10 to 100) μA (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	69 μA/A + 40 pA 36 μA/A + 40 pA 23 μA/A + 0.10 nA 25 μA/A + 0.80 nA 25 μA/A + 5.0 nA 25 μA/A + 50 nA 47 μA/A + 500 nA 0.013 % + 10 μA	HP 3458A
DC Resistance – Generate <sup>3</sup>	(0 to 10.9999) Ω (11 to 32.9999) Ω (33 to 109.9999) Ω (110 to 329.9999) Ω (0.33 to 1.099999) kΩ (1.1 to 3.299999) kΩ (3.3 to 10.99999) kΩ (11 to 32.99999) kΩ (33 to 109.9999) kΩ (110 to 329.9999) kΩ (0.33 to 1.099999) MΩ (1.1 to 3.299999) MΩ (3.3 to 10.99999) MΩ (11 to 32.99999) MΩ (33 to 109.9999) MΩ (110 to 329.9999) MΩ (330 to 1100) MΩ	46 μΩ/Ω + 0.78 mΩ 52 μΩ/Ω + 1.2 mΩ 34 μΩ/Ω + 1.1 mΩ 29 μΩ/Ω + 1.6 mΩ 27 μΩ/Ω + 1.6 mΩ 28 μΩ/Ω + 16 mΩ 27 μΩ/Ω + 16 mΩ 29 μΩ/Ω + 0.16 Ω 27 μΩ/Ω + 0.16 Ω 25 μΩ/Ω + 1.6 Ω 26 μΩ/Ω + 1.6 Ω 49 μΩ/Ω + 23 Ω 0.11 mΩ/Ω + 39 Ω 0.20 mΩ/Ω + 1.9 kΩ 0.40 mΩ/Ω + 2.3 kΩ 2.4 mΩ/Ω + 78 kΩ 12 mΩ/Ω + 390 kΩ	Fluke 5520A w/ SC1100
Fixed Points	1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	0.13 mΩ 0.18 mΩ 0.23 mΩ 0.43 mΩ 1.1 mΩ 2.0 mΩ 8.2 mΩ 16 mΩ 85 mΩ 0.15 Ω 1.0 Ω 2.7 Ω 50 Ω 60 Ω 4.0 kΩ 8.5 kΩ 26 kΩ	Fluke 5720A

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
DC Resistance – Measure <sup>3</sup>	(0 to 10) Ω (10 to 100) Ω 100 Ω to 1 kΩ (1 to 10) kΩ (10 to 100) kΩ 100 kΩ to 1 MΩ (1 to 10) MΩ (10 to 100) MΩ 100 MΩ to 1.2 GΩ	12 μΩ/Ω + 39 μΩ 9.5 μΩ/Ω + 0.39 mΩ 8.0 μΩ/Ω + 0.39 mΩ 7.8 μΩ/Ω + 3.9 mΩ 8.0 μΩ/Ω + 39 mΩ 12 μΩ/Ω + 1.6 Ω 39 μΩ/Ω + 78 Ω 0.40 mΩ/Ω + 0.78 kΩ 4.7 mΩ/Ω + 7.8 kΩ	HP 3458A, opt 002
AC Voltage – Generate <sup>3</sup>			
Up to 2.2 mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	1.2 mV/V + 5.0 μV 0.91 mV/V + 5.0 μV 0.9 mV/V + 5.0 μV 1.5 mV/V + 5.0 μV 2.1 mV/V + 6.0 μV 3.6 mV/V + 12 μV 5.2 mV/V + 25 μV 6.6 mV/V + 25 μV	Fluke 5720A w/ Fluke 5725A
(2.2 to 22) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.32 mV/V + 5.0 μV 0.18 mV/V + 5.0 μV 0.18 mV/V + 5.0 μV 0.33 mV/V + 5.0 μV 0.59 mV/V + 6.0 μV 1.3 mV/V + 12 μV 1.7 mV/V + 25 μV 3.2 mV/V + 25 μV	
(22 to 220) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.39 mV/V + 15 μV 99 μV/V + 8.0 μV 87 μV/V + 8.0 μV 0.21 mV/V + 8.0 μV 0.49 mV/V + 20 μV 0.89 mV/V + 25 μV 1.4 mV/V + 30 μV 2.7 mV/V + 60 μV	

Parameter/Range	Frequency	CMC <sup>2, 4</sup> (±)	Comments
AC Voltage – Generate <sup>3</sup> (cont)			
220 mV to 2.2 V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.54 mV/V + 50 μV 88 μV/V + 20 μV 46 μV/V + 10 μV 80 μV/V + 12 μV 0.14 mV/V + 40 μV 0.42 mV/V + 0.10 mV 0.96 mV/V + 0.25 mV 1.7 mV/V + 0.40 mV	Fluke 5720A w/ Fluke 5725A
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.40 mV/V + 0.50 mV 95 μV/V + 0.20 mV 48 μV/V + 70 μV 82 μV/V + 0.12 mV 0.12 mV/V + 0.25 mV 0.30 mV/V + 0.80 mV 1.0 mV/V + 2.5 mV 1.7 mV/V + 4.0 mV	
(22 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.46 mV/V + 5.0 mV 99 μV/V + 2.0 mV 71 μV/V + 0.70 mV 0.15 mV/V + 1.2 mV 0.23 mV/V + 3.0 mV 0.87 mV/V + 20 mV 4.2 mV/V + 50 mV 8.6 mV/V + 100 mV	Fluke 5720A w/ Fluke 5725A
(220 to 250) V	(15 to 50) Hz	0.28 mV/V + 20 mV	
(220 to 1100) V	50 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	79 μV/V + 4.0 mV 0.13 mV/V + 6.0 mV 0.49 mV/V + 11 mV	
(220 to 750) V	(30 to 50) kHz (50 to 100) kHz	0.47 mV/V + 11 mV 1.8 mV/V + 45 mV	

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Voltage – Generate <sup>3</sup> (cont)			
Wideband Output Up to 1.1 mV	(10 to 30) Hz 30 Hz to 120 kHz (0.12 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	2.4 mV/V 0.98 mV/V 2.2 mV/V + 2.3 μV 3.7 mV/V + 2.3 μV 5.5 mV/V + 2.3 μV 13 mV/V + 12 μV	Fluke 5720A
(1.1 to 3) mV	(10 to 30) Hz 30 Hz to 120 kHz (0.12 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	2.4 mV/V 0.94 mV/V 1.3 mV/V + 2.3 μV 2.2 mV/V + 2.3 μV 4.9 mV/V + 2.3 μV 13 mV/V + 2.3 μV	
(3 to 11) mV	(10 to 30) Hz 30 Hz to 120 kHz (0.12 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	2.4 mV/V 0.94 mV/V 1.12 mV/V + 2.3 μV 2.0 mV/V + 2.3 μV 3.9 mV/V + 2.3 μV 8.6 mV/V + 2.3 μV	
(11 to 33) mV	(10 to 30) Hz 30 Hz to 120 kHz (0.12 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	2.4 mV/V 0.91 mV/V 1.0 mV/V + 2.3 μV 2.0 mV/V + 2.3 μV 3.8 mV/V + 2.3 μV 8.5 mV/V + 2.3 μV	
(33 to 110) mV	(10 to 30) Hz 30 Hz to 120 kHz (0.12 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	2.4 mV/V 0.87 mV/V 1.1 mV/V + 2.3 μV 2.0 mV/V + 2.3 μV 3.8 mV/V + 2.3 μV 8.1 mV/V + 2.3 μV	

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Voltage – Generate <sup>3</sup> (cont)			
Wideband Output (110 to 330) mV	(10 to 30) Hz 30 Hz to 120 kHz (0.12 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	2.3 mV/V 0.84 mV/V 1.1 mV/V + 2.3 μV 2.0 mV/V + 2.3 μV 3.8 mV/V + 2.3 μV 8.5 mV/V + 2.3 μV	Fluke 5720A
330 mV to 1.1 V	(10 to 30) Hz 30 Hz to 120 kHz (0.12 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	2.3 mV/V 0.84 mV/V 1.0 mV/V + 2.3 μV 2.0 mV/V + 2.3 μV 3.8 mV/V + 2.3 μV 8.5 mV/V + 2.3 μV	
(1.1 to 3.5) V	(10 to 30) Hz 30 Hz to 120 kHz (0.12 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	2.3 mV/V 0.84 mV/V 1.0 mV/V + 2.3 μV 2.0 mV/V + 2.3 μV 3.8 mV/V + 2.3 μV 8.5 mV/V + 2.3 μV	
AC Voltage – Measure <sup>3</sup>			
(1 to 10) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (0.10 to 1) MHz	0.38 μV/mV + 2.3 μV 0.19 μV/mV + 0.85 μV 0.25 μV/mV + 0.85 μV 0.79 μV/mV + 0.85 μV 3.9 μV/mV + 0.85 μV 31 μV/mV + 3.9 μV	HP 3458A
(10 to 100) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 2) MHz	0.3 μV/mV + 3.1 μV 0.09 μV/mV + 1.6 μV 0.12 μV/mV + 1.6 μV 0.26 μV/mV + 1.6 μV 0.63 μV/mV + 1.6 μV 2.3 μV/mV + 7.8 μV 7.9 μV/mV + 7.8 μV	

Parameter/Range	Frequency	CMC <sup>2, 4</sup> (±)	Comments
AC Voltage – Measure <sup>3</sup> (cont)			
(0.1 to 1) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 2) MHz	0.28 mV/V + 31 μV 0.07 mV/V + 16 μV 0.12 mV/V + 16 μV 0.26 mV/V + 16 μV 0.63 mV/V + 16 μV 2.3 mV/V + 78 μV 7.9 mV/V + 78 μV	HP 3458A
(1 to 10) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 2) MHz	0.28 mV/V + 0.31 mV 0.07 mV/V + 0.16 mV 0.12 mV/V + 0.16 mV 0.26 mV/V + 0.16 mV 0.63 mV/V + 0.16 mV 2.3 mV/V + 0.78 mV 7.9 mV/V + 0.78 mV	
(10 to 100) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	0.31 mV/V + 3.1 mV 0.16 mV/V + 1.6 mV 0.16 mV/V + 1.6 mV 0.3 mV/V + 1.6 mV 0.94 mV/V + 1.6 mV 3.1 mV/V + 7.8 mV 12 mV/V + 7.8 mV	
(100 to 700) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.42 mV/V + 31 mV 0.31 mV/V + 16 mV 0.47 mV/V + 16 mV 0.94 mV/V + 16 mV 2.3 mV/V + 16 mV	
AC Current – Generate <sup>3</sup>			
Up to 220 μA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.4 mA/A + 16 nA 0.17 mA/A + 10 nA 0.12 mA/A + 8 nA 0.33 mA/A + 12 nA 1.0 mA/A + 65 nA	Fluke 5720A
220 μA to 2.2 mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.36 mA/A + 40 nA 0.19 mA/A + 35 nA 0.15 mA/A + 35 nA 0.22 mA/A + 110 nA 1.0 mA/A + 650 nA	
(2.2 to 22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.37 mA/A + 400 nA 0.17 mA/A + 350 nA 0.12 mA/A + 350 nA 0.20 mA/A + 550 nA 1.0 mA/A + 5.0 μA	

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Current – Generate <sup>3</sup> (cont)			
(22 to 220) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.38 mA/A + 4.0 µA 0.17 mA/A + 3.5 µA 0.13 mA/A + 2.5 µA 0.20 mA/A + 3.5 µA 1.0 mA/A + 10 µA	Fluke 5720A
220 mA to 2.2 A	20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.27 mA/A + 35 µA 0.41 mA/A + 80 µA 6.3 mA/A + 160 µA	Fluke 5520A w/ SC1100
(2.2 to 11) A	40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.4 mA/A + 170 µA 0.76 mA/A + 380 µA 2.9 mA/A + 750 µA	
(1.1 to 2.99999) A	Up to 45 Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	1.4 mA/A + 78 µA 0.49 mA/A + 78 µA 4.7 mA/A + 0.78 mA 19 mA/A + 3.9 mA	
(3 to 10.9999) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.47 mA/A + 1.6 mA 0.78 mA/A + 1.6 mA 23 mA/A + 1.6 mA	
(11 to 20.5) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.94 mA/A + 3.9 mA 1.2 mA/A + 3.9 mA 23 mA/A + 3.9 mA	
AC Current – Measure <sup>3</sup>			
(0 to 100) µA	(10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz 100 Hz to 5 kHz	0.47 % + 30 nA 0.18 % + 30 nA 0.071 % + 30 nA 0.07 % + 30 nA	HP 3458A
(0.1 to 1) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz (5 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.47 % + 200 nA 0.18 % + 200 nA 0.071 % + 200 nA 0.036 % + 200 nA 0.071 % + 200 nA 0.47 % + 400 nA 0.65 % + 1.5 µA	



Parameter/Range	Frequency	CMC <sup>2,4,5</sup> (±)	Comments
AC Current – Measure <sup>3</sup> (cont)			
(1 to 10) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz (5 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.47 % + 2 μA 0.18 % + 2 μA 0.071 % + 2 μA 0.036 % + 2 μA 0.071 % + 2 μA 0.47 % + 4 μA 0.65 % + 15 μA	HP 3458A
(10 to 100) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz (5 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.47 % + 20 μA 0.18 % + 20 μA 0.071 % + 20 μA 0.036 % + 20 μA 0.071 % + 20 μA 0.47 % + 40 μA 0.65 % + 150 μA	
(0.1 to 1) A	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz (5 to 20) kHz (20 to 50) kHz	0.47 % + 200 μA 0.19 % + 200 μA 0.095 % + 200 μA 0.12 % + 200 μA 0.35 % + 200 μA 1.2 % + 400 μA	

Parameter/Equipment	Range	CMC <sup>2,4,5</sup> (±)	Comments
Oscilloscopes <sup>3</sup> –			
Square Wave 50 Ω Load	±1 mV to ±6.6 V <sub>p-p</sub> 10 Hz to 10 kHz	1.6 mV/V + 31 μV	Fluke 5520A/SC1100;
1 MΩ Load	±1 mV to ±130 V <sub>p-p</sub> 10 Hz to 10 kHz	0.78 mV/V + 31 μV	
Rise Time	Nominal 300 ps	19 ps	
Leveled Sine Wave Relative to 50 kHz [5 mV to 5.5 V] <sub>p-p</sub>	5 mV to 5.5 V <sub>p-p</sub> 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz	2.8 % 3.0 % 4.0 %	
[5 mV to 3.5 V] <sub>p-p</sub>	5 mV to 3.5 V <sub>p-p</sub> (600 to 1100) MHz	4.7 %	

Parameter/Equipment	Range	CMC <sup>2, 4, 5</sup> (±)	Comments
Oscilloscopes <sup>3</sup> – (cont)			
Time Marker into 50 Ω Load – Generate	Cardinal Points 1 ns to 20 ms	2.1 μs/s	Fluke 5520A/SC1100;
	Non-Cardinal Points 1 ns to 20 ms	39 μs/s	
	Any value in range 50 ms to 5 s	(19 + 39t) μs/s	t = time in seconds

Parameter/Range	Frequency	CMC <sup>2, 4</sup> (±)	Comments
Capacitance – Generate <sup>3</sup>			
(0.19 to 1.1) nF	10 Hz to 10 kHz	4.1 pF/nF + 7.8 pF	Fluke 5520A w/ SC1100
(1.1 to 3.3) nF	10 Hz to 3 kHz	4.0 pF/nF + 7.8 pF	
(3.3 to 11) nF	(0.01 to 1) kHz	2.3 pF/nF + 7.8 pF	
(11 to 110) nF	(0.01 to 1) kHz	2.3 pF/nF + 78 pF	
(110 to 330) nF	(0.01 to 1) kHz	2.3 pF/nF + 0.23 nF	
(0.33 to 1.1) μF	(10 to 600) Hz	2.3 nF/μF + 0.78 nF	
(1.1 to 3.29) μF	(10 to 300) Hz	2.3 nF/μF + 2.3 nF	
(3.3 to 11) μF	(10 to 150) Hz	2.3 nF/μF + 7.8 nF	
(11 to 33) μF	(10 to 120) Hz	3.4 nF/μF + 23 nF	
(33 to 110) μF	(10 to 80) Hz	3.7 nF/μF + 78 nF	
(110 to 330) μF	(10 to 50) Hz	3.5 nF/μF + 0.23 μF	
(0.33 to 1.1) mF	(10 to 20) Hz	3.5 μF/mF + 0.78 μF	
(1.1 to 3.3) mF	(0 to 6) Hz	3.5 μF/mF + 2.3 μF	
(3.3 to 11) mF	(0 to 2) Hz	3.5 μF/mF + 7.8 μF	
(11 to 33) mF	(0 to 0.6) Hz	5.8 μF/mF + 23 μF	
(33 to 110) mF	(0 to 0.2) Hz	8.5 μF/mF + 78 μF	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Calibration of Thermocouple Indicators <sup>3</sup> –			
Type B	(600 to 800) °C (800 to 1000) °C (1000 to 1550) °C (1550 to 1820) °C	0.34 °C 0.27 °C 0.24 °C 0.26 °C	Fluke 5520A/SC1100
Type C	(0 to 150) °C (150 to 650) °C (650 to 1000) °C (1000 to 1800) °C (1800 to 2316) °C	0.24 °C 0.21 °C 0.25 °C 0.39 °C 0.65 °C	
Type E	(-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1000) °C	0.39 °C 0.13 °C 0.11 °C 0.13 °C 0.17 °C	
Type J	(-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C	0.25 °C 0.13 °C 0.12 °C 0.14 °C 0.18 °C	
Type K	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1000) °C (1000 to 1372) °C	0.26 °C 0.14 °C 0.13 °C 0.20 °C 0.31 °C	
Type N	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1300) °C	0.31 °C 0.17 °C 0.15 °C 0.14 °C 0.21 °C	
Type R	(0 to 250) °C (250 to 400) °C (400 to 1000) °C (1000 to 1767) °C	0.45 °C 0.28 °C 0.26 °C 0.32 °C	
Type S	(0 to 250) °C (250 to 1000) °C (1000 to 1400) °C (1400 to 1767) °C	0.38 °C 0.28 °C 0.29 °C 0.36 °C	
Type T	(-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C	0.50 °C 0.19 °C 0.13 °C 0.11 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Calibration of Thermocouple Indicators <sup>3</sup> (cont) –  Type U	(-200 to 0) °C (0 to 600) °C	0.57 °C 0.29 °C	Fluke 5520A/SC1100
Electrical Calibration of RTDs <sup>3</sup> –  Pt 385 (100 Ω)	(-200 to -80) °C (-80 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C	0.085 °C 0.12 °C 0.11 °C 0.097 °C 0.11 °C 0.20 °C	Fluke 5520A/SC 1100
Pt 3926 (100 Ω)	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C	0.064 °C 0.076 °C 0.075 °C 0.089 °C 0.095 °C 0.17 °C	
Pt 3916 (100 Ω)	(-200 to -190) °C (-190 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.21 °C 0.060 °C 0.068 °C 0.070 °C 0.077 °C 0.084 °C 0.090 °C 0.13 °C 0.19 °C	
Pt 385 (200 Ω)	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.041 °C 0.043 °C 0.044 °C 0.051 °C 0.098 °C 0.11 °C 0.11 °C 0.13 °C	
Pt 385 (500 Ω)	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 400) °C (400 to 600) °C (600 to 630) °C	0.036 °C 0.043 °C 0.044 °C 0.051 °C 0.066 °C 0.073 °C 0.088 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Calibration of RTDs <sup>3</sup> – (cont)			
Pt 385 (1000 Ω)	(-200 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.029 °C 0.036 °C 0.042 °C 0.050 °C 0.18 °C 0.057 °C 0.18 °C	Fluke 5520A/SC 1100
PtNi 385 (120 Ω)	(-80 to 0) °C (0 to 260) °C	0.081 °C 0.11 °C	
Cu 427 (10 Ω)	(-100 to 260) °C	0.69 °C	

## II. Mechanical

Parameter/Equipment	Range	CMC <sup>2, 7</sup> (±)	Comments
Torque <sup>3</sup> – Measuring Equipment	(1 to 10) in·lbf (10 to 50) in·lbf (25 to 250) in·lbf (250 to 1000) in·lbf (83 to 250) ft·lbf (250 to 600) ft·lbf	0.024 in·lbf 0.15 in·lbf 1.2 in·lbf 3.0 in·lbf 0.73 ft·lbf 1.9 ft·lbf	CDI 2000 4 to 1 Transducer  CDI 600 ft·lbf xducer

<sup>1</sup> This accreditation covers commercial calibration service and field calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

- <sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- <sup>4</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC are expressed as either a specific value that covers the full range or as a fraction/percent of the reading plus a fixed floor specification.
- <sup>5</sup> In the statement of CMC, percentages are to be read as percent of reading unless otherwise noted. In the statement of CMC, LSD is the least significant digit.
- <sup>6</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.
- <sup>7</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



## Accredited Laboratory

A2LA has accredited

**TEKTRONIX, INC.**

*San Jose, CA*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NC SL Z540-1-1994 and the requirements of ANSI/NC SLI Z540.3-2006 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 26<sup>th</sup> day of January 2021.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 2357.05  
Valid to May 31, 2023  
Revised January 5, 2022

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*