



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

AUTO TECHNOLOGY COMPANY
20026 Progress Drive
Strongsville, OH 44149
Kevin Smith Phone: 440 572 7800

CALIBRATION

Valid To: December 31, 2023

Certificate Number: 2563.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 6}:

I. Chemical

Parameter/Equipment	Range	CMC ^{2, 7} (±)	Comments
Gas Analyzer Calibration ³ –			
H ₂ S	(0 to 2000) parts per 10 ⁹	2.6 %	Mass flow meter /w calibration gas
SO ₂	(0 to 2000) parts per 10 ⁹	2.6 %	
NO ₂	(0 to 2000) parts per 10 ⁹	3.3 %	
Cl ₂	(0 to 50) parts per 10 ⁹	5.4 %	

II. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2, 4} (±)	Comments
DC Voltage ³ – Measure	(0 to 110) mV (0.11 to 1.1) V (1.1 to 11) V (11 to 110) V (110 to 300) V	0.029 % + 20 µV 0.029 % + 0.065 mV 0.029 % + 0.66 mV 0.058 % + 6.4 mV 0.057 % + 21 mV	Fluke 743B

Parameter/Equipment	Range	CMC ^{2,4} (\pm)	Comments
DC Voltage ³ – Generate	(0 to 110) mV (0.11 to 1.1) V (1.1 to 11) V	0.011 % + 6.7 μ V 0.011 % + 0.065 mV 0.012 % + 0.87 mV	Fluke 743B
DC Current ³ – Generate	(2 to 22) mA	0.012 % + 3.7 μ A	Fluke 743B
Resistance ³ – Measure	(0 to 110) Ω (0.11 to 1.1) k Ω (1.1 to 11) k Ω	0.058 % + 0.06 Ω 0.058 % + 0.58 Ω 0.12 % + 12 Ω	Fluke 743B
Resistance ^{3,4} – Generate	(0 to 11) Ω (11 to 110) Ω (0.11 to 1.1) k Ω (1.1 to 11) k Ω	0.011 % + 0.023 Ω 0.011 % + 0.047 Ω 0.023 % + 0.59 Ω 0.030 % + 6 Ω	Fluke 743B
Thermocouple Simulation ³ –			
Type J	(-210 to -100) °C (-100 to 800) °C (800 to 1200) °C	0.40 °C 0.30 °C 0.28 °C	Fluke 743B
Type K	(-200 to -100) °C (-100 to 400) °C (400 to 1200) °C (1200 to 1372) °C	0.52 °C 0.41 °C 0.40 °C 0.40 °C	
Type T	(-250 to -200) °C (-200 to 0) °C (0 to 400) °C	1.2 °C 0.51 °C 0.40 °C	
RTD ³ – Simulation			
100 Ω , Pt 3926	(-200 to 0) °C (0 to 630) °C	0.26 °C 0.33 °C	Fluke 743B
100 Ω , Pt 385	(-200 to 0) °C (0 to 400) °C (400 to 800) °C	0.26 °C 0.33 °C 0.52 °C	
100 Ω , Pt 3916	(-200 to -190) °C (-190 to 0) °C (0 to 360) °C	0.42 °C 0.26 °C 0.33 °C	

III. Fluid Quantities

Parameter/Equipment	Range	CMC ^{2, 5, 7} (\pm)	Comments
Gas Flow ³ – Measure	(0.002 to 250) SLPM	1.3 %	Mass flow meter

IV. Mechanical

Parameter/Equipment	Range	CMC ² (\pm)	Comments
Gage Pressure ³ – Pneumatic Gages and Transducers	(0 to 200) psig	0.19 psig	Fluke 743B and Fluke 700PD7

V. Thermodynamics

Parameter/Equipment	Range	CMC ^{2, 5, 7} (\pm)	Comments
Relative Humidity ³ – Measure	(10 to 90) % RH (90.1 to 95) % RH	1.6 % 2.6 %	Vaisala temperature and humidity indicator
Temperature ³ – Measure	(70 to 200) °C (-40 to 125) °C	1.2 °C 0.67 °C	Fluke 743B w/ type T SLE TC Fluke 743B w/ PRT

VI. Time & Frequency

Parameter/Equipment	Range	CMC ^{2, 5} (\pm)	Comments
Timers – Measuring Equipment ³	(1 to 60) min	1.3 seconds	Stopwatch

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ The measurands stated are generated with the Fluke 743B and Fluke 700PXX series of instruments. This capability is suitable for the calibration of the devices intended to measure the stated measurand in the ranges indicated. CMCs are expressed as either a specific value that covers the full range or as a fraction of the reading plus a fixed floor specification.

⁵ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

⁶ This scope meets A2LA's *P112 Flexible Scope Policy*.

⁷ In the statement of CMC, percentages are percentage of reading, unless otherwise indicated.



Accredited Laboratory

A2LA has accredited

AUTO TECHNOLOGY COMPANY

Strongsville, OH

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 29th day of November 2021.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 2563.01
Valid to December 31, 2023

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.