



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994 & ANSI/NCSL Z540.3-2006

TEKTRONIX, INC.
322 Lake Hazeltine Drive
Chaska, MN 55318
Jeff McNamara Phone: 952-368-7894

CALIBRATION

Valid To: August 31, 2023

Certificate Number: 2357.27

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1,5}:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
DC Voltage – Generate ³	(0 to 220) mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	8.3 μV/V + 0.39 μV 5.3 μV/V + 0.62 μV 3.9 μV/V + 2.3 μV 3.9 μV/V + 3.9 μV 5.2 μV/V + 39 μV 6.7 μV/V + 0.39 mV	Fluke 5720A
DC Current – Generate ³	Up to 220 μA 220 mA to 2.2 mA (2.2 to 22) mA (22 to 220) mA 220 mA to 2.2 A (2.2 to 3) A (3 to 11) A (11 to 20.5) A (16.5 to 149.999) A (150 to 1025) A	39 μA/A + 5.4 nA 31 μA/A + 6.2 nA 32 μA/A + 39 nA 39 μA/A + 0.62 μA 70 μA/A + 12 μA 0.30 mA/A + 31 μA 0.40 mA/A + 0.39 mA 0.78 mA/A + 0.58 mA 4.4 mA/A + 0.11 mA 5.1 mA/A + 0.39 mA	Fluke 5720A Fluke 5522A Fluke 5522A w/ 5500A coil

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
DC Resistance – Generate ³	(0 to 10.9999) Ω	38 μΩ/Ω + 0.78 mΩ	Fluke 5522A
	(11 to 32.9999) Ω	27 μΩ/Ω + 1.2 mΩ	
	(33 to 109.9999) Ω	23 μΩ/Ω + 1.1 mΩ	
	(110 to 329.9999) Ω	23 μΩ/Ω + 1.6 mΩ	
	(0.33 to 1.099 999) kΩ	22 μΩ/Ω + 1.6 mΩ	
	(1.1 to 3.2 999 99) kΩ	22 μΩ/Ω + 16 mΩ	
	(3.3 to 10.999 99) kΩ	22 μΩ/Ω + 16 mΩ	
	(11 to 32.999 99) kΩ	24 μΩ/Ω + 0.16 Ω	
	(33 to 109.9999) kΩ	23 μΩ/Ω + 0.16 Ω	
	(110 to 329.9999) kΩ	27 μΩ/Ω + 1.6 Ω	
	(0.33 to 1.09 9999) MΩ	27 μΩ/Ω + 1.6 Ω	
	(1.1 to 3.2 999 99) MΩ	59 μΩ/Ω + 23 Ω	
	(3.3 to 10.999 99) MΩ	0.10 mΩ/Ω + 39 Ω	
	(11 to 32.999 99) MΩ	0.20 mΩ/Ω + 1.9 kΩ	
	(33 to 109.9999) MΩ	0.40 mΩ/Ω + 2.3 kΩ	
	(110 to 329.9999) MΩ	2.4 mΩ/Ω + 78 kΩ	
	(330 to 1100) MΩ	12 mΩ/Ω + 0.39 MΩ	
Fixed Points ³	1 Ω	0.11 mΩ	Fluke 5720A
	1.9 Ω	0.16 mΩ	
	10 Ω	0.21 mΩ	
	19 Ω	0.41 mΩ	
	100 Ω	0.96 mΩ	
	190 Ω	1.8 mΩ	
	1 kΩ	8.6 mΩ	
	1.9 kΩ	16 mΩ	
	10 kΩ	80 mΩ	
	19 kΩ	0.15 Ω	
	100 kΩ	1.2 Ω	
	190 kΩ	2.2 Ω	
	1 MΩ	20 Ω	
	1.9 MΩ	38 Ω	
	10 MΩ	0.37 kΩ	
	19 MΩ	0.84 kΩ	
100 MΩ	12 kΩ		
DC Voltage – Measure ³	Up to 100 mV	8.0 μV/V + 0.23 μV	Keysight 3458A
	(0.1 to 1) V	4.7 μV/V + 2.3 μV	
	(1 to 10) V	4.7 μV/V + 23 μV	
	(10 to 100) V	6.2 μV/V + 0.23 mV	
	(100 to 1000) V	6.4 μV/V + 2.3 mV	

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
DC Current – Measure ³	Up to 100 nA (0.1 to 1) μA (1 to 10) μA (10 to 100) μA (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A	0.24 mA/A + 31 pA 40 μA/A + 31 pA 17 μA/A + 0.078 nA 20 μA/A + 0.62 nA 18 μA/A + 3.9 nA 18 μA/A + 39 nA 30 μA/A + 390 nA 87 μA/A + 7.8 μA	Keysight 3458A
DC Resistance – Measure ³	(0 to 10) Ω (10 to 100) Ω 100 Ω to 1 kΩ (1 to 10) kΩ (10 to 100) kΩ (0.1 to 1) MΩ (1 to 10) MΩ (10 to 100) MΩ (100 to 1200) MΩ	14 μΩ/Ω + 39 μΩ 12 μΩ/Ω + 0.39 mΩ 10 μΩ/Ω + 0.39 mΩ 10 μΩ/Ω + 3.9 mΩ 11 μΩ/Ω + 39 mΩ 16 μΩ/Ω + 1.6 Ω 53 μΩ/Ω + 78 Ω 0.40 mΩ/Ω + 0.78 kΩ 4.8 mΩ/Ω + 7.8 kΩ	Keysight 3458A

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage – Generate ³			
Up to 2.2 mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.55 mV/V + 3.9 μV 0.48 mV/V + 3.9 μV 0.51 mV/V + 3.9 μV 0.5 mV/V + 3.9 μV 2.3 mV/V + 4.7 μV 1.2 mV/V + 9.3 μV 1.7 mV/V + 19 μV 3.3 mV/V + 19 μV	Fluke 5720A
(2.2 to 22) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.25 mV/V + 3.9 μV 0.12 mV/V + 3.9 μV 0.12 mV/V + 3.9 μV 0.22 mV/V + 3.9 μV 0.49 mV/V + 4.7 μV 1.0 mV/V + 9.3 μV 1.4 mV/V + 19 μV 2.7 mV/V + 19 μV	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage – Generate ³ (cont)			
(22 to 220) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.41 mV/V + 12 µV 130 µV/V + 6.2 µV 83 µV/V + 6.2 µV 0.20 mV/V + 6.2 µV 0.47 mV/V + 16 µV 0.86 mV/V + 19 µV 1.3 mV/V + 23 µV 2.6 mV/V + 47 µV	Fluke 5720A
(0.22 to 2.2) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.50 mV/V + 39 µV 86 µV/V + 16 µV 42 µV/V + 7.8 µV 71 µV/V + 9.3 µV 0.13 mV/V + 31 µV 0.39 mV/V + 78 µV 0.93 mV/V + 0.19 mV 1.6 mV/V + 0.31 mV	
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.38 mV/V + 0.39 mV 91 µV/V + 0.16 mV 42 µV/V + 54 µV 71 µV/V + 93 µV 94 µV/V + 0.19 mV 0.25 mV/V + 0.62 mV 0.93 mV/V + 1.9 mV 1.4 mV/V + 3.1 mV	
(22 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.44 mV/V + 3.9 mV 87 µV/V + 1.6 mV 52 µV/V + 0.54 mV 79 µV/V + 0.93 mV 0.14 mV/V + 2.3 mV 0.85 mV/V + 16 mV 4.2 mV/V + 39 mV 7.8 mV/V + 78 mV	
(220 to 1100) V	(15 to 50) Hz 50 Hz to 1 kHz	0.28 mV/V + 16 mV 71 µV/V + 3.1 mV	
(330 to 1020) V	(0.045 to 1) kHz (1 to 5) kHz (5 to 10) kHz	0.23 mV/V + 7.8 mV 0.20 mV/V + 7.8 mV 0.24 mV/V + 7.8 mV	Fluke 5522A

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage – Measure ³			
(0 to 10) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz 100 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz	0.27 μV/V + 2.3 μV 0.18 μV/V + 0.85 μV 0.25 μV/V + 0.85 μV 0.78 μV/V + 0.85 μV 3.9 μV/V + 0.85 μV 31 μV/V + 3.9 μV 54 μV/V + 5.4 μV 0.16 mV/V + 6.2 μV	Keysight 3458A
(10 to 100) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz (8 to 10) MHz	0.13 mV/V + 3.1 μV 71 μV/V + 1.6 μV 0.11 μV/V + 1.6 μV 0.23 μV/V + 1.6 μV 0.62 μV/V + 1.6 μV 2.3 μV/V + 7.8 μV 7.8 μV/V + 7.8 μV 31 μV/V + 54 μV 31 μV/V + 62 μV 0.12 V/V + 78 μV	
(0.1 to 1) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz (8 to 10) MHz	0.28 mV/V + 31 μV 98 μV/V + 16 μV 0.14 mV/V + 16 μV 0.25 mV/V + 16 μV 0.63 mV/V + 16 μV 2.9 mV/V + 78 μV 7.9 mV/V + 78 μV 31 mV/V + 0.54 mV 31 mV/V + 0.62 mV 0.12 V/V + 0.78 mV	
(1 to 10) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz (8 to 10) MHz	0.28 mV/V + 0.31 mV 97 μV/V + 0.16 mV 0.13 mV/V + 0.16 mV 0.25 mV/V + 0.16 mV 0.63 mV/V + 0.16 mV 2.9 mV/V + 0.78 mV 7.9 mV/V + 0.78 mV 31 mV/V + 5.4 mV 31 mV/V + 6.2 mV 0.12 V/V + 7.8 mV	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage ³ – Measure (cont)			
(10 to 100) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	0.31 mV/V + 3.1 mV 0.17 mV/V + 1.6 mV 0.17 mV/V + 1.6 mV 0.28 mV/V + 1.6 mV 0.93 mV/V + 1.6 mV 3.5 mV/V + 7.8 mV 12 mV/V + 7.8 mV	Keysight 3458A
(100 to 700) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.42 mV/V + 31 mV 0.32 mV/V + 16 mV 0.47 mV/V + 16 mV 0.93 mV/V + 16 mV 2.3 mV/V + 16 mV	
AC Current – Generate ³			
Up to 220 µA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.39 mA/A + 16 nA 0.17 mA/A + 10 nA 0.12 mA/A + 8.0 nA 0.28 mA/A + 12 nA 1.0 mA/A + 65 nA	Fluke 5720A
(0.22 to 2.2) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.35 mA/A + 40 nA 0.19 mA/A + 35 nA 0.16 mA/A + 35 nA 0.23 mA/A + 0.11 µA 1.0 mA/A + 0.65 µA	
(2.2 to 22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.34 mA/A + 0.40 µA 0.17 mA/A + 0.35 µA 0.13 mA/A + 0.35 µA 0.20 mA/A + 0.55 µA 1.0 mA/A + 5.0 µA	
(22 to 220) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.35 mA/A + 4.0 µA 0.17 mA/A + 3.5 µA 0.12 mA/A + 2.5 µA 0.20 mA/A + 3.5 µA 1.0 mA/A + 10 µA	
(0.22 to 2.2) A	20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.27 mA/A + 35 µA 0.41 mA/A + 80 µA 6.0 mA/A + 0.16 mA	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Current – Generate ³ (cont)			
(29 to 329.99) µA (0.33 to 3.2999) mA (3.3 to 32.999) mA (33 to 329.99) mA	(10 to 30) kHz (10 to 30) kHz (10 to 30) kHz (10 to 30) kHz	12 mA/A + 0.31 µA 7.8 mA/A + 0.47 µA 3.4 mA/A + 3.1 µA 3.1 mA/A + 0.16 mA	Fluke 5522A
(2.2 to 10.9999) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.48 mA/A + 1.6 mA 0.78 mA/A + 1.6 mA 23 mA/A + 1.6 mA	
(11 to 20.5) A	(45 to 100) Hz 100 Hz to 1 kHz (1 to 5) kHz	0.95 mA/A + 3.9 mA 1.2 mA/A + 3.9 mA 23 mA/A + 3.9 mA	
Clamp-On Meters: Torodial Type Clamps (16.5 to 149.999) A (16.5 to 149.999) A (150 to 1025) A (150 to 1025) A	(45 to 65) Hz (65 to 440) Hz (45 to 65) Hz (65 to 440) Hz	0.38 % 0.84 % 0.37 % 0.83 %	Fluke 5522A w/ 5500 coil
Non-Torodial Type Clamps (16.5 to 149.999) A (16.5 to 149.999) A (150 to 1025) A (150 to 1025) A	(45 to 65) Hz (65 to 440) Hz (45 to 65) Hz (65 to 440) Hz	0.76 % 1.2 % 1.2 % 1.6 %	
AC Current – Measure ³			
100 µA Range	(10 to 20) Hz (20 to 45) Hz 45 Hz to 5 kHz	3.2 mA/A + 23 nA 1.2 mA/A + 23 nA 0.47 mA/A + 23 nA	Keysight 3458A
1 mA Range	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	3.4 mA/A + 0.16 µA 1.3 mA/A + 0.16 µA 0.53 mA/A + 0.16 µA 0.25 mA/A + 0.16 µA	
10 mA Range	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	3.4 mA/A + 1.6 µA 1.3 mA/A + 1.6 µA 0.52 mA/A + 1.6 µA 0.26 mA/A + 1.6 µA	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Current – Measure ³ (cont)			
100 mA Range	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	3.5 mA/A + 16 µA 1.3 mA/A + 16 µA 0.52 mA/A + 16 µA 0.26 mA/A + 16 µA	Keysight 3458A
1 A Range	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz	3.5 mA/A + 0.16 mA 1.3 mA/A + 0.16 mA 0.66 mA/A + 0.16 mA 0.80 mA/A + 0.16 mA	

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical Calibration of Thermocouples ³ – Generate and Measure			
Type B	(600 to 800) °C (800 to 1000) °C (1000 to 1550) °C (1550 to 1820) °C	0.34 °C 0.27 °C 0.24 °C 0.26 °C	Fluke 5522A
Type C	(0 to 150) °C (150 to 650) °C (650 to 1000) °C (1000C to 1800) °C (1800 to 2316) °C	0.24 °C 0.21 °C 0.25 °C 0.39 °C 0.65 °C	

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical Calibration of Thermocouples ³ – Generate and Measure (cont)			
Type E	(-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1000) °C	0.39 °C 0.13 °C 0.11 °C 0.13 °C 0.17 °C	Fluke 5522A
Type J	(-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1200) °C	0.25 °C 0.13 °C 0.12 °C 0.14 °C 0.18 °C	
Type K	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1000) °C (1000 to 1372) °C	0.26 °C 0.14 °C 0.13 °C 0.20 °C 0.31 °C	
Type N	(-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1300) °C	0.31 °C 0.17 °C 0.15 °C 0.14 °C 0.21 °C	
Type R	(0 to 250) °C (250 to 400) °C (400 to 1000) °C (1000 to 1767) °C	0.45 °C 0.28 °C 0.26 °C 0.32 °C	

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical Calibration of Thermocouples ³ – Generate and Measure (cont)			
Type S	(0 to 250) °C (250 to 1000) °C (1000 to 1400) °C (1400 to 1767) °C	0.38 °C 0.28 °C 0.29 °C 0.36 °C	Fluke 5522A
Type T	(-250 to -150) °C (-150 to 0) °C (0 to 120) °C (120 to 400) °C	0.50 °C 0.19 °C 0.13 °C 0.11 °C	
Type U	(-200 to 0) °C (0 to 600) °C	0.44 °C 0.21 °C	
Electrical Calibration of RTD Indicators ³			
Pt 385, 100 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C (630 to 800) °C	0.085 °C 0.12 °C 0.12 °C 0.11 °C 0.097 °C 0.11 °C 0.20 °C	Fluke 5522A
Pt 3926, 100 Ω	(-200 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 400) °C (400 to 630) °C	0.064 °C 0.076 °C 0.075 °C 0.089 °C 0.095 °C 0.17 °C	
Pt 3916, 100 Ω	(-200 to -190) °C (-190 to -80) °C (-80 to 0) °C (0 to 100) °C (100 to 260) °C (260 to 300) °C (300 to 400) °C (400 to 600) °C (600 to 630) °C	0.21 °C 0.060 °C 0.068 °C 0.070 °C 0.077 °C 0.084 °C 0.090 °C 0.13 °C 0.19 °C	

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical Calibration of RTD Indicators ³ (cont)			
Pt 385, 200 Ω	(-200 to -80) °C	0.041 °C	Fluke 5522A
	(-80 to 0) °C	0.043 °C	
	(0 to 100) °C	0.044 °C	
	(100 to 260) °C	0.051 °C	
	(260 to 300) °C	0.098 °C	
	(300 to 400) °C	0.11 °C	
	(400 to 600) °C	0.11 °C	
	(600 to 630) °C	0.13 °C	
Pt 385, 500 Ω	(-200 to -80) °C	0.036 °C	
	(-80 to 0) °C	0.043 °C	
	(0 to 100) °C	0.044 °C	
	(100 to 260) °C	0.051 °C	
	(260 to 300) °C	0.066 °C	
	(300 to 400) °C	0.066 °C	
	(400 to 600) °C	0.073 °C	
	(600 to 630) °C	0.088 °C	
Pt 385, 1000 Ω	(-200 to -80) °C	0.029 °C	
	(-80 to 0) °C	0.029 °C	
	(0 to 100) °C	0.036 °C	
	(100 to 260) °C	0.042 °C	
	(260 to 300) °C	0.050 °C	
	(300 to 400) °C	0.18 °C	
	(400 to 600) °C	0.057 °C	
	(600 to 630) °C	0.18 °C	
PtNi 385, 120 Ω	(-80 to 0) °C	0.081 °C	
	(0 to 100) °C	0.11 °C	
	(100 to 260) °C	0.11 °C	
Cu 427, 10 Ω	(-100 to 260) °C	0.69 °C	

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
Capacitance – Generate ³	(0.19 to 1.1) nF (1.1 to 3.3) nF (3.3 to 11) nF (11 to 110) nF (110 to 330) nF (0.33 to 1.1) μF (1.1 to 3.29) μF (3.3 to 11) μF (11 to 33) μF (33 to 110) μF (110 to 330) μF (0.33 to 1.1) mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF	4.0 mF/F + 7.8 pF 4.0 mF/F + 7.8 pF 2.0 mF/F + 7.8 pF 2.1 mF/F + 78 pF 2.0 mF/F + 0.23 nF 2.1 mF/F + 0.78 nF 2.1 mF/F + 2.3 nF 2.1 mF/F + 7.8 nF 3.2 mF/F + 23 nF 3.6 mF/F + 78 nF 3.5 mF/F + 0.23 μF 3.5 mF/F + 0.78 μF 3.5 mF/F + 2.3 μF 3.5 mF/F + 7.8 μF 5.8 mF/F + 23 μF 8.5 mF/F + 78 μF	Fluke 5522A

¹ This laboratory offers commercial and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMCs are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

⁵ This scope meets A2LA's *P112 Flexible Scope Policy*.

⁶ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



Accredited Laboratory

A2LA has accredited

TEKTRONIX, INC.

Chaska, MN

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and the requirements of ANSI/NCSL Z540.3-2006 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 16th day of August 2021.

A blue ink signature of the Vice President of Accreditation Services, written over a horizontal line.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 2357.27
Valid to August 31, 2023

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.