



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

ACTCO METROLOGY SERVICES  
202 Westview Drive  
Meadville, PA 16335  
Ryan Weingard Phone: 800 382 0393

CALIBRATION

Valid To: February 28, 2022

Certificate Number: 1347.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1, 5</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2, 3</sup> (±)	Comments
Gage Blocks	Up to 4 in	(6 + 4.8L) μin	Standard gage blocks and P&W Labmaster™ laser interferometer system
Pin and Plug Gages	Up to 6 in	(6 + 4.8L) μin	Standard gage blocks and P&W Labmaster™ laser interferometer system
Ring Gages	Up to 4 in	(12 + 5.3L) μin	Standard ring gages and P&W Labmaster™ laser interferometer system

II. Dimensional Testing/Calibration<sup>4</sup>

Parameter/Equipment	Range	CMC <sup>2, 3</sup> (±)	Comments
3-Dimensional Length	Up to 39 in	(210 + 4L) μin	Touch CMM
	Up to 24 in	(380 + 10L) μin	Optical CMM

Parameter/Equipment	Range	CMC <sup>2</sup> ( $\pm$ )	Comments
Dimensional Measurements – Fixtures, Parts	Up to 24 in	0.0043 in	Faro articulating arm CMM

<sup>1</sup> This laboratory offers commercial dimensional testing/calibration service.

<sup>2</sup> Calibration and Measurement Capability uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> In the statement of CMC,  $L$  is the numerical value of the nominal length in inches.

<sup>4</sup> This laboratory meets R205 – *Specific Requirements: Calibration Laboratory Accreditation Program* for the types of dimensional tests listed above and is considered equivalent to that of a calibration.

<sup>5</sup> This scope meets A2LA's P112 *Flexible Scope Policy*.



# Accredited Laboratory

A2LA has accredited

## ACTCO METROLOGY SERVICES

*Meadville, PA*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 21<sup>st</sup> day of January 2020.

A blue ink signature of a person, written over a horizontal line.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1347.01  
Valid to February 28, 2022

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*