



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

INDUSTRIAL SCALES & SYSTEMS, INC.
4295 Cromwell Rd., Suite 615
Chattanooga, TN 37421
Kelly Black Phone: 423 499 2210

CALIBRATION

Valid To: May 31, 2024

Certificate Number: 4017.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 6}:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2, 4} (\pm)	Comments
Calipers ³	Up to 80 in	$(43 + 7.1L) \mu\text{in} + 0.6R$	Gage blocks
Height Gages ³	Up to 48 in	$(19 + 10L) \mu\text{in} + 0.6R$	Gage blocks
Indicators ³	Up to 1 in	$60 \mu\text{in} + 0.6R$	Gage blocks
Micrometers ³	Up to 48 in	$(18 + 3.3L) \mu\text{in} + 0.6R$	Gage blocks
Thickness Gages ³	Up to 1 in	$100 \mu\text{in} + 0.6R$	Gage blocks

II. Mechanical

Parameter/Equipment	Range	CMC ^{2, 4, 5} (\pm)	Comments
Pressure ³ – Measuring Equipment	Up to 300 psi (300 to 3000) psi	$0.24 \text{ psi} + 0.6R$ $0.06 \% + 0.6R$	Crystal engineering pressure module

Parameter/Equipment	Range	CMC ^{2, 4} (\pm)	Comments
Scales and Balances ³	Up to 500 mg	0.021 mg + 0.6R	Class 1 weights
	>500 mg to 10 g	0.044 mg + 0.6R	
	(>10 to 50) g	0.18 mg + 0.6R	
	(>50 to 100) g	0.26 mg + 0.6R	
	(>100 to 300) g	0.45 mg + 0.6R	
	(>300 to 1000) g	0.58 mg + 0.6R	Class 2 weights
	(>1000 to 3000) g	14 mg + 0.6R	
	(>3 to 5) kg	0.22 g + 0.6R	Class 6 and Class F weights
	(>5 to 30) kg	0.7 g + 0.6R	
	(>30 to 50) kg	2.5 g + 0.6R	
	(>50 to 100) kg	3.3 g + 0.6R	
	(>100 to 200) kg	4.1 g + 0.6R	
	(>200 to 250) kg	4.8 g + 0.6R	
	(0 to 200) lb	0.06 lb + 0.6R	
	(>200 to 500) lb	0.14 lb + 0.6R	
	(>500 to 5000) lb	1.2 lb + 0.6R	

¹ This laboratory offers commercial and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ In the statement of CMC, L is the numerical value of the nominal length of the device measured in inches; R represents the resolution of the device, and percentages are to be read as percent of reading, unless otherwise noted.

⁵ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

⁶ This scope meets A2LA's *P112 Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

INDUSTRIAL SCALES & SYSTEMS, INC./ISSI METROLOGY GROUP

Chattanooga, TN

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 23rd day of March 2022.

A blue ink signature of the Vice President, Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 4017.01
Valid to May 31, 2024

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.