

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

FRANK BACON MACHINERY SALES CO. 21251 Ryan Road Warren, MI 48091 Katrina Nelson Phone: 586 756 4280

CALIBRATION

Valid To: June 30, 2025

Certificate Number: 6166.01

Page 1 of 2

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2, 4} (±)	Comments
Linear Displacement – Encoder ³	(0.1 to 48) in	0.016 % IV	ASTM E2309
Linear Displacement – Digital Indicator ³	(0.01 to 1) in	0.037 % IV	ASTM E2309
Extensometer ³ –			ASTM E83-with:
Displacement	(0.002 to 2) in	0.012 % IV	Precision micrometer
	(0.0001 to 4) in	0.010 % IV	Gauge block set
Gauge Length	(0.2 to 6) in	0.037 % IV	Caliper
	(0.1 to 48) in	0.016 % IV	Digital encoder

(A2LA Cert. No. 6166.01) 08/07/2023

5202 Presidents Court, Suite 220 | Frederick, MD 21703-8515 | Phone: 301 644 3248 | Fax: 240 454 9449 | www.A2LA.org

II. Mechanical

Parameter/Equipment	Range	CMC ^{2, 4} (±)	Comments
Force – Force Machines ³ Compression & Tension	(0.02 to 250) lbf (20 to 1000) lbf (250 to 10 000) lbf (1131 to 60 000) lbf (32 564 to 600 000) lbf	0.049 % IV 0.009 % IV 0.025 % IV 0.03 % IV 0.048 % IV	ASTM E4 Class 6 certified weights ASTM E74 Class A Load Cells

III. Time & Frequency

Parameter/Equipment	Range	CMC ^{2, 4} (±)	Comments
Speed ³ – Linear Travel	(0.001 to 30) in/min	0.0091 % IV	ASTM E2658

¹ This laboratory offers commercial calibration service.

- ² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- ³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ In the statement of CMC, IV represents "Indicated Value".

Page 2 of 2

(A2LA Cert. No. 6166.01) 08/07/2023





Accredited Laboratory

A2LA has accredited

FRANK BACON MACHINERY SALES CO.

Warren, MI

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 7th day of August 2023

Mr. Trace McInturff, Vice President, Accreditation Services For the Accreditation Council Certificate Number 6166.01 Valid to June 30, 2025

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.