



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

OMNI-TECH CORPORATION
1460 Torrey Road Suite A
Fenton, MI 48430-3323
Daryl Barlow Phone: 810 750 6474

CALIBRATION

Valid To: February 28, 2026

Certificate Number: 1791.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 5}:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2, 4} (\pm)	Comments
Measuring Machines and Machine Tools ³ –			
Linear Displacement Accuracy	Displacement from (0.1 to 10) m	$(0.13 + 1.3L) \mu\text{m}$	Laser
Temperature Compensated Machine	Up to 1000 mm	$(0.27 + 1.2L) \mu\text{m}$	ISO 10360-2 using gauge block
Non-Temperature Compensated Machine	Up to 1000 mm	$(0.27 + 1.4L) \mu\text{m}$	
Measurement Repeatability	Sphere \varnothing from (12 to 55) mm	0.32 μm	ANSI/ASME B89.4.1 – 1997 Section: 5.3 repeatability and ISO 10360-5 using sphere
Volumetric Performance	Ball bar lengths from (0.1 to 1) m	0.37 μm	Ball bar

¹ This laboratory offers commercial calibration service and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ In the statement of CMC, L is the numerical value of the nominal length of the device measured in inches or meters.

⁵ This scope meets A2LA's *P112 Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

OMNI-TECH CORPORATION

Fenton, MI

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 4th day of January 2024.

A blue ink signature of Mr. Trace McInturff.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 1791.01
Valid to February 28, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.