



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

VERMONT GAGE
(DIVISION OF VERMONT PRECISION TOOLS, INC.)
10 Precision Lane
Swanton, VT 05488
Joe Chevalier Phone: 800 451 4026

CALIBRATION

Valid To: January 31, 2027

Certificate Number: 2514.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations^{1,3}:

I. Dimensional

Parameter/Equipment	Range	CMC ² (±)	Comments
External Diameter – Pins, Plugs, Trilocks, and Taperlocks	Up to 1 in (1 to 2) in (2 to 3) in (3 to 4) in (4 to 5) in (5 to 6) in	10 µin 13 µin 15 µin 18 µin 21 µin 24 µin	Master gage pins with Mahr PLM 600-2
Internal Diameter – Ring Gages	Up to 1 in (1 to 2) in (2 to 3) in (3 to 4) in (4 to 5) in (5 to 6) in	11 µin 12 µin 15 µin 17 µin 20 µin 22 µin	Master rings with Labmaster U306270

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ This scope meets A2LA's *P112 Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

VERMONT GAGE (DIVISION OF VERMONT PRECISION TOOLS, INC.)

Swanton, VT

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 22nd day of November 2024.

A blue ink signature of Mr. Trace McInturff, written over a horizontal line.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 2514.01
Valid to January 31, 2027

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.