



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

JAPAN QUALITY ASSURANCE ORGANIZATION  
 Kyushu Testing Office  
 2-33, Miyanojin 3-Chome, Kurume-shi, Fukuoka, 839-0801 Japan  
 Shinji Wakisaka Phone: 81 942 48 7763

CALIBRATION

Valid To: March 31, 2026

Certificate Number: 1400.05

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,7</sup>:

I. Chemical Quantities

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
pH Indicator <sup>3</sup> –			
Instrumental Error	(0 to 14) pH units	0.0050 pH units	DC voltage
Linearity	(0 to 14) pH units	0.0060 pH units	
Repeatability	(0 to 14) pH units	0.0040 pH units	
pH Detector <sup>3</sup> –	(-414.12 to 414.12) mV		
Instrumental Error	4 pH units & 9 pH units	0.40 mV/pH	Standard solutions DC voltage
Linearity	4 pH units, 7 pH units & 9 pH units	0.90 mV	
Repeatability	4 pH units 7 pH units 9 pH units	0.90 mV 0.80 mV 1.9 mV	
pH Meters <sup>3</sup> –			
Three – Point Calibration Instrumental Error	4 pH units 7 pH units 9 pH units	0.022 pH units 0.023 pH units 0.030 pH units	Standard solutions
Two – Point Calibration Instrumental Error	4 pH units, 9 pH units 7 pH units	0.024 pH units 0.013 pH units	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
pH Meters <sup>3</sup> – (cont)			
One – Point Calibration Instrumental Error	4 pH units, 7 pH units & 9 pH units	0.19 pH units	Standard solutions
Analyzer of Atmospheres Concerned <sup>3</sup>			
Nitrogen Oxide	(25 to 5000) parts in 10 <sup>6</sup>	2.4 % of full scale	Standard gases
Carbon Monoxide	(5 to 50 000) parts in 10 <sup>6</sup>	2.4 % of full scale	
Sulfur Dioxide	(25 to 5000) parts in 10 <sup>6</sup>	2.4 % of full scale	
Oxygen	(5 to 25) vol %	2.4 % of full scale	

## II. Dimensional

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Outside Micrometers <sup>3</sup> – Resolution 0.1µm	Up to 25 mm Up to 50 mm (50 to 200) mm (200 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm (600 to 700) mm (700 to 800) mm (800 to 900) mm (900 to 1000) mm	0.60 µm 2.0 µm 4.0 µm 6.0 µm 8.0 µm 9.0 µm 10 µm 12 µm 13 µm 15 µm 16 µm	Gauge blocks
Calipers <sup>3</sup>	Up to 300 mm (300 to 500) mm (500 to 600) mm (600 to 1000) mm	0.030 mm 0.040 mm 0.050 mm 0.070 mm	Gauge blocks, step gauges

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Height Gauges <sup>3</sup> –  Resolution 0.01 mm  Resolution 0.001 mm  Resolution 0.0001 mm	Up to 200 mm (200 to 500) mm (500 to 600) mm (600 to 1000) mm  Up to 300 mm (300 to 600) mm (600 to 1000) mm  Up to 300 mm (300 to 600) mm (600 to 1000) mm	0.020 mm 0.030 mm 0.040 mm 0.050 mm  0.0090 mm 0.013 mm 0.018 mm  0.0092 mm 0.013 mm 0.018 mm	Gauge blocks, step gauges
Digital Indicators <sup>3</sup>	Up to 100 mm	3.0 µm	Gauge blocks
Depth Micrometer <sup>3</sup>	Up to 50 mm (50 to 150) mm (150 to 300) mm	3.0 µm 4.0 µm 7.0 µm	Gauge blocks, depth micro checker
Dial Gauges <sup>3</sup>	Up to 10 mm (10 to 25) mm (25 to 50) mm (50 to 100) mm	1.8 µm 1.9 µm 2.2 µm 3.2 µm	Gauge blocks, dial gauge tester
Depth Gauges <sup>3</sup>	Up to 300 mm (300 to 600) mm	0.040 mm 0.050 mm	Gauge blocks, step gauges, depth micro-checker
Inside Micrometers <sup>3</sup> –  Calipers	Up to 200 mm (200 to 300) mm (300 to 400) mm (400 to 500) mm	5.0 µm 7.0 µm 8.0 µm 10 µm	Gauge blocks
Inside Micrometers <sup>3</sup> –  Bar Inside	Up to 150 mm (150 to 200) mm (200 to 300) mm (300 to 400) mm (400 to 500) mm (500 to 600) mm (600 to 700) mm (700 to 800) mm (800 to 900) mm (900 to 1000) mm	4.0 µm 5.0 µm 7.0 µm 8.0 µm 10 µm 11 µm 12 µm 13 µm 15 µm 17 µm	Gauge blocks

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Dial Caliper Gauges <sup>3</sup>	Up to 25 mm (25 to 80) mm (80 to 150) mm (150 to 200) mm	0.030 mm 0.040 mm 0.050 mm 0.060 mm	Gauge blocks
Standard Bars	Up to 50 mm (50 to 100) mm (100 to 200) mm (200 to 300) mm (300 to 400) mm (400 to 500) mm	0.80 µm 0.90 µm 1.1 µm 1.5 µm 1.8 µm 2.2 µm	Gauge blocks, electrical – comparators
Thickness Gauges <sup>3</sup> –  Dial Resolution 0.1 mm  Digital Resolution 0.01 mm  Resolution 0.001 mm	Up to 50 mm  Up to 50 mm  Up to 20 mm (20 to 50) mm	0.020 mm  0.020 mm  0.0020 mm 0.0030 mm	Gauge blocks
Cylinder Gauges <sup>3</sup>	Up to 1.6 mm	5.0 µm	Dial gauge tester, dial gauge Note: range is the effective measuring range
Precision Surface Plate <sup>3</sup> (Flatness Only)  Diagonal Length	Up to 430 mm (>430 to 1420) mm (>1420 to 2500) mm	3.5 µm 3.9 µm 4.8 µm	Electronic level
Measuring Microscope <sup>3</sup> –  Run & Parallelism  Straightness  Squareness  Linear Accuracy	Up to 200 mm  Up to 200 mm  Up to 200 mm  Up to 200 mm	7.4 µm  4.6 µm  4.6 µm  7.9 µm	Glass line standard, scales, electrical comparators, square

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Measuring Projector <sup>3</sup> – Linear Accuracy Magnification	Up to 200 mm Up to 600 mm (Screen Size)	4.3 μm 0.05 % of projected reading	Glass line standard, scales
Coordinate Measuring Machines <sup>3</sup>	Up to 1000 mm	(2.0 + L/1200) μm	Step gauges

### III. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
DC Voltage – Generate <sup>3</sup>	(1 to 10) mV (10 to 100) mV (0.10 to 1) V (1 to 10) V (10 to 100) V (100 to 1000) V  (1 to 10) kV	0.051 % + 3.5 μV 0.013 % + 3.5 μV 69 μV/V + 5.8 μV 66 μV/V + 58 μV 74 μV/V + 0.58 mV 74 μV/V + 1.8 mV  0.25 %	Fluke Calibrator      Multimeter with high voltage divider
DC Voltage – Measure <sup>3</sup>	(1 to 10) mV (10 to 100) mV (0.10 to 1) V (1 to 10) V (10 to 100) V (100 to 1000) V  (1 to 10) kV (10 to 50) kV	0.093 % 0.011 % 23 μV/V 19 μV/V 24 μV/V 21 μV/V  0.28 % 0.10 %	Multimeter      High voltage digital meter
DC Current – Generate <sup>3</sup>  Simulated Current	(0.10 to 3) mA (3 to 30) mA (30 to 300) mA (0.3 to 2) A (2 to 10) A (10 to 20) A (20 to 30) A  (20 to 100) A (100 to 500) A (500 to 1000) A	0.021 % + 0.03 μA 0.016 % + 0.06 μA 0.020 % + 0.29 μA 0.020 % + 2.9 μA 0.073 % + 580 μA 0.12 % + 870 μA 0.080 % + 2.1 mA  0.39 % + 110 mA 0.38 % + 110 mA 0.41 % + 110 mA	Calibration of current measuring devices using: Fluke Calibrator      Calibrator with current coil

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
DC Current – Measure <sup>3</sup>	(10 to 100) μA (0.10 to 1) mA (1 to 10) mA (10 to 100) mA (0.10 to 1) A (1 to 2) A (2 to 20) A (20 to 50) A (50 to 100) A	27 μA/A 26 μA/A 28 μA/A 28 μA/A 0.011 % 81 μA/A 0.015 % 0.015 % 0.017 %	Calibration of current generators using multimeter with shunts
DC Resistance – Generate <sup>3</sup>			Calibration of resistance meters using:
Fixed Values	1 mΩ 0.01 Ω 0.1 Ω 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ	80 μΩ/Ω 51 μΩ/Ω 28 μΩ/Ω 18 μΩ/Ω 20 μΩ/Ω 20 μΩ/Ω 16 μΩ/Ω 16 μΩ/Ω 20 μΩ/Ω 23 μΩ/Ω	Standard resistors using with multimeter & generator
Ranges	(0.10 to 1) Ω (1 to 10) Ω (10 to 100) Ω (0.10 to 1) kΩ (1 to 10) kΩ (10 to 100) kΩ (0.10 to 1) MΩ (1 to 100) MΩ (0.10 to 1) GΩ (1 to 10) GΩ (10 to 100) GΩ	0.025 % 72 μΩ/Ω 72 μΩ/Ω 72 μΩ/Ω 72 μΩ/Ω 72 μΩ/Ω 0.023 % 0.027 % 0.24 % 0.32 % 1.1 %	Decade resistors
	(1 to 11) Ω (11 to 33) Ω (0.033 to 33) kΩ (33 to 110) kΩ (110 to 330) kΩ (0.33 to 1.1) MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ	0.035 % 0.018 % 0.015 % 0.017 % 0.018 % 0.035 % 0.035 % 0.076 % 0.12 % 0.58 %	Calibrator (Simulated)

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
DC Resistance – Measure <sup>3</sup>			Calibration of resistors using:
Fixed Values	1 mΩ 0.01 Ω 0.1 Ω 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ 1 MΩ	0.012 % 83 μΩ/Ω 64 μΩ/Ω 18 μΩ/Ω 18 μΩ/Ω 18 μΩ/Ω 18 μΩ/Ω 18 μΩ/Ω 21 μΩ/Ω 21 μΩ/Ω	Multimeter with standard resistors
Ranges	(1 to 10) mΩ (10 to 100) mΩ (0.10 to 1) Ω (1 to 10) Ω (10 to 100) Ω (0.10 to 1) kΩ (1 to 10) kΩ (10 to 100) kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ	0.012 % 90 μΩ/Ω 84 μΩ/Ω 27 μΩ/Ω 27 μΩ/Ω 27 μΩ/Ω 31 μΩ/Ω 48 μΩ/Ω 48 μΩ/Ω 0.035 % 0.13 %	

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Voltage – Generate <sup>3</sup>			Calibration of AC voltmeters using:
(10 to 100) mV (0.10 to 1) V	50 Hz to 1 kHz	0.061 % + 24 μV 0.038 % + 70 μV	Fluke calibrator
(1 to 10) V (10 to 100) V (0.1 to 1) kV	50 Hz to 1 kHz	0.049 % + 0.70 mV 0.060 % + 3.5 mV 0.060 % + 24 mV	
(1 to 10) kV	(50, 60) Hz	0.40 %	Multimeter with high voltage divider
(10 to 100) mV (0.10 to 1) V (1 to 10) V (10 to 100) V (0.10 to 1) kV	(1 to 10) kHz	0.066 % + 24 μV 0.041 % + 70 μV 0.038 % + 0.70 mV 0.094 % + 11 mV 0.11 % + 0.024 V	

Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Voltage – Measure <sup>3</sup>			
(10 to 100) mV (0.1 to 0.3) V (0.3 to 1) V	50 Hz to 1 kHz (except 50, 60, 400 Hz)	0.041 % 0.025 % 0.023 %	Calibration of voltage generators using: multimeter
(1 to 10) V (10 to 100) V (100 to 1000) V	50 Hz to 1 kHz (except 50, 60, 400 Hz)	0.022 % 0.023 % 0.034 %	
(10 to 100) mV (100 to 300) mV	(1 to 10) kHz	0.047 % 0.038 %	
(0.30 to 1) V (1 to 10) V (10 to 100) V (100 to 1000) V	(1 to 10) kHz	0.024 % 0.023 % 0.024 % 0.034 %	
(10 to 100) mV (0.1 to 0.3) V (0.3 to 1) V	(50, 60, 400) Hz	0.042 % 0.025 % 0.023 %	
(1 to 10) V (10 to 100) V (100 to 1000) V	(50, 60, 400) Hz	0.022 % 0.023 % 0.034 %	
(1 to 10) kV (10 to 30) kV	(50, 60) Hz	0.57 % 0.51 %	
AC Current – Generate <sup>3</sup>			
100 µA to 3 mA (3 to 30) mA (30 to 300) mA (0.30 to 2) A (2 to 10) A (10 to 20) A (20 to 50) A	50 Hz to 1 kHz	0.14 % + 0.18 µA 0.084 % + 2.4 µA 0.084 % + 24 µA 0.091 % + 0.16 mA 0.14 % + 2.4 mA 0.19 % + 5.8 mA 0.18 % + 5.8 mA	Calibration of AC current measuring devices using:  Fluke calibrator
(20 to 100) A (100 to 500) A (500 to 1000) A	(50, 60) Hz	0.40 % + 110 mA 0.40 % + 110 mA 0.42 % + 110 mA	Calibrator with current coil (ex. clampmeter)



Parameter/Range	Frequency	CMC <sup>2,4</sup> (±)	Comments
AC Current – Measure <sup>3</sup>  100 μA to 2 mA (2 to 10) mA (10 to 100) mA (0.10 to 1) A (1 to 2) A (2 to 14) A  (14 to 20) A (20 to 50) A (50 to 100) A	50 Hz to 1 kHz       (50, 60) Hz	0.048 % 0.038 % 0.041 % 0.045 % 0.054 % 0.062 %  0.049 % 0.049 % 0.063 %	Calibration of current generators using multimeter with shunts
AC Resistance – Generate <sup>3</sup>  Fixed Points 0.01 Ω 0.1 Ω 0.2 Ω 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ	50 Hz to 1 kHz         1 kHz	0.082 % 0.082 % 0.082 % 0.082 % 0.082 % 0.082 % 0.082 % 0.082 % 0.082 %	Calibration of AC resistance meters using: standard AC resistors
AC Power – Generate <sup>3</sup>  10 W to 6 kW  (6 to 950) kW clamp type	(50, 60) Hz	0.18 %  0.40 %	Calibration of watt meters:  Calibrator  Calibrator & current coil
High Frequency Voltage <sup>3</sup> –  (0.01 to 0.1) V  (0.10 to 1) V  (10 to 120) dB/μV	10 kHz to 1 MHz (1 to 10) MHz  10 kHz to 1 MHz (1 to 10) MHz  (10 to 500) MHz (500 to 1000) MHz	1.8 % 2.6 %  1.8 % 2.0 %  0.25 dB 0.30 dB	Rohde & Schwartz URV5   Measuring receiver

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Calibration of Thermocouple Temperature Indicators <sup>3</sup> –			
Type E	(-250 to -220) °C	0.81 °C	Calibration of temperature meters using: Fluke calibrator
	(-220 to -180) °C	0.70 °C	
	(-180 to -100) °C	0.66 °C	
	(-100 to -70) °C	0.37 °C	
	(-70 to -25) °C	0.37 °C	
	(-25 to 350) °C	0.36 °C	
	(350 to 650) °C	0.37 °C	
Type J	(650 to 1000) °C	0.40 °C	
	(-210 to -170) °C	0.54 °C	
	(-170 to -120) °C	0.49 °C	
	(-120 to -20) °C	0.42 °C	
	(-20 to 150) °C	0.41 °C	
	(150 to 760) °C	0.43 °C	
Type K	(760 to 1200) °C	0.46 °C	
	(-200 to -170) °C	0.68 °C	
	(-170 to -100) °C	0.55 °C	
	(-100 to -25) °C	0.39 °C	
	(-25 to 120) °C	0.37 °C	
Type R	(120 to 1000) °C	0.44 °C	
	(1000 to 1372) °C	0.56 °C	
	(0 to 50) °C	1.2 °C	
	(50 to 250) °C	0.92 °C	
	(250 to 400) °C	0.77 °C	
	(400 to 800) °C	0.70 °C	
Type S	(800 to 1000) °C	0.61 °C	
	(1000 to 1760) °C	0.66 °C	
	(0 to 50) °C	1.2 °C	
	(50 to 250) °C	0.92 °C	
	(250 to 400) °C	0.77 °C	
	(400 to 800) °C	0.70 °C	
Type T	(800 to 1000) °C	0.70 °C	
	(1000 to 1400) °C	0.71 °C	
	(1400 to 1760) °C	0.77 °C	
	(-250 to -230) °C	1.4 °C	
	(-230 to -170) °C	0.92 °C	
	(-170 to -150) °C	0.83 °C	
	(-150 to -90) °C	0.48 °C	
	(-90 to -20) °C	0.42 °C	
Type T	(-20 to 0) °C	0.42 °C	
	(0 to 120) °C	0.37 °C	
	(120 to 400) °C	0.36 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Calibration of RTDs <sup>3</sup> –  Pt385, 100 Ω & JPt-100	(-200 to 850) °C	0.13 °C	Decade resistors

#### IV. Mechanical

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Scales & Balances <sup>3</sup> –	1 mg 2 mg (2 to 5) mg 5 mg (5 to 10) mg 10 mg 20 mg (20 to 50) mg 50 mg (50 to 100) mg 100 mg 200 mg (200 to 500) mg 500 mg (500 to 1000) mg 1 g 2 g (2 to 5) g 5 g (5 to 10) g 10 g 20 g (20 to 50) g 50 g (50 to 100) g 100 g (100 to 200) g 200 g (200 to 20 000) g (20 to 160) kg (160 to 300) kg	3600 µg/g 1800 µg/g 1800 µg/g 710 µg/g 1400 µg/g 470 µg/g 300 µg/g 360 µg/g 150 µg/g 220 µg/g 94 µg/g 59 µg/g 59 µg/g 30 µg/g 45 µg/g 18 µg/g 12 µg/g 12 µg/g 5.9 µg/g 8.7 µg/g 3.6 µg/g 2.4 µg/g 2.8 µg/g 1.2 µg/g 1.8 µg/g 1.0 µg/g 1.4 µg/g 1.0 µg/g 1.1 µg/g 30 µg/g 31 µg/g	Calibrations performed using reference weights

Parameter/Equipment	Frequency	CMC <sup>2</sup> (±)	Comments
Vibration Meter –  Acceleration (1 to 50) m/s <sup>2</sup>  Velocity (1 to 50) mm/s  Displacement (0.1 to 10) mm	(5 to 2000) Hz  (5 to 2000) Hz  (5 to 160) Hz	3.0 % of measurement value  3.0 % of measurement value  3.0 % of measurement value	Secondary vibration calibration system
Vibration Exciter <sup>3</sup> –  Acceleration (1 to 200) m/s <sup>2</sup>  Velocity (0.01 to 1) m/s  Displacement (1 to 10) mm  Frequency	(5 to 10) Hz (10 to 5000) Hz  (5 to 10) Hz (10 to 160) Hz (160 to 1000) Hz (1 to 3) kHz  (5 to 10) Hz (10 to 20) Hz (20 to 40) Hz (40 to 80) Hz (80 to 160) Hz (160 to 315) Hz  (5 to 2000) Hz (2 to 4) kHz (4 to 5) kHz	4.0 % of measurement value 3.0 % of measurement value  6.0 % of measurement value 3.0 % of measurement value 7.0 % of measurement value 8.0 % of measurement value  3.0 % of measurement value 4.0 % of measurement value 4.0 % of measurement value 5.0 % of measurement value 9.0 % of measurement value 16 % of measurement value  0.02 Hz 0.03 Hz 0.04 Hz	Secondary vibration calibration system       Frequency counter
Vibration Accelerometer  Voltage sensitivity 0.02 to 10 <sup>3</sup> V/(m/s <sup>2</sup> )  Charge sensitivity 10 <sup>-4</sup> to 10 <sup>6</sup> pC/(m/s <sup>2</sup> )	(20 to 50) Hz (50 to 80) Hz (80 to 100) Hz (100 to 160) Hz (160 to 1600) Hz (1.6 to 2) kHz (2 to 2.5) kHz (2.5 to 5) kHz  (20 to 100) Hz (100 to 160) Hz (160 to 200) Hz (200 to 315) Hz (315 to 1600) Hz (1.6 to 2) kHz (2 to 2.5) kHz (2.5 to 3.15) kHz (3.15 to 5) kHz	1.8 % 1.7 % 1.5 % 1.4 % 1.5 % 1.6 % 1.9 % 2.5 %  1.4 % 1.3 % 1.4 % 1.5 % 1.4 % 1.6 % 1.9 % 2.4 % 2.5 %	Secondary vibration accelerometer calibration system

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Pressure –			
Hydraulic – Gauge Pressure <sup>3</sup>	Up to 100 MPa	0.0042 MPa or 0.021 % of measurement value, whichever is largest	Dead-weight pressure balance, digital pressure gauge
Air – Gauge Pressure <sup>3</sup>	(-95 to 0) kPa	0.040 kPa	
	(0 to 350) kPa	0.063 kPa or 0.030 % of measurement value, whichever is largest	
	(350 to 2000) kPa	0.20 kPa or 0.040 % of measurement value, whichever is largest	
	(2000 to 7000) kPa	1.7 kPa	
Air – Absolute Pressure	750 hPa to 1150 hPa	0.50 hPa	
Dead-Weight Pressure Balances –			
Pressure Medium – Oil <sup>3</sup>	Up to 100 MPa	0.0054 MPa or 0.018 % of measurement value, whichever is largest	
Pressure Medium – Air <sup>3</sup>	Up to 350 kPa	0.074 kPa or 0.035 % of measurement value, whichever is largest	
	(350 to 2000) kPa	0.40 kPa or 0.040 % of measurement value, whichever is largest	
	(2000 to 7000) kPa	2.0 kPa	
Liquid Column Monometer –			
Liquid – Mercury <sup>3</sup>	Up to 220 kPa	0.15 kPa	
Liquid – Water <sup>3</sup>	Up to 20 kPa	0.060 kPa	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments	
Volumetric Ware –				
One Mark Pipettes	(0.5 to 2) ml (2 to 5) ml (5 to 10) ml (10 to 25) ml (25 to 50) ml (50 to 100) ml	0.010 ml 0.015 ml 0.020 ml 0.030 ml 0.050 ml 0.080 ml	Scales & balances, liquid-in-glass thermometer	
Graduated Pipettes	(0.1 to 1) ml (1 to 2) ml (2 to 5) ml (5 to 10) ml (10 to 25) ml (25 to 50) ml	0.010 ml 0.015 ml 0.030 ml 0.050 ml 0.10 ml 0.20 ml		
Burettes	(1 to 5) ml (5 to 10) ml (10 to 25) ml (25 to 50) ml (50 to 100) ml	0.010 ml 0.020 ml 0.030 ml 0.050 ml 0.10 ml		
One Mark Volumetric Flask	(5 to 10) ml (10 to 25) ml (25 to 50) ml (50 to 100) ml (100 to 250) ml (250 to 500) ml (500 to 1000) ml (1000 to 2000) ml	0.025 ml 0.040 ml 0.040 ml 0.060 ml 0.15 ml 0.25 ml 0.40 ml 0.60 ml		
Graduated Measuring Cylinders	Up to 5 ml (5 to 25) ml (25 to 100) ml (100 to 200) ml (200 to 300) ml (300 to 500) ml (500 to 1000) ml (1000 to 2000) ml	0.10 ml 0.20 ml 0.50 ml 1.0 ml 1.5 ml 2.5 ml 5.0 ml 10 ml		
Torque Tools <sup>3</sup>	(0.2 to 500) N·m	0.80 % of reading		Torque wrench tester, torque tester

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Indirect Verification of Rockwell Hardness Testing Machine <sup>3</sup> –  B Scale  C Scale	Low Medium High  Low Medium High	0.87 HRB 0.79 HRB 0.63 HRB  0.67 HRC 0.54 HRC 0.41 HRC	Indirect calibration method per JIS B 7726 or ASTM E18 (ISO 6508-2);  Standard blocks
Direct Verification of Rockwell Hardness Testing Machine <sup>3</sup> –  Verification of the Test Force  Verification of the Measuring Device	98.07 N Preliminary test force  588.4 N to 1471 N (60 to 150) kgf B scale C scale  (-50 to 200) μm B scale C scale	0.90 N  1.9 N  0.76 μm	Direct calibration method per JIS B 7726 or ASTM E18 (ISO 6508-2);  Force-proving instrument Note: this is only a partial direct calibration  Index master
Indirect Verification of Micro Vickers Hardness Testing Machine <sup>3</sup>	100 HV to 900 HV	5.1 HV	Indirect calibration method per JIS B 7725 or ASTM E92 (ISO6507-2); standard blocks
Direct Verification of Micro Vickers Hardness Testing Machine <sup>3</sup> –  Verification of the Test Force  Verification of the Measuring Device	98.07 mN to 19.61 N 10 gf to 2 kgf  (0 to 1) mm	0.046 N  1.3 μm	Direct calibration method per JIS B 7725 or ASTM E92 (ISO 6507-2); Scale & balances, force-proving instrument  Stage micrometer Note: this is only a partial direct calibration
Indirect Verification of Vickers Hardness Testing Machine <sup>3</sup>	(100 to 900) HV	5.3 HV	Indirect calibration method per JIS B 7725 or ASTM E92 (ISO 6507-2); standard blocks

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Direct Verification of Vickers Hardness Testing Machine <sup>3</sup> –			Direct calibration method per JIS B 7725 or ASTM E92 (ISO 6507-2); Force-proving Instrument
Verification of the Test Force	(19.61 to 490.3) N (2 to 50) kgf	1.0 N	
Verification of the Measuring Device	(0 to 1) mm	1.3 µm	Stage micrometer Note: this is only a partial direct calibration
Indirect Verification of Brinell Hardness Testing Machine <sup>3</sup>	(250 to 450) HBW (100 to 250) HBW (225 to 450) HBW (125 to 225) HBW	8.8 HBW 6.4 HBW 9.4 HBW 6.6 HBW	Indirect calibration method per JIS B 7724 standard blocks  Indirect calibration method per ASTM E10 standard blocks
Direct Verification of Brinell Hardness Testing Machine <sup>3</sup> –			Direct calibration method per JIS B 7724;
Verification of the Test Force	(4.903 to 29.42) kN (500 to 3000) kgf	0.095 kN	Force-proving instrument
Verification of the Measuring Device	(0 to 8) mm	6.6 µm	Stage micrometer
Verification of the Test Force	(4.903 to 29.42) kN (500 to 3000) kgf	0.15 kN	Direct calibration method per ASTM E10
Verification of the Measuring Device	(0 to 8) mm	7.0 µm	Force-proving Instrument
Direct Verification	(225 to 450) HBW (125 to 225) HBW	4.0 HBW 2.3 HBW	Stage micrometer
Uniaxial Testing Machine <sup>3</sup> –			Calibration method per JIS B 7721;
Tension Force	0.1 N to 100 kN	0.30 %	Force-proving instrument, load cell, & dead weight;
Compression Force	0.1 N to 2 MN 2 MN to 5 MN 10 N to 2 MN	0.30 % 0.40 % 0.40 %	Calibration method per ASTM E4;  Force-proving instrument, load cell, & dead weight



V. Thermodynamics

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Digital Thermometer	(-50 to 350) °C	0.080 °C	Digital thermometer, ice point, calibration device
Surface Temperature Measurement Type	(-50 to 500) °C	1.8 °C	
Liquid-in-Glass Thermometer	(-50 to 350) °C	0.090 °C	Digital thermometer, ice point
Mechanical Thermometer	(-50 to 350) °C	0.20 °C	Digital thermometer, ice point
Temperature Installations-Furnaces, Ovens, Incubators, Stirred Water Baths, Fridges & Freezers <sup>3,5</sup>	(-80 to < -50) °C (-50 to 100) °C (>100 to 300) °C (>300 to 950) °C	0.10 °C 0.05 °C 0.11 °C 2.5 °C	PRT (includes uncertainty of transfer from measurand) JTM K09, JTM K07, IEC 60068-3-6, IEC 60068-3-5, T/C's
Humidity Controlled Chamber <sup>3,5</sup>	(10 to 98) % RH	1.6 % RH	Dew point hygrometer & Reference PRT JTM K09, JTM K07, IEC 60068-3-6, IEC 60068-3-5
Hygrometer	30 % RH (>30 to 40) % RH (>40 to 60) % RH (>60 to 70) % RH (>70 to 80) % RH (>80 to 95) % RH	1.1 % RH 1.4 % RH 1.9 % RH 2.2 % RH 2.5 % RH 2.9 % RH	Optical dew point meter

## VI. Time & Frequency

Parameter/Equipment	Range	CMC <sup>2,6</sup> ( $\pm$ )	Comments
Frequency – Measure <sup>3</sup>	1 mHz to 1300 MHz	1.5 parts in $10^7$ Hz/Hz	Frequency counter 53220A, OCXO reference clock, OCXO-MV89A
Frequency – Measuring Equipment <sup>3</sup>	1 mHz to 600 MHz	1.5 parts in $10^7$ Hz/Hz	Frequency counter 53220A, OCXO reference clock, OCXO-MV89A
Stopwatch <sup>3</sup>	(0 to 388.8) s/day	0.02 s/day	Watch tester, OCXO reference clock, OCXO-MV89A

<sup>1</sup> This laboratory offers commercial and field calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC. CMC for the on-site calibrations assume that the environmental conditions are maintained as in the primary laboratory.

<sup>4</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMCs are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

<sup>5</sup> The contributions from the "best existing device" are not included in the CMC claim.

<sup>6</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

<sup>7</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



# Accredited Laboratory

A2LA has accredited

## JAPAN QUALITY ASSURANCE ORGANIZATION KYUSHU TESTING OFFICE

*Fukuoka-ken, JAPAN*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system  
*(refer to joint ISO-ILAC-IAF Communiqué dated April 2017).*



Presented this 20<sup>th</sup> day of March 2024.

A blue ink signature of Trace McInturff, written over a horizontal line.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1400.05  
Valid to March 31, 2026  
Revised August 15, 2024

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*