



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

SCALAR LABORATORIES LTD  
 03 Ha-Metsuda St  
 POB 11425  
 Azor, Israel 5800147  
 Snir Alaluf Phone: 972 3-9442202

CALIBRATION

Valid To: September 30, 2026

Certificate Number: 4968.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations<sup>1, 8:</sup>

I. Chemical

Parameter/Equipment	Range	CMC <sup>2, 9</sup> (±)	Comments <sup>4</sup>
pH <sup>3</sup> – Measuring Equipment	(2, 4, 7, 10, 12) pH	0.04 pH	Standard solution
pH <sup>3</sup> – Electrical Simulation	(0 ≤ pH ≤ 14) pH	0.01 pH	Electrical calibrator
Conductivity <sup>3</sup> – Measuring Equipment	10 μS/cm 84 μS/cm 100 μS/cm 1000 μS/cm 1413 μS/cm 10000 μS/cm 100000 μS/cm	0.27 μS/cm 1.4 μS/cm 1.5 μS/cm 13 μS/cm 18 μS/cm 0.21 mS/cm 1.3 mS/cm	Standard solution
Conductivity <sup>3</sup> – Electrical Simulation	(0.1 ≤ C ≤ 100000) μS/cm	0.3 % + 0.0012 μS/cm	Electrical calibrator
Conductivity <sup>3</sup> – Measure	(0.1 < C ≤ 10000) μS/cm	0.8 % + 0.011 μS/cm	Conductivity meter

Parameter/Equipment	Range	CMC <sup>2,9</sup> (±)	Comments <sup>4</sup>
CO <sub>2</sub> <sup>3</sup> – Measure	(0 < CO <sub>2</sub> ≤ 2.5) % CO <sub>2</sub> (2.5 < CO <sub>2</sub> ≤ 5) % CO <sub>2</sub> (5 < CO <sub>2</sub> ≤ 10) % CO <sub>2</sub>	0.29 % CO <sub>2</sub> 0.26 % CO <sub>2</sub> 0.27 % CO <sub>2</sub>	Gas analyzer
O <sub>2</sub> <sup>3</sup> – Measure	(0 < O <sub>2</sub> ≤ 99) % O <sub>2</sub>	0.4 % O <sub>2</sub>	Gas analyzer
Gas Analyzers, Data Loggers, Transmitters & Detection Equipment <sup>3</sup>	1 % CO <sub>2</sub> 2.5 % CO <sub>2</sub> 5 % CO <sub>2</sub> 10 % CO <sub>2</sub> 20 % CO <sub>2</sub>  1 % O <sub>2</sub> 5 % O <sub>2</sub> 10 % O <sub>2</sub> 20.8 % O <sub>2</sub> 23.5 % O <sub>2</sub>  10 ppm CO 250 ppm CO 500 ppm CO 1000 ppm CO  50 ppm H <sub>2</sub> S 100 ppm H <sub>2</sub> S 250 ppm H <sub>2</sub> S 500 ppm H <sub>2</sub> S  0.5% CH <sub>4</sub> (10% LEL) 2.5% CH <sub>4</sub> (50% LEL) 4.75% CH <sub>4</sub> (95% LEL)	0.10 % CO <sub>2</sub> 0.10 % CO <sub>2</sub> 0.12 % CO <sub>2</sub> 0.16 % CO <sub>2</sub> 0.32 % CO <sub>2</sub>  0.11 % O <sub>2</sub> 0.14 % O <sub>2</sub> 0.18 % O <sub>2</sub> 0.28 % O <sub>2</sub> 0.39 % O <sub>2</sub>  1.0 ppm CO 4.1 ppm CO 7.9 ppm CO 15 ppm CO  1.9 ppm H <sub>2</sub> S 2.3 ppm H <sub>2</sub> S 6.1 ppm H <sub>2</sub> S 8.5 ppm H <sub>2</sub> S  0.04 % CH <sub>4</sub> (0.8 % LEL) 0.05 % CH <sub>4</sub> (1 % LEL) 0.11 % CH <sub>4</sub> (2.2 % LEL)	Reference gases

## II. Dimensional

Parameter/Equipment	Range <sup>7</sup>	CMC <sup>2</sup> (±)	Comments <sup>4</sup>
Calipers <sup>3</sup>	$L \leq 150$ mm $150 < L \leq 300$ mm	11 μm (430 μin) 18 μm (710 μin)	Caliper checker gage blocks: DIN 862

Parameter/Equipment	Range <sup>7</sup>	CMC <sup>2</sup> (±)	Comments <sup>4</sup>
Thickness Gages (Mechanical, Electronical, Voil, Ultrasonic) <sup>3</sup>	$L \leq 50$ mm	2.3 $\mu\text{m}$ (91 $\mu\text{in}$ )	Gage blocks
Micrometers <sup>3</sup>	$L \leq 50$ mm $50 < L \leq 100$ mm $100 < L \leq 200$ mm $200 < L \leq 275$ mm	1.7 $\mu\text{m}$ (67 $\mu\text{in}$ ) 2.4 $\mu\text{m}$ (95 $\mu\text{in}$ ) 4.3 $\mu\text{m}$ (170 $\mu\text{in}$ ) 5.8 $\mu\text{m}$ (228 $\mu\text{in}$ )	Gage blocks: DIN 863

### III. Fluid Quantities

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments <sup>4</sup>
Volume – Fixed Points (Volumetric Apparatus, Pipettes) <sup>3</sup>	1 $\mu\text{l}$ 2 $\mu\text{l}$ 5 $\mu\text{l}$ 10 $\mu\text{l}$ 20 $\mu\text{l}$ 50 $\mu\text{l}$ 100 $\mu\text{l}$ 200 $\mu\text{l}$ 500 $\mu\text{l}$ 1 ml 2 ml 5 ml 10 ml 20 ml 50 ml 100 ml	0.04 $\mu\text{l}$ 0.04 $\mu\text{l}$ 0.04 $\mu\text{l}$ 0.06 $\mu\text{l}$ 0.11 $\mu\text{l}$ 0.20 $\mu\text{l}$ 0.21 $\mu\text{l}$ 0.23 $\mu\text{l}$ 0.44 $\mu\text{l}$ 0.67 $\mu\text{l}$ 1.7 $\mu\text{l}$ 2.7 $\mu\text{l}$ 5.1 $\mu\text{l}$ 8.0 $\mu\text{l}$ 15 $\mu\text{l}$ 17 $\mu\text{l}$	Gravimetric method with analytical balance: ISO 8655-1; ISO 8655-2; ISO 8655-3; ISO 8655-4; ISO 8655-5; ISO 8655-6

#### IV. Industry Specific Calibrations

Parameter/Equipment	Range <sup>7</sup>	CMC <sup>2</sup> ( $\pm$ )	Comments <sup>4</sup>
Sieves – 2D	$L \leq (150 \times 50)$ mm	7.6 $\mu$ m	Profile projector ASTM E11; ISO 3310; ISO 9044; ISO 4782; ISO 4783

#### V. Mechanical

Parameter/Equipment	Range <sup>7</sup>	CMC <sup>2, 9</sup> ( $\pm$ )	Comments <sup>4</sup>
Pressure Gages, Pressure Transducers, Pressure Indicators, Pressure Transmitters <sup>3</sup>			
Absolute	( $0.1 \leq P \leq 50$ ) Pa ( $0.05 < P \leq 2$ ) kPa ( $3.5 \leq P \leq 200$ ) kPa ( $0.2 < P \leq 0.4$ ) MPa ( $0.4 < P \leq 1.2$ ) MPa ( $1.2 < P \leq 2$ ) MPa ( $2 < P \leq 20$ ) MPa ( $20 < P \leq 40$ ) MPa	2 % + 0.02 Pa 1.5 % + 0.5 Pa 55 Pa 0.15 kPa 0.20 kPa 0.25 kPa 25 kPa 29 kPa	Pressure transducer OIML/R101; EA-10/17 Israeli Standard 697
Pneumatic			
Gage & Differential	( $-2400 < P \leq 2400$ ) Pa	0.24 % + 0.15 Pa	
Vacuum	( $-90 \leq P \leq 0$ ) kPa	0.16 kPa	
Hydraulic & Pneumatic	( $0 \leq P \leq 0.1$ ) MPa ( $0.1 < P \leq 0.4$ ) MPa ( $0.4 < P \leq 1.2$ ) MPa ( $1.2 < P \leq 2$ ) MPa ( $2 < P \leq 5$ ) MPa ( $5 < P \leq 7$ ) MPa ( $7 < P \leq 35$ ) MPa ( $35 < P \leq 50$ ) MPa ( $50 < P \leq 70$ ) MPa	0.13 kPa 0.15 kPa 0.20 kPa 0.25 kPa 1.2 kPa 1.6 kPa 12 kPa 18 kPa 25 kPa	

Parameter/Equipment	Range <sup>7</sup>	CMC <sup>2, 9</sup> ( $\pm$ )	Comments <sup>4</sup>
Balances <sup>3, 6</sup> (Includes Analytical Balances)	$1 \text{ mg} < m \leq 500 \text{ g}$ $500 \text{ g} < m \leq 50 \text{ kg}$ $50 \text{ kg} < m \leq 200 \text{ kg}$ $200 \text{ kg} < m \leq 600 \text{ kg}^{10}$ $600 \text{ kg} < m \leq 800 \text{ kg}^{10}$ $800 \text{ kg} < m \leq 1000 \text{ kg}^{10}$	2 LSVD 1 LSVD 2 LSVD 2 LSVD 3 LSVD 4 LSVD	Mass standards: Class E2, F1, M1, OIML R76-1, EURAMET/cg 18/v.02, USP 41
Torque – Torque Wrenches & Torque Drivers	$\text{TRQ} \leq 2.5 \text{ N}\cdot\text{m}$ $2.5 < \text{TRQ} \leq 50 \text{ N}\cdot\text{m}$ $50 < \text{TRQ} \leq 150 \text{ N}\cdot\text{m}$ $150 < \text{TRQ} \leq 500 \text{ N}\cdot\text{m}$ $500 < \text{TRQ} \leq 1000 \text{ N}\cdot\text{m}$	0.012 N·m 0.13 N·m 0.24 N·m 1.4 N·m 3.8 N·m	Torque transducers

## VI. Thermodynamics

Parameter/Equipment	Range <sup>7</sup>	CMC <sup>2</sup> ( $\pm$ )	Comments <sup>4</sup>
Thermocouple Calibration – Type K, T, N, J, R, S	$-196 \text{ }^\circ\text{C}$ $(-100 \leq T \leq 400) \text{ }^\circ\text{C}$ $(400 < T \leq 600) \text{ }^\circ\text{C}$ $(600 < T \leq 800) \text{ }^\circ\text{C}$ $(800 < T \leq 1200) \text{ }^\circ\text{C}$	0.16 $^\circ\text{C}$ 0.16 $^\circ\text{C}$ 0.92 $^\circ\text{C}$ 1.5 $^\circ\text{C}$ 1.8 $^\circ\text{C}$	ASTM E220
RTD (PRT) Probes	$-196 \text{ }^\circ\text{C}$	0.04 $^\circ\text{C}$	ASTM E644
Calibration	$(-100 \leq T < 0) \text{ }^\circ\text{C}$ $(0 \leq T \leq 165) \text{ }^\circ\text{C}$ $(165 < T \leq 400) \text{ }^\circ\text{C}$	0.04 $^\circ\text{C}$ 0.02 $^\circ\text{C}$ 0.04 $^\circ\text{C}$	ASTM E 1137

Parameter/Equipment	Range <sup>7</sup>	CMC <sup>2,9</sup> (±)	Comments <sup>4</sup>
Temperature Measuring  Instruments – Liquid in Glass Thermometers  Mechanical & Electrical Indicators <sup>3</sup>	 (-35 ≤ T ≤ 165) °C (165 < T ≤ 400) °C  -196 °C (-100 ≤ T ≤ 400) °C (400 < T ≤ 600) °C (600 < T ≤ 800) °C (800 < T ≤ 1200) °C	 0.09 °C 0.17 °C  0.13 °C 0.13 °C 1.3 °C 1.5 °C 1.8 °C	Health procedure 126 <sup>12</sup> ; IS 1291 <sup>13</sup>  ASTM E1  ASTM E77; ISO 1770; ISO 1771
Temperature – Measure <sup>3</sup>  Liquid Baths, Heaters	 (-196 ≤ T < -100) °C (-80 ≤ T ≤ 165) °C (165 < T ≤ 400) °C (400 < T ≤ 600) °C (600 < T ≤ 800) °C (800 < T ≤ 1200) °C	 0.04 °C 0.02 °C 0.04 °C 1.2 °C 1.3 °C 1.5 °C	RTD probes, thermocouples
Furnaces, Ovens, Freezers, Autoclaves (Uniformity Surveys) <sup>3</sup>	 (-196 ≤ T < -80) °C (-80 ≤ T ≤ 180) °C (180 < T ≤ 400) °C (400 < T ≤ 600) °C (600 < T ≤ 800) °C (800 < T ≤ 1200) °C	 0.28 °C 0.23 °C 0.28 °C 2.9 °C 3.9 °C 5.8 °C	AMS 2750; ISO 17665-1; ISO 17665-2 DKD-R 5-7 Health procedure 126 <sup>12</sup> ; IS 1291 <sup>13</sup>
Electrical Simulation of RTDs & Thermocouples <sup>3</sup>  RTD  Thermocouples  Type K Type J Type R Type S Type T	 (-200 ≤ T ≤ 800) °C     (-200 ≤ T ≤ 1260) °C (-200 ≤ T ≤ 1000) °C (100 ≤ T ≤ 1580) °C (100 ≤ T ≤ 1580) °C (-200 ≤ T ≤ 400) °C	 0.10 °C    0.14 °C 0.20 °C 0.23 °C 0.22 °C 0.15 °C	EURAMET/cg-11
Relative Humidity – Measure <sup>3</sup> & Measuring Equipment	(10 ≤ H ≤ 95) %RH	1.2 % RH	Humidity chamber & indicator

Parameter/Equipment	Range <sup>7</sup>	CMC <sup>2, 9</sup> (±)	Comments <sup>4</sup>
Relative Humidity – Probe Measuring Equipment	10 % RH 35 % RH 50 % RH 80 % RH 95 % RH	0.59 % RH 0.54 % RH 0.69 % RH 0.78 % RH 0.86 % RH	Humidity standards
Radiation Thermometry	(-35 ≤ T < 0) °C (0 ≤ T ≤ 50) °C (50 < T ≤ 160) °C  (160 < T ≤ 400) °C (400 < T ≤ 600) °C (600 < T ≤ 800) °C (800 < T ≤ 1000) °C	1.1 °C 0.37 °C 1.4 °C  3.6 °C 5.6 °C 7.6 °C 9.7 °C	Blackbody target ε=0.9 to 1.0 λ=(8 to 14) μm ASTM E2847
Dew Point Temperature – Measuring Equipment	(-70 ≤ DP < -40) °C DP (-40 ≤ DP ≤ 20) °C DP	0.38 °C DP 0.31 °C DP	Chilled mirror DP: dew point

## VII. Time & Frequency

Parameter/Equipment	Range <sup>7</sup>	CMC <sup>2</sup> (±)	Comments <sup>4</sup>
Timers	5 sec ≤ t ≤ 9:45 hrs	0.27 sec	Timers
Speed <sup>3</sup>  Rotational (RPM) Oscillation (OPM)	(20 < ω ≤ 90) RPM (90 < ω ≤ 900) RPM (900 < ω ≤ 2000) RPM (2000 < ω ≤ 6000) RPM (6000 < ω ≤ 20 000) RPM (20 000 < ω ≤ 80 000) RPM	0.16 RPM 0.77 RPM 3.4 RPM 3.7 RPM 6.6 RPM 21 RPM	Optical tachometer  OPM is oscillation per minute.
Digital Tachometers	(18 ≤ ω ≤ 600) RPM (600 < ω ≤ 6000) RPM (6000 < ω ≤ 99 999) RPM	0.087 RPM 0.61 RPM 1.4 RPM	Electrical calibrator

<sup>1</sup> This laboratory offers commercial calibration and field calibration services.

- <sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- <sup>3</sup> Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- <sup>4</sup> Calibration can be also performed to manufacturer or specific customer requirements
- <sup>5</sup> LSVD represents the least significant valid displayed division of the device subject to calibration
- <sup>6</sup> Analytical balance is a balance with one LSVD less of 0.001 gram.
- <sup>7</sup> In the statement of the range,  $P$  is pressure,  $m$  is mass,  $T$  is temperature,  $H$  is humidity,  $t$  is time,  $\omega$  is rotation,  $L$  is length,  $C$  is conductivity,  $pH$  is pH.
- <sup>8</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.
- <sup>9</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.
- <sup>10</sup> Calibration by means of substitution method.
- <sup>11</sup> Without testing of flatness of anvils.
- <sup>12</sup> Israel's Ministry of health procedure 126 – "Guidance for the storage and transport of pharmaceutical products"
- <sup>13</sup> IS 1291 Israeli standard (Hebrew version) "Vehicle for food products transport in a controlled temperature."





# Accredited Laboratory

A2LA has accredited

## SCALAR LABORATORIES LTD

*Azor, ISRAEL*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 24<sup>th</sup> day of September 2024.

A blue ink signature of Mr. Trace McInturff, written over a horizontal line.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 4968.01  
Valid to September 30, 2026  
Revised February 5, 2025

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*



*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*