



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSLI Z540-1-1994 & ANSI/NCSLI Z540.3-2006

TEKTRONIX & FLUKE CORPORATION
Sanki Yamato Building B 3rd Floor, 7-10-1 Chuurinkan, Yamato-shi
Kanagawa Pref. 242-007 JAPAN
Hideki Yuyama Phone: +81 46 200 8851

CALIBRATION

Valid To: December 31, 2023

Certificate Number: 2357.02

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted at the location listed above as well as the one satellite location listed below to perform the following calibrations^{1, 5}:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2, 4} (±)	Comments
DC Voltage – Generate	(0 to 220) mV (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1000) V	6.0 μV/V + 0.60 μV 6.0 μV/V + 0.60 μV 4.0 μV/V + 3.0 μV 4.0 μV/V + 2.0 μV 5.0 μV/V + 40 μV 8.0 μV/V + 300 μV	Fluke 57X0A
DC Voltage – Measure	(0 to 0.1) V (0.1 to 1) V (1 to 10) V (10 to 100) V (100 to 1100) V 0 V 0.329 V 1 V 3.29 V 10 V (32.9 to 50) V 329 V 334 V 900 V 1000 V	5.0 μV/V + 0.70 μV 6.0 μV/V + 0.30 μV 4.0 μV/V + 10 μV 7.0 μV/V + 40 μV 20 μV/V 0.31 μV 1.4 μV 3.5 μV 14 μV 14 μV 50 μV 0.73 mV 0.84 mV 0.83 mV 0.86 mV	HP 3458A, OPT 002 HP 3458A, OPT 002 HP 3458A, OPT 002 w/ 752A

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
DC Voltage – Measure (cont)	0.1 V 1 V 10 V 100 V 1000 V	1.2 µV 2.6 µV 16 µV 250 µV 2.6 mV	HP 3458A, OPT 002 w/ 752A, 732B
DC Voltage – Measure ³	0.02 V 0.1 V 0.2 V 1 V 2 V 10 V 20 V 100 V 200 V	37 µV 34 µV 34 µV 44 µV 58 µV 0.3 mV 1.1 mV 3.4 mV 12 mV	Keithley DMM7510
Thermocouple Simulation – Measure 0 °C 100 °C 1000 °C 10 000 °C	0.1 mV 1 mV 10 mV 100 mV	0.43 µV 0.45 µV 0.44 µV 0.72 µV	HP 3458A, OPT 002 w/ 752A, 732B
Thermocouple Simulation – Generate 0 °C 10 000 °C	0 mV 100 mV	0.11 °C 0.29 °C	Fluke 5500A, HP 3458A, OPT 002
DC Current – Generate	(0.2 to 2) pA (2 to 20) pA (20 to 200) pA (0.2 to 2) nA (2 to 20) nA (20 to 200) nA (1 to 220) µA (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A	0.08 % + 0.9 fA 0.1 % + 0.6 fA 0.06 % + 8 fA 0.06 % + 4 fA 0.02 % + 0.8 pA 0.02 % + 0.4 fA 40 µA/A + 6.0 nA 30 µA/A + 6.0 nA 30 µA/A + 40 nA 70 µA/A + 0.30 µA 80 µA/A + 10 µA	Fluke 57X0A Keithley 5156 Fluke 57X0A

Parameter/Equipment	Range	CMC ^{2, 4, 7} (\pm)	Comments
DC Current – Generate (cont)	(2.2 to 11) A (11 to 50) A (110 to 500) A	0.040 % + 0.50 mA 1.7 % 0.85 %	Fluke 57X0A/5725A 5500A/COIL
DC Current – Measure	(0 to 20) pA (20 to 200) pA (0.2 to 2) nA (2 to 20) nA (20 to 200) nA (0 to 100) nA (0.1 to 1) μ A (1 to 10) μ A (10 to 100) μ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A (1 to 2) A (2 to 11) A (0 to 1) pA (5 to 10) pA (50 to 100) pA (0.5 to 1) nA (5 to 10) nA (50 to 100) nA (0.5 to 1) μ A (5 to 10) μ A (50 to 100) μ A (0.5 to 1) mA (5 to 10) mA (50 to 100) mA 190 μ A 329 μ A 1.9 mA 3.29 mA 19 mA 32.9 mA 190 mA 329 mA 1.09 A 2.99 A 10.9 A 20 A	0.96 % + 10 fA 1 % + 4 fA 0.13 % + 2 pA 0.2 % + 0.3 pA 0.2 % + 20 fA 3.3 % + 50 pA 0.33 % + 30 pA 0.030 % + 30 pA 40 μ A/A + 0.8 nA 40 μ A/A + 5 nA 40 μ A/A + 50 nA 50 μ A/A + 0.6 μ A 0.010 % + 10 μ A 0.020 % + 30 μ A 0.040 % + 0.40 mA 0.25 % + 1.0 fA 0.12 % + 0.1 fA 0.03 % + 0.4 fA 0.03 % + 2 fA 0.01 % + 30 fA 0.01 % + 0.4 pA 0.02 % + 5 pA 0.01 % + 50 pA 0.04 % + 8 nA 0.04 % + 50 nA 0.05 % + 0.5 μ A 0.06 % + 5 μ A 3.0 nA 3.6 nA 30 nA 35 nA 0.31 μ A 0.39 μ A 3.1 μ A 4.0 μ A 15 μ A 44 μ A 390 μ A 790 μ A	Keithley 6514 HP 3458A Fluke 8508A Keithley 2002 Keithley 7177 HP 3458A, OPT 002 w/ Fluke 742A-1k w/ Fluke 742A-100 w/ Fluke 742A-10 w/ Fluke 742A-1 w/ Guildline 9230 w/ Fluke Y5020

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
DC Current – Measure (cont)	100 µA 1 mA 10 mA 100 mA 1 A	0.51 nA 6.7 nA 0.1 µA 1.2 µA 12 µA	HP3458A, OPT 002 w/ 742A-1k w/ 742A-100 w/ 742A-10 w/ 742A-1 w/ Guildline 9230
DC Current – Measure ³	0.1 nA 0.5 nA 1 nA 5 nA 10 nA 50 nA 100 nA 0.5 µA 1 µA 5 µA 10 µA 50 µA 100 µA 0.5 mA 1 mA 5 mA 10 mA 50 mA 100 mA 0.5 A 1 A	0.31 pA 0.86 pA 1.3 pA 5.8 pA 8.3 pA 34 pA 68 pA 86 pA 0.12 nA 3.0 nA 2.4 nA 7.2 nA 13 nA 59 nA 0.11 µA 0.58 µA 0.97 µA 5.5 µA 17 µA 190 µA 0.20 mA	Keithley DMM7510 Keithley 5880-SRU
DC Resistance – Generate, Fixed Points	1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	90 µΩ 0.17 mΩ 0.23 mΩ 0.43 mΩ 1.1 mΩ 2.0 mΩ 9.3 mΩ 0.017 Ω 0.093 Ω 0.17 Ω 1.1 Ω 2.1 Ω 19 Ω 38 Ω 0.44 kΩ 0.95 kΩ 12 kΩ	Fluke 57X0A

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
DC Resistance – Generate (cont)	(0 to 10.9999) Ω (11 to 32.9999) Ω (33 to 109.9999) Ω (110 to 329.9999) Ω (0.33 to 32.099 999) kΩ (1.1 to 3.299 999) kΩ (3.3 to 10.999 99) kΩ (11 to 32.999 99) kΩ (33 to 109.9999) kΩ (110 to 329.9999) kΩ (0.33 to 1.099 999) MΩ (1.1 to 3.299 999) MΩ (3.3 to 10.999 99) MΩ (11 to 32.999 99) MΩ (33 to 109.9999) MΩ (110 to 329.9999) MΩ (330 to 1100) MΩ	31 μΩ/Ω + 0.78 mΩ 24 μΩ/Ω + 1.2 mΩ 22 μΩ/Ω + 1.1 mΩ 22 μΩ/Ω + 1.5 mΩ 22 μΩ/Ω + 1.6 mΩ 23 μΩ/Ω + 15 mΩ 23 μΩ/Ω + 16 mΩ 22 μΩ/Ω + 0.16 Ω 23 μΩ/Ω + 0.14 Ω 27 μΩ/Ω + 1.4 Ω 25 μΩ/Ω + 2.0 Ω 50 μΩ/Ω + 21 Ω 0.1 mΩ/Ω + 41 Ω 0.2 mΩ/Ω + 1.9 kΩ 0.4 mΩ/Ω + 2.4 kΩ 2.3 mΩ/Ω + 77 kΩ 12 mΩ/Ω + 0.43 MΩ	Fluke 57X0A
DC Resistance ³ – Measure	(0.1 to 10) Ω (10 to 100) Ω (0.1 to 1) kΩ (1 to 10) kΩ (10 to 100) kΩ (0.1 to 1) MΩ (1 to 10) MΩ (10 to 100) MΩ	20 μΩ/Ω + 0.060 mΩ 10 μΩ/Ω + 0.60 mΩ 10 μΩ/Ω + 0.60 mΩ 10 μΩ/Ω + 5.6 mΩ 10 μΩ/Ω + 0.070 Ω 20 μΩ/Ω + 2.3 Ω 60 μΩ/Ω + 120 Ω 0.060 % + 1.1 kΩ	HP 3458A

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
DC Resistance ³ – Measure (cont)			
Fixed Points	0 Ω	0.53 μΩ	HP 3458A
	2 Ω	0.82 μΩ	
	10.9 Ω	0.61 mΩ	
	11.9 Ω	0.62 mΩ	
	19 Ω	0.86 mΩ	
	30 Ω	0.95 mΩ	
	33 Ω	0.97 mΩ	
	109 Ω	2.1 mΩ	
	119 Ω	5.3 mΩ	
	190 Ω	5.6 mΩ	
	300 Ω	60 mΩ	
	330 Ω	60 mΩ	
	1.09 kΩ	0.015 Ω	
	1.19 kΩ	0.015 Ω	
	1.9 kΩ	0.025 Ω	
	3 kΩ	0.032 Ω	
	3.3 kΩ	0.077 Ω	
	10.9 kΩ	0.15 Ω	
	11.9 kΩ	0.53 Ω	
	19 kΩ	0.55 Ω	
	30 kΩ	0.59 Ω	
	33 kΩ	0.92 Ω	
	109 kΩ	1.5 Ω	
	119 kΩ	1.0 Ω	
	190 kΩ	1.1 Ω	
	300 kΩ	0.93 Ω	
	330 kΩ	0.92 Ω	
	1.09 MΩ	3 Ω	
	1.19 MΩ	14 Ω	
	1.9 MΩ	35 Ω	
	3 MΩ	23 Ω	HP 3458 w/ 742A-1M
	3.3 MΩ	0.15 kΩ	
	10.9 MΩ	0.17 kΩ	
	11.9 MΩ	0.13 kΩ	
	19 MΩ	1.3 kΩ	
	30 MΩ	0.52 kΩ	
	33 MΩ	0.53 kΩ	w/ 742A-10M

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
DC Resistance ³ – Measure (cont)			
Fixed Points	109 MΩ 119 MΩ 290 MΩ 400 MΩ 640 MΩ 1090 MΩ	3.3 kΩ 7.1 kΩ 21 kΩ 0.31 MΩ 1.9 MΩ 0.34 MΩ	HP 3458A w/ 742A-100M
	1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	21 μΩ 18 μΩ 29 μΩ 0.23 mΩ 0.29 mΩ 1.0 mΩ 1.3 mΩ 10 mΩ 11 mΩ 0.19 Ω 0.20 Ω 2.3 Ω 2.8 Ω 25 Ω 800 Ω 1.3 kΩ 9.0 kΩ	HP 8508A w/ 742A

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage – Generate (0.2 to 2.2) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.010 % + 5.0 μV 0.020 % + 4.0 μV 60 μV/V + 4.0 μV 0.020 % + 4.0 μV 0.030 % + 6.0 μV 0.11 % + 10 μV 0.11 % + 20 μV 0.22 % + 20 μV	Fluke 57X0A

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage – Generate (cont)			
(2.2 to 22) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.060 % + 4.0 μV 0.040 % + 4.0 μV 0.030 % + 4.0 μV 0.020 % + 4.0 μV 0.080 % + 5.0 μV 0.14 % + 10 μV 0.14 % + 20 μV 0.27 % + 20 μV	Fluke 57X0A
(22 to 220) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.030 % + 10 μV 0.020 % + 10 μV 90 μV/V + 7.0 μV 0.020 % + 7.0 μV 0.050 % + 20 μV 0.090 % + 20 μV 0.13 % + 30 μV 0.26 % + 50 μV	
220 mV to 2.2 V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.030 % + 40 μV 0.020 % + 40 μV 90 μV/V + 20 μV 70 μV/V + 10 μV 0.010 % + 30 μV 0.040 % + 80 μV 0.10 % + 200 μV 0.15 % + 400 μV	
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (0.5 to 1) MHz	0.030 % + 0.40 mV 0.020 % + 0.40 mV 90 μV/V + 0.20 mV 70 μV/V + 90 μV 0.010 % + 0.20 mV 0.030 % + 0.60 mV 0.090 % + 0.90 mV 0.14 % + 3.1 mV	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage – Generate (cont)			
(22 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz	0.030 % + 3.5 mV 0.020 % + 3.8 mV 90 µV/V + 1.5 mV 80 µV/V + 0.90 mV 0.14 % + 1.6 mV 0.090 % + 15 mV	Fluke 57X0A
(220 to 700) V	(15 to 50) kHz 50 Hz to 1 kHz	90 µV/V + 1.6 mV 80 µV/V	
(700 to 1100) V	40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz	0.010 % + 0.30 mV 0.020 % + 3.3 mV 0.060 % + 10 mV	Fluke 57X0A/5725A
AC Voltage – Measure			
220 µV to 10 mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz	0.030 % + 4.0 µV 0.020 % + 3.0 µV 0.030 % + 3.0 µV 0.10 % + 3.0 µV 0.60 % + 2.0 µV 4.6 % + 3.0 µV	HP 3458A
(10 to 100) mV	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	70 µV/V + 50 µV 0.010 % + 9.0 µV 0.010 % + 9.0 µV 0.030 % + 8.0 µV 0.090 % + 5.0 µV 0.35 % + 10 µV 1.2 % + 30 µV	
(0.1 to 1) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	70 µV/V + 60 µV 70 µV/V + 40 µV 0.020 % + 40 µV 0.030 % + 40 µV 0.090 % + 30 µV 0.35 % + 100 µV 1.2 % + 300 µV	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage – Measure (cont)			
(1 to 10) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (0.3 to 1) MHz	70 µV/V + 0.70 mV 60 µV/V + 0.60 mV 0.010 % + 0.60 mV 0.030 % + 0.50 mV 0.090 % + 0.40 mV 0.35 % + 1.3 mV 1.2 % + 5.6 mV	HP 3458A
(10 to 100) V	(1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.020 % + 5.7 mV 0.020 % + 3.8 mV 0.020 % + 3.8 mV 0.040 % + 3.4 mV 0.14 % + 3.3 mV	Fluke 8508A
(100 to 700) V	40 Hz to 1 kHz	0.050 % + 30 mV	
(700 to 1000) V	40 Hz to 10 kHz (10 to 30) kHz	0.030 % + 24 mV 0.020 % + 57 mV	
3 mV 30 mV	(50 to 60) Hz 10 Hz 50 Hz 60 Hz to 20 kHz 50 kHz 100 kHz 450 kHz	2 µV 9.1 µV 3.7 µV 3.6 µV 6.1 µV 0.11 mV 30 µV	Fluke 5790B
33 mV 300 mV	1 kHz 10 Hz 50 Hz 60 Hz 1 kHz 10 kHz 20 kHz 50 kHz 100 kHz 500 kHz	3.9 µV 68 µV 20 µV 12 µV 18 µV 19 µV 12 µV 18 µV 28 µV 0.11 mV	
0.33 V 3 V	1 kHz 10 Hz 50 Hz 60 Hz 1 kHz 10 kHz	18 µV 0.93 mV 0.14 mV 0.12 mV 0.10 mV 0.19 mV	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage – Measure (cont)			
0.33 V 3 V	20 kHz 50 kHz 100 kHz 450 kHz	85 µV 0.15 mV 0.25 mV 1.2 mV	Fluke 5790B
3.3 V 30 V	1 kHz 10 Hz 50 Hz 60 Hz (1 to 10) kHz 20 kHz 50 kHz 90 kHz	0.35 mV 6.1 mV 1.0 mV 1.1 mV 1.0 mV 1.1 mV 1.8 mV 2.9 mV	
33 V 300 V	1 kHz 50 Hz 60 Hz (1 to 18) kHz 50 kHz 80 kHz	2.5 mV 14 mV 15 mV 14 mV 40 mV 0.15 V	
330 V 1000 V	1 kHz (50 to 60) Hz 1 kHz 5 kHz 8 kHz	16 mV 44 mV 48 mV 43 mV 44 mV	
10 mV	50 Hz 60 Hz 1 kHz 5 kHz	12 µV 2.2 µV 2.9 µV 3.4 µV	
300 mV	10 kHz 30 kHz 10 Hz 45 Hz 50 Hz 60 Hz 1 kHz 5 kHz	2.8 µV 2.6 µV 64 µV 12 µV 15 µV 12 µV 18 µV 12 µV	
3 V	10 kHz 30 kHz 10 Hz 45 Hz	19 µV 18 µV 0.93 mV 0.11 mV	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage – Measure (cont)			
3 V	50 Hz	0.15 mV	Fluke 5790B
	60 Hz	0.14 mV	
	1 kHz	0.11 mV	
	5 kHz	0.10 mV	
	10 kHz	0.20 mV	
	30 kHz	0.16 mV	
	5 V	10 Hz	
45 Hz		0.35 mV	
50 Hz		0.31 mV	
60 Hz		0.28 mV	
1 kHz		0.22 mV	
5 kHz		0.19 mV	
10 kHz		1.1 mV	
2 mV		10 Hz	3.7 μV
	50 Hz	1.8 μV	
	60 Hz	1.7 μV	
	400 Hz to 20 kHz	1.7 μV	
	50 kHz	2.8 μV	
	100 kHz	3.8 μV	
	300 kHz	6.7 μV	
	1 MHz	12 μV	
20 mV	10 Hz	5.8 μV	
	(50 to 60) Hz	2.9 μV	
	400 Hz to 20 kHz	2.8 μV	
	50 kHz	4.9 μV	
	100 kHz	7.0 μV	
	300 kHz	16 μV	
	1 MHz	34 μV	
200 mV	10 Hz	36 μV	
	50 Hz	11 μV	
	60 Hz	8.4 μV	
	400 Hz	7.8 μV	
	1 kHz	7.6 μV	
	20 kHz	7.7 μV	
	50 kHz	13 μV	
	100 kHz	28 μV	
	300 kHz	45 μV	
	1 MHz	0.17 mV	
2 V	10 Hz	1.6 mV	
	50 Hz	84 μV	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage – Measure (cont)			
2 V	60 Hz 400 Hz 1 kHz 20 kHz 50 kHz 100 kHz 300 kHz 1 MHz	46 µV 42 µV 39 µV 40 µV 73 µV 0.12 mV 0.26 mV 1.5 mV	Fluke 5790B
20 V	10 Hz 50 Hz 60 Hz 400 Hz 1 kHz 20 kHz 50 kHz 100 kHz 300 kHz 1 MHz	16 mV 0.48 mV 0.63 mV 0.48 mV 0.47 mV 0.46 mV 0.77 mV 1.3 mV 3.0 mV 19 mV	
200 V	10 Hz 50 Hz 60 Hz 400 Hz (1 to 20) kHz 50 kHz 100 kHz	36 mV 8.1 mV 6.7 mV 5.4 mV 5.3 mV 11 mV 16 mV	
1000 V	50 Hz 60 Hz 400 Hz to 1 kHz	39 mV 57 mV 34 mV	
AC Current – Generate			
(0 to 0.22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.020 % + 20 nA 0.020 % + 20 nA 0.010 % + 20 nA 0.020 % + 20 nA 1.0 % + 60 nA	Fluke 57X0A
(0.22 to 2.2) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz	0.020 % + 40 nA 0.020 % + 40 nA 0.020 % + 30 nA 0.020 % + 100 nA	

Parameter/Range	Frequency	CMC ^{2, 4, 7} (\pm)	Comments
AC Current – Generate (cont)			
(0.22 to 2.2) mA	(5 to 10) kHz	0.10 % + 600 nA	Fluke 57X0A
(2.2 to 22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.020 % + 200 nA 0.020 % + 200 nA 0.020 % + 100 nA 0.020 % + 300 nA 0.090 % + 3.0 μ A	
(22 to 220) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.020 % + 2.0 μ A 0.020 % + 4.0 μ A 0.020 % + 3.0 μ A 0.020 % + 3.0 μ A 0.10 % + 10 μ A	
220 mA to 2.2 A	20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.030 % + 30 μ A 0.050 % + 80 μ A 0.72 % + 0.20 mA	
(2.2 to 11) A	40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.05 % + 0.20 mA 0.10 % + 0.30 mA 0.36 % + 1.3 mA	Fluke 57X0A/5725A
(11 to 500) A	(45 to 65) Hz	1.5 %	Fluke 5500A/COIL
AC Current – Measure			
(10 to 100) μ A	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 1) kHz	0.46 % + 0.040 μ A 0.17 % + 0.040 μ A 0.070 % + 0.040 μ A 0.070 % + 0.040 μ A	HP 3458A
(0.1 to 1) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz	0.46 % + 0.30 μ A 0.17 % + 0.30 μ A 0.060 % + 0.40 μ A 0.030 % + 0.40 μ A	
(1 to 10) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz	0.46 % + 3.0 μ A 0.16 % + 4.0 μ A 0.050 % + 5.0 μ A 0.020 % + 5.0 μ A	
(10 to 100) mA	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz	0.46 % + 30 μ A 0.16 % + 40 μ A 0.060 % + 40 μ A 0.020 % + 40 μ A	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Current – Measure (cont)			
(0.1 to 1) A	(10 to 20) Hz (20 to 45) Hz (45 to 100) Hz (0.1 to 5) kHz	0.45 % + 400 µA 0.15 % + 700 µA 0.060 % + 800 µA 0.090 % + 700 µA	HP 3458A
(1 to 2) A	10 Hz to 2 kHz (2 to 10) kHz	0.060 % + 0.20 mA 0.080 % + 0.20 mA	
(2 to 11) A	10 Hz to 2 kHz (2 to 10) kHz	0.080 % + 2.0 mA 0.26 % + 2.0 mA	Fluke 8508A
33 µA	(1 to 10) kHz	0.036 µA	Fluke 5522A, 5790B w/ 1 kΩ metal film resistor
329 µA	10 Hz 45 Hz 50 Hz 60 Hz to 30 kHz	0.13 µA 0.11 µA 0.16 µA 0.11 µA	
0.33 mA	(1 to 5) kHz	0.13 µA	
3.29 mA	10 Hz 45 Hz 50 Hz 60 Hz to 10 kHz 30 kHz	0.97 µA 0.64 µA 0.65 µA 0.64 µA 0.66 µA	
3.3 mA	(1 to 5) kHz	0.92 µA	
32.9 mA	10 Hz 45 Hz to 10 kHz 30 kHz	9.5 µA 6.2 µA 6.8 µA	
33 mA	1 kHz 5 kHz 30 kHz	9.1 µA 9.2 µA 9.3 µA	
329 mA	10 Hz 45 Hz (50 to 60) Hz (1 to 10) kHz 30 kHz	0.11 mA 0.069 mA 0.070 mA 0.069 mA 0.087 mA	
0.33 A	1 kHz (5 to 10) kHz	0.17 mA 0.16 mA	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Current – Measure (cont)			
2.99 A	50 Hz to 5 kHz	0.0017 A	Fluke 5790B
3.3 A	500 Hz 5 kHz	0.0031 A 0.0032 A	
10 A	(50 to 60) Hz 1 kHz 5 kHz	0.0080 A 0.0083 A 0.011 A	
20 A	50 Hz to 1 kHz 5 kHz	0.030 A 0.034 A	
190 µA	50 Hz 60 Hz 400 Hz to 10 kHz	0.026 µA 0.027 µA 0.026 µA	
1.9 mA	50 Hz 60 Hz 400 Hz to 10 kHz	0.24 µA 0.24 µA 0.21 µA	
19 mA	(50 to 60) Hz 400 Hz to 5 kHz 10 kHz	2.4 µA 2.2 µA 2.1 µA	
190 mA	50 Hz to 10 kHz	23 µA	
1.9 A	(50 to 60) Hz 400 Hz to 1 kHz 5 kHz 10 kHz	0.3 mA 0.29 mA 0.36 mA 1 mA	
AC Voltage Phase – Measure			
3 V/ 0 deg	65 Hz to 1 kHz 5 kHz 10 kHz 30 kHz	0.022 deg 0.048 deg 0.056 deg 0.063 deg	Clarke-Hess 6000A
3 V/ 60 deg	60 Hz 400 Hz 1 kHz 5 kHz 10 kHz 30 kHz	0.026 deg 0.027 deg 0.025 deg 0.048 deg 0.058 deg 0.063 deg	

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Voltage Phase – Measure (cont) 3 V/ 90 deg 30 V/ 90 deg 50 V/ 90 deg	(60 to 65) Hz 400 Hz 1 kHz 5 kHz 10 kHz 30 kHz 65 Hz 65 Hz	0.022 deg 0.027 deg 0.027 deg 0.048 deg 0.058 deg 0.067 deg 0.022 deg 0.035 deg	Clarke-Hess 6000A
AC Current Phase – Measure 300 mA/ 0 deg 2 A/ 0 deg 5 A/ 0 deg 300 mA/ 60 deg 2 A/ 60 deg 20 A/ 60 deg 20 A/ 60 deg 300 mA/ 90 deg 2 A/ 90 deg 20 A/ 90 deg 20 A/ 90 deg	65 Hz 1 kHz 30 kHz 65 Hz 65 Hz 400 Hz 65 Hz	0.025 deg 0.027 deg 0.066 deg 0.024 deg 0.030 deg 0.030 deg 0.027 deg 0.026 deg 0.025 deg 0.032 deg 0.027 deg 0.025 deg 0.034 deg 0.039 deg	For 300 mA and 2A; A40, 6000 A For 5 A and 20 A; Y5020, 6000 A
AC Capacitance – Measure 0.22 nF 0.35 nF 0.48 nF 0.6 nF 1 nF 2 nF 7 nF 10.9 nF 20 nF	5 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz	0.46 pF 0.39 pF 0.46 pF 0.48 pF 0.63 pF 1.2 pF 3.8 pF 6.6 pF 11 pF	PM 6304

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
AC Capacitance – Measure (cont) 70 nF 109 nF 200 nF 300 nF 0.7 μF 1.09 μF 2 μF 3 μF 7 μF 10.9 μF 20 μF 30 μF 70 μF 109 μF	1 kHz 1 kHz 1 kHz 1 kHz 100 Hz 100 Hz 100 Hz 100 Hz 100 Hz 100 Hz 100 Hz 100 Hz 100 Hz 100 Hz 50 Hz 50 Hz	39 pF 66 pF 0.11 nF 0.17 nF 0.40 nF 0.78 nF 1.2 nF 1.7 nF 3.9 nF 6.9 nF 12 nF 16 nF 41 nF 68 nF	PM 6304
DC Capacitance – Measure 200 μF 300 μF 1 mF 3 mF 10 mF 30 mF 50 mF 100 mF		0.098 μF 0.11 μF 0.24 μF 0.45 μF 1.1 μF 3.4 μF 5 μF 9 μF	HP 3458A, 5522A
4 Terminal Parallel Capacitance – Generate ³ 1 pF 10 pF 100 pF	1 MHz 100 kHz 1 MHz 10 kHz 100 kHz 1 MHz	12 fF 0.014 pF 0.012 pF 0.18 pF 0.038 pF 0.042 pF	5880-SRU set of standard capacitors from 1 pF to 1000 pF

Parameter/Range	Frequency	CMC ^{2,4} (±)	Comments
4 Terminal Parallel Capacitance – Generate ³ (cont)			
1000 pF	10 kHz 100 kHz 1 MHz	0.88 pF 0.25 pF 0.82 pF	5880-SRU set of standard capacitors from 1 pF to 1000 pF

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
Oscilloscopes –			
Short Circuit Output	0 V	15 μV	Wavetek 9500, Fluke 9500B
Amplitude – DC Volts, 50 Ω and 1 MΩ Load	(1 to 100) mV 100 mV to 1.0 V (1.0 to 5.56) V	0.050 % + 26 μV 0.022 % + 65 μV 0.026 % + 50 μV	
1 MΩ Load	(5 to 200) V	0.030 %	Wavetek 9500, Fluke 9500B
Amplitude - Sinewave, V _{p-p} ; 50 Ω Load	4.4 mV to 5.6 V 1 Hz to 550 MHz 4.4 mV to 3.4 V 550 MHz to 2.5 GHz 4.4 mV to 2.2 V (2.5 to 3.2) GHz	0.033 V/V 0.063 V/V 0.11 V/V	Wavetek/Fluke 9530
AC Voltage – Measure 50 Ω, V _{RMS}			
20 MHz	(0 to 18) mV _{RMS}	0.048 μV/V _{RMS} + 0.14 μV _{RMS}	
(175 to 350) MHz	(0 to 22) mV _{RMS}	0.016 μV/V _{RMS} + 0.12 μV _{RMS}	
(1.5 to 2) GHz	(0.15 to 40) mV _{RMS}	0.0016 μV/V _{RMS} + 0.035 μV _{RMS}	
(2 to 4) GHz	(0.09 to 24) mV _{RMS}	0.02 μV/V _{RMS} + 0.035 μV _{RMS}	Anritsu MG369x/B/C
(5 to 8) GHz	(0.10 to 42) mV _{RMS}	0.011 μV/V _{RMS} + 0.47 μV _{RMS}	

Parameter/Equipment	Range	CMC ² (±)	Comments
Oscilloscopes – (cont)			
1 M Ω , V _{RMS} 20 MHz (250 to 500) MHz	(0.04 to 18) mV _{RMS} (0.1 to 29) mV _{RMS}	0.038 μ V/V _{RMS} + 0.36 μ V _{RMS} 0.0014 μ V/V _{RMS} + 0.67 μ V _{RMS}	Wavetek 9500, Fluke 9500B Wavetek/Fluke 9530
Leveled Sinewave Flatness 50 Ω Load 50 kHz to 10 MHz Reference, V _{p-p}			Wavetek 9500, Fluke 9500B
1 Hz to 100 MHz	4.4 mV to 5.6 V	0.22 dB	Synthesized signal generator w/ power sensor and power meter
(100 to 550) MHz	4.4 mV to 5.6 V	0.27 dB	
550 MHz to 1.1 GHz	4.4 mV to 3.4 V	0.37 dB	
(1.1 to 2.5) GHz	4.4 mV to 3.4 V	0.47 dB	
(2.5 to 3.2) GHz	4.4 mV to 2.2 V	0.48 dB	
(0.05 to 6.0) GHz	4.4 mV to 2.2 V	0.71 dB	
(6.0 to 8.0) GHz	Up to 300 mV	0.30 dB	
(6.0 to 8.0) GHz	> 300 mV	0.31 dB	
(8.0 to 12.5) GHz	Up to 300 mV	0.37 dB	
(8.0 to 12.5) GHz	> 300 mV	0.38 dB	
(12.5 to 16) GHz	Up to 300 mV	0.46 dB	
(12.5 to 16) GHz	> 300 mV	0.42 dB	

Parameter/Equipment	Range	CMC ^{2, 4, 7} (\pm)	Comments
Oscilloscopes – (cont)			
Leveled Sinewave Flatness 50 Ω Load 50 kHz to 10 MHz Reference, V_{p-p}			Wavetek 9500, Fluke 9500B
(16 to 18) GHz	Up to 300 mV	0.52 dB	Synthesized signal generator w/ power sensor and power meter
(16 to 18) GHz	> 300 mV	0.42 dB	
(18 to 20) GHz	Up to 300 mV	0.56 dB	
(18 to 20) GHz	> 300 mV	0.40 dB	
(20 to 23) GHz	Up to 300 mV	0.55 dB	
(20 to 23) GHz	> 300 mV	0.58 dB	
(23 to 25) GHz	Up to 300 mV	0.57 dB	
(23 to 25) GHz	> 300 mV	0.61 dB	
(25 to 33) GHz	Up to 120 mV	0.87 dB	
(25 to 33) GHz	> 120 mV	0.84 dB	
Time Marker ³	1 ns to 100 ms	0.20 μ s/s + 3 ps	
Rise Time – Measuring Equipment	(14 \pm 3) ps	4.7 ps	
Rise Time – Measure	17.5 ps to 10 ns	5.6 ps	TDS 8200/80E04
Input Resistance – Measure	50 Ω 75 Ω 1 M Ω	0.13 % 0.13 Ω 1.2 k Ω	Fluke 9500B/9530
Pulse Characterization – Measure, Fixed Points ³			
Overshoot	0.5 V 2 V 5 V 20 V	0.50 % 0.53 % 0.46 % 0.45 %	Oscilloscope Tektronix MSO44
Preshoot	5 V 20 V	0.46 % 0.48 %	

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
Pulse Characterization – Measure, Fixed Points ³ (cont)			
Pulse Width	500 ns	0.11 ns	Oscilloscope Tektronix MSO44
RMS Jitter	0 s	0.017 ns	
Slew Rate	0.5 V/ns 0.2 V/ns	0.0061 V/ns 0.0025 V/ns	

II. Fluid Quantities

Parameter/Equipment	Frequency	CMC ^{2,6,7} (±)	Comments
Gas Flow – Measuring Equipment	(1 to 10) SCCM (20 to 200) SCCM (0.1 to 1) SLM (0.5 to 5) SLM	0.24 % 0.37 % 0.24 % 0.33 %	MOLBLOC 1E1 MOLBLOC 2E2 MOLBLOC 1E3 MOLBLOC 5E3 w/ MOLBOX 1

III. Mechanical

Parameter/Equipment	Frequency	CMC ^{2,6,7} (±)	Comments
Torque – Measuring Equipment	(3.5 to 35) N·cm (35 to 141) N·cm (1.41 to 5.65) N·m (5.65 to 45.2) N·m (45.2 to 113) N·m (113 to 339) N·m (339 to 813.6) N·m	0.32 % 0.34 % 0.34 % 0.32 % 0.35 % 0.30 % 0.33 %	Snap-on TTC4 Snap-on TTC5 CDI 2000-400-02 (8400-A) (8400-B) (8400-C) (8400-D) CDI 2000-12-02 w/ 2000-810-01

Parameter/Equipment	Frequency	CMC ^{2,6} (±)	Comments
Gauge Pressure – Measuring Equipment (Pneumatic)	(-90 to 0) kPag (0 to 100) kPag (100 to 200) kPag (200 to 700) kPag (700 to 2000) kPag	0.041 kPa 0.024 kPa 0.033 kPa 0.14 kPa 0.26 kPa	Fluke PM500 w/ BG100K w/ BG200K w/ BG700K w/ BG2M; 6270A

IV. Time & Frequency

Parameter/Equipment	Frequency	CMC ^{2,6} (±)	Comments
Frequency – Measuring Equipment	12 kHz to 3.2 GHz	0.27 µHz/Hz	Wavetek 9500 option 100, Fluke 9500B
Measure	10 MHz	0.33 mHz	Keysight 53181A- 010 w/ GPS
Measure	119 Hz 120 Hz 100 kHz	1.2 µHz 9.3 µHz 0.58 mHz	FCA3103 w/ GPS
	10 Hz 15 Hz 100 Hz 0.2 kHz 0.5 kHz 1 kHz 5 kHz 10 kHz 0.14 MHz 0.2 MHz 0.5 MHz 1 MHz	0.000 12 Hz 0.000 11 Hz 0.000 13 Hz 0.000 10 Hz 0.000 17 Hz 0.000 67 Hz 0.0011 Hz 0.0021 Hz 0.025 Hz 0.036 Hz 0.12 Hz 0.27 Hz	FCA3103 w/ GPS

Satellite Facility

TEKTRONIX Korea, Ltd.
 7th Floor, 517 Yeongdong-daero, Gangnam-gu
 Seoul, 06164, Republic of Korea
 Scott Her Phone: +82 10 3896 1455

CALIBRATION

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
DC Voltage – Measure ³	0 V (100 to 200) mV (1 to 2) V (10 to 20) V (100 to 200) V	12 µV 5 µV/V + 10 µV 20 µV/V + 8 µV 10 µV/V + 0.2 mV 30 µV/V + 1.3 mV	Keithley 2002
DC Current – Measure ³	(0 to 1) pA (5 to 10) pA (50 to 100) pA (0.5 to 1) nA (5 to 10) nA (50 to 100) nA (0.5 to 1) µA (5 to 10) µA (50 to 100) µA (0.5 to 1) mA (5 to 10) mA (50 to 100) mA	0.25 % + 1.0 fA 0.12 % + 0.1 fA 0.03 % + 0.4 fA 0.03 % + 2 fA 0.01 % + 30 fA 0.01 % + 0.4 pA 0.02 % + 5 pA 0.01 % + 50 pA 0.04 % + 8 nA 0.04 % + 50 nA 0.05 % + 0.5 µA 0.06 % + 5 µA	Keithley 2002 Keithley 7177



Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
Oscilloscopes –			
Short Circuit Output	0 V	15 µV	Wavetek 9500, Fluke 9500B
Amplitude – DC Volts, 50 Ω and 1 MΩ Load	(1 to 100) mV 100 mV to 1.0 V (1.0 to 5.56) V	0.050 % + 26 µV 0.022 % + 65 µV 0.026 % + 50 µV	
1 MΩ Load	(5.6 to 222.4) V	0.030 %	
Amplitude - Sinewave, V _{p-p} 50 Ω Load	4.4 mV to 5.6 V 1 Hz to 550 MHz 4.4 mV to 3.4 V 550 MHz to 2.5 GHz 4.4 mV to 2.2 V (2.5 to 3.2) GHz	0.033 V/V 0.063 V/V 0.11 V/V	Wavetek 9500, Fluke 9500B Wavetek/Fluke 9530
AC Voltage – Measure 50 Ω, V _{RMS} 20 MHz (175 to 350) MHz (1.5 to 2) GHz	(0 to 18) mV _{RMS} (0 to 22) mV _{RMS} (0.15 to 40) mV _{RMS}	0.048 µV/V _{RMS} + 0.14 µV _{RMS} 0.016 µV/V _{RMS} + 0.12 µV _{RMS} 0.0016 µV/V _{RMS} + 0.035 µV _{RMS}	Wavetek 9500, Fluke 9500B Wavetek/Fluke 9530
(2 to 4) GHz (5 to 8) GHz	(0.09 to 24) mV _{RMS} (0.10 to 42) mV _{RMS}	0.02 µV/V _{RMS} + 0.035 µV _{RMS} 0.011 µV/V _{RMS} + 0.47 µV _{RMS}	Anritsu MG369x/B/C
1 MΩ, V _{RMS} 20 MHz (250 to 500) MHz	(0.04 to 18) mV _{RMS} (0.1 to 29) mV _{RMS}	0.038 µV/V _{RMS} + 0.36 µV _{RMS} 0.0014 µV/V _{RMS} + 0.67 µV _{RMS}	Wavetek 9500, Fluke 9500B Wavetek/Fluke 9530

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
Oscilloscopes – (cont)			
Leveled Sinewave Flatness, 50 Ω Load 50 kHz, 10 MHz Reference, V _{p-p}			
1 Hz to 100 MHz	4.4 mV to 5.6 V	0.22 dB	Synthesized signal generator w/ power sensor and power meter
(100 to 550) MHz	4.4 mV to 5.6 V	0.27 dB	
550 MHz to 1.1 GHz	4.4 mV to 5.6 V	0.37 dB	
(1.1 to 2.5) GHz	4.4 mV to 3.4 V	0.47 dB	
(2.5 to 3.2) GHz	4.4 mV to 3.4 V	0.48 dB	
(0.05 to 6.0) GHz	4.4 mV to 2.2 V	0.71 dB	
(6.0 to 8.0) GHz	Up to 300 mV	0.30 dB	
(6.0 to 8.0) GHz	> 300 mV	0.31 dB	
(8.0 to 12.5) GHz	Up to 300 mV	0.37 dB	
(8.0 to 12.5) GHz	> 300 mV	0.38 dB	
(12.5 to 16) GHz	Up to 300 mV	0.46 dB	
(12.5 to 16) GHz	> 300 mV	0.42 dB	
(16 to 18) GHz	Up to 300 mV	0.52 dB	
(16 to 18) GHz	> 300 mV	0.42 dB	
(18 to 20) GHz	Up to 300 mV	0.56 dB	
(18 to 20) GHz	> 300 mV	0.40 dB	
(20 to 23) GHz	Up to 300 mV	0.55 dB	
(20 to 23) GHz	> 300 mV	0.58 dB	
(23 to 25) GHz	Up to 300 mV	0.57 dB	
(23 to 25) GHz	> 300 mV	0.61 dB	
(25 to 33) GHz	Up to 120 mV	0.87 dB	
(25 to 33) GHz	> 120 mV	0.84 dB	
Time Marker ³	1 ns to 100 ms	0.20 μs/s + 3 ps	Fluke 9500B/9530

Parameter/Equipment	Range	CMC ^{2, 4, 7} (±)	Comments
Input Resistance – Measure	50 Ω 75 Ω 1 MΩ	0.13 % 0.13 Ω 1.2 kΩ	Fluke 9500B/9530

II. Time & Frequency

Parameter/Equipment	Range	CMC ^{2, 6} (±)	Comments
Frequency – Measuring Equipment	12 kHz to 3.2 GHz	0.27 μHz/Hz	Wavetek 9500 option 100, Fluke 9500B

¹ This laboratory offers commercial calibration service at all laboratory locations listed in this scope of accreditation.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC uncertainty due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

⁵ This scope meets A2LA's *P112 Flexible Scope Policy*.

⁶ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

⁷ In the statement of CMC, percent is defined as percent of reading.



Accredited Laboratory

A2LA has accredited

TEKTRONIX & FLUKE CORPORATION

Kanagawa Pref. Japan

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994, the requirements of ANSI/NCSL Z540.3-2006, and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 2nd day of September 2021.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 2357.02
Valid to December 31, 2023

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.