



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

NATIONAL CALIBRATION INC.  
 1395 N Hayden Rd  
 Scottsdale, AZ 85257  
 Bruce Demaree Phone: 602 437 0114

CALIBRATION

Valid To: August 31, 2026

Certificate Number: 1892.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations<sup>1, 8</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
Calipers <sup>3</sup> – OD Resolution 0.0005”	Up to 8 in (8 to 60) in (60 to 80) in	(290 + 0.48L) µin (290 + 0.83L) µin (6.1L) µin	Gage blocks, gage block accessories
ID Resolution 0.0005”	Up to 2 in	0.0003 in	
Step Resolution 0.0005”	Up to 2 in	0.0003 in	
Depth Resolution 0.0005”	Up to 2 in	0.0003 in	
Gage Blocks	Up to 2 in (2 to 4) in (4 to 6) in (6 to 8) in (8 to 12) in	(3.4 + 1.2L) µin (4.0 + 1.4L) µin (3.0 + 1.9L) µin (2.7 + 2.0L) µin (2.5 + 2.1L) µin	LabMaster™, standard gage blocks
Height Gage <sup>3</sup>	Up to 40 in	(110 + 1.6L) µin	Gage blocks

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Height Master	Up to 36 in	(76 + 1.7L) μin	Gage blocks
Indicators <sup>3</sup>	Up to 6 in	(29 + 0.8L) μin	Labmaster™, Gage blocks, Supermicrometer™
Length Standards	Up to 12 in	(4 + 1.9L) μin	Labmaster™, gage blocks,
	(12 to 40) in	(23 + 2.8L) μin	Electronic height gage, surface plate, gage blocks (standardization)
Micrometers <sup>3</sup> –			
ID Mics	Up to 8 in (8 to 16) in (16 to 24) in (24 to 32) in (32 to 40) in	(58 + 0.72L) μin (49 + 2.3L) μin (48 + 2.3L) μin (43 + 2.6L) μin (32 + 2.9L) μin	Gage blocks
OD Mics	Up to 8 in (8 to 16) in (16 to 24) in (24 to 32) in (32 to 40) in (40 to 48) in (48 to 56) in	(28 + 1.3L) μin (19 + 3.1L) μin (24 + 2.8L) μin (22 + 2.9L) μin (14 + 3.1L) μin (33 + 2.6L) μin (20 + 2.9L) μin	Gage blocks
Depth Mics	Up to 6 in (6 to 12) in	(29 + 1.1L) μin (8.8 + 4.4L) μin	Gage blocks, triangle base
Tri-Mics	(0.01 to 0.018) in (0.15 to 6) in	(37 + 0.025D) μin (120 + 4.0D) μin	Ring gages
V-Anvil Mics	Up to 2 in	(61 + 7.6L) μin	Plug gages
Pin Gauges/Plug Gages	(0.011 to 0.060) in (0.061 to 0.250) in (0.251 to 0.500) in (0.501 to 0.625) in (0.626 to 1.000) in (1.000 to 4.000) in	(8.1 + 32D) μin (8.7 + 1.4D) μin (7.6 + 5.6D) μin (8.6 + 3.7D) μin (4.1 + 8.1D) μin (8.6 + 6.4D) μin	LabMaster™

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
Digital Protractors	Up to 45°	0.005°	Sine Plate & Gage Blocks
Plain Ring Gage	(0.04 to 0.125) in (0.125 to 0.25) in (0.25 to 1) in (1 to 10) in	(6.6 + 14.7D) μin (8.4 + 24.2D) μin (17 + 9.5D) μin (2 + 8.4D) μin	LabMaster™
Ruler Calibrator	Up to 40 in	0.0005 in	Gage blocks
Rules	Up to 40 in	0.002 in	Rule calibrator
Thread Plug Gages – (Pitch & Major Diameter)	Up to 6 in Pitch Diameter Major Diameter	(120 + 1.9D) μin (21 + 1.3D) μin	Supermicrometer™
Supermicrometer™  Up to 10 in total length	Up to 1 in travel from reference	(6.8 + 6.7L) μin	Gage blocks, force gage
LabMaster™  Up to 12 in total length	Up to 2 in travel from reference  Ratiometric / Repeatability 2 in 3 in 4 in	3.0 μin + 3.9 μin/in from Datum  6.2 μin 8.2 μin 10 μin	Gage blocks, force gage
Feeler Gages	Up to 0.25 in	(14 + 0.1L) μin	Supermicrometer™, Labmaster™

II. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC <sup>2, 5, 6</sup> (±)	Comments
Oscilloscopes <sup>3</sup> –			
Square Wave Signal 10 Hz to 10 kHz 50Ω Impedance	(1 to 24.999) mV <sub>p-p</sub> (25 to 109.999) mV <sub>p-p</sub> (0.11 to 2.1999) V <sub>p-p</sub> (2.2 to 6.6) V <sub>p-p</sub>	0.10 mV <sub>p-p</sub> 0.34 mV <sub>p-p</sub> 0.0058 V <sub>p-p</sub> 0.017 V <sub>p-p</sub>	Fluke 5522A- SC1100
10 Hz to 10 kHz 1MΩ Impedance	(1 to 24.999) mV <sub>p-p</sub> (25 to 109.999) mV <sub>p-p</sub> (0.11 to 2.1999) V <sub>p-p</sub> (2.2 to 10.999) V <sub>p-p</sub> (11 to 130) V <sub>p-p</sub>	0.069 mV <sub>p-p</sub> 0.16 mV <sub>p-p</sub> 0.0024 V <sub>p-p</sub> 0.012 V <sub>p-p</sub> 0.15 V <sub>p-p</sub>	
Level Sine Wave Flatness 50 kHz Ref	Up to 10 mV (10 to 40) mV (40 to 400) mV (0.4 to 1.3) V (1.3 to 5.5) V	0.7 mV 0.5 mV 4.3 mV 0.02 V 0.1 V	
50 kHz to 100 MHz	Up to 10 mV (10 to 40) mV (40 to 400) mV (0.4 to 1.3) V (1.3 to 5.5) V	0.25 mV 0.55 mV 4.6 mV 0.015 V 0.075 V	
(100 to 300) MHz	Up to 10 mV 10 to 40 mV 40 to 400 mV 0.4 to 1.3 V 1.3 to 5.5V	0.3 mV 0.7 mV 6.1 mV 0.021 V 0.1 V	
(300 to 600) MHz	Up to 10 mV (10 to 40) mV (40 to 400) mV (0.4 to 1.3) V (1.3 to 5.5) V	0.5 mV 1.3 mV 4.3 mV 0.041 V 0.2 V	
600 MHz to 1.1 GHz	Up to 10 mV (10 to 40) mV (40 to 400) mV (0.4 to 1.3) V	0.6 mV 1.6 mV 15 mV 0.05 V	

Parameter/Equipment	Range	CMC <sup>2, 5, 6</sup> ( $\pm$ )	Comments
Oscilloscopes <sup>3</sup> – (cont)			
Amplitude Time Markers (5-2-1 Sequence) into a 50 $\Omega$ Load	(1 to 2) ns (5 to 100) ns 100 ns to 20 ms 50 ms to 5 s	0.0009 ns 0.0041 ns + 0.0006 ns/ns 0.0042 ms 0.0041 ms + 0.007 s/s	Fluke 5522A-SC1100
Rise Time 10 mV to 2.5 V	1 ns to 1 ms	350 ps	
DC Voltage <sup>3</sup> – Generate	(0 to 220) mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V	9.5 $\mu$ V/V + 0.4 $\mu$ V 5.9 $\mu$ V/V + 0.7 $\mu$ V 4.1 $\mu$ V/V + 2.5 $\mu$ V 4.1 $\mu$ V/V + 4 $\mu$ V 5.9 $\mu$ V/V + 40 $\mu$ V 7.6 $\mu$ V/V + 0.4 mV	Fluke 5730A
DC Voltage <sup>3</sup> – Measure	(0 to 200) mV 200 mV to 2 V (2 to 20) V (20 to 200) V (200 to 1000) V	7.5 $\mu$ V/V + 0.1 $\mu$ V 6.6 $\mu$ V/V + 0.4 $\mu$ V 3.5 $\mu$ V/V + 4 $\mu$ V 5.3 $\mu$ V/V + 40 $\mu$ V 5.2 $\mu$ V/V + 0.4 mV	Fluke 8508A
DC High Voltage – Measure	(1 to 10) kV (10 to 35) kV (35 to 100) kV	0.18 + 0.49 mV/V 0.18 + 0.72 mV/V 0.96 + 1.2 mV/V	Vitrek 4700 with HVL100
DC Current <sup>3</sup> – Generate	(0 to 220) $\mu$ A 220 $\mu$ V to 2.2 mA (2.2 to 22) mA (22 to 220) mA 220 mA to 2.2 A (2.2 to 11) A	48 $\mu$ A/A + 6 nA 42 $\mu$ A/A + 7 nA 42 $\mu$ A/A + 40 nA 53 $\mu$ A/A + 0.7 $\mu$ A 0.12 mA/A + 12 $\mu$ A 0.42 mA/A + 0.48 mA	Fluke 5730A
Clamp-On Only <sup>3</sup>	(11 to 20) A (16.5 to 149.999) A (150 to 1025) A	0.82 mA/A + 0.75 mA 0.39 % + 15 mA 0.48 % + 50 mA	Fluke 5522A Fluke 5522A w/ 5500A Coil

Parameter/Equipment	Range	CMC <sup>2, 6</sup> (±)	Comments
DC Current <sup>3</sup> – Measure	(0 to 200) $\mu$ A 200 $\mu$ A to 2 mA (2 to 20) mA (20 to 200) mA 200 mA to 2 A (2 to 20) A  (20 to 1000) A	14 $\mu$ A/A + 0.4 nA 14 $\mu$ A/A + 4 nA 15 $\mu$ A/A + 40 nA 43 $\mu$ A/A + 0.8 $\mu$ A 0.2 mA/A + 1.6 $\mu$ A 0.44 mA/A + 200 $\mu$ A  0.24 A/A + 0.64 A	Fluke 8508A       Clamp-on ammeter
Resistance <sup>3</sup> – Generate, Fixed Points	0 $\Omega$ 1 $\Omega$ 1.9 $\Omega$ 10.0 $\Omega$ 19.0 $\Omega$ 100 $\Omega$ 190 $\Omega$ 1.0 k $\Omega$ 1.9 k $\Omega$ 10 k $\Omega$ 19 k $\Omega$ 100 k $\Omega$ 190 k $\Omega$ 1.0 M $\Omega$ 1.9 M $\Omega$ 10 M $\Omega$ 19 M $\Omega$ 100 M $\Omega$	46 $\mu\Omega$ 0.11 m $\Omega$ 0.11 m $\Omega$ 0.13 m $\Omega$ 0.27 m $\Omega$ 1.2 m $\Omega$ 1.2 m $\Omega$ 9.8 m $\Omega$ 9.8 m $\Omega$ 0.1 $\Omega$ 0.1 $\Omega$ 1.3 $\Omega$ 1.3 $\Omega$ 24 $\Omega$ 25 $\Omega$ 0.52 k $\Omega$ 0.58 k $\Omega$ 13 k $\Omega$	Fluke 5730A
Resistance <sup>3</sup> – Measure	(0 to 2) $\Omega$ (2 to 20) $\Omega$ (20 to 200) $\Omega$ 200 $\Omega$ to 2 k $\Omega$ (2 to 20) k $\Omega$ (20 to 200) k $\Omega$ 200 k $\Omega$ to 2 M $\Omega$ (2 to 20) M $\Omega$ (20 to 200) M $\Omega$ 200 M $\Omega$ to 2 G $\Omega$ (2 to 20) G $\Omega$	17 $\mu\Omega/\Omega$ + 4 $\mu\Omega$ 11 $\mu\Omega/\Omega$ + 14 $\mu\Omega$ 9.8 $\mu\Omega/\Omega$ + 50 $\mu\Omega$ 13 $\mu\Omega/\Omega$ + 0.5 m $\Omega$ 10 $\mu\Omega/\Omega$ + 5 m $\Omega$ 8.8 $\mu\Omega/\Omega$ + 50 m $\Omega$ 11 $\mu\Omega/\Omega$ + 1 $\Omega$ 18 $\mu\Omega/\Omega$ + 100 $\Omega$ 58 $\mu\Omega/\Omega$ + 10 k $\Omega$ 1.0 m $\Omega/\Omega$ + 1 M $\Omega$ 0.14 % + 10 M $\Omega$	Fluke 8508A

Parameter/Range	Frequency	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
AC Voltage <sup>3</sup> – Generate			
(0.22 to 2.2) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz	1.0 mV/V + 4 $\mu$ V 0.98 mV/V + 4 $\mu$ V 0.91 mV/V + 4 $\mu$ V 0.78 mV/V + 4 $\mu$ V 0.78 mV/V + 5 $\mu$ V 1.3 mV/V + 10 $\mu$ V 1.8 mV/V + 20 $\mu$ V 3.4 mV/V + 20 $\mu$ V	Fluke 5730A
(2.2 to 22) mV	500 kHz to 1 MHz (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	1.4 mV/V + 4 $\mu$ V 0.20 mV/V + 4 $\mu$ V 0.30 mV/V + 46 $\mu$ V 0.30 mV/V + 4 $\mu$ V 0.60 mV/V + 5 $\mu$ V 1.2 mV/V + 10 $\mu$ V 1.6 mV/V + 20 $\mu$ V 3.1 mV/V + 20 $\mu$ V	
(22 to 220) mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.30 mV/V + 12 $\mu$ V 0.10 mV/V + 7 $\mu$ V 0.081 mV/V + 7 $\mu$ V 0.23 mV/V + 7 $\mu$ V 0.50 mV/V + 17 $\mu$ V 1.0 mV/V + 20 $\mu$ V 1.6 mV/V + 25 $\mu$ V 3.1 mV/V + 45 $\mu$ V	
220 mV to 2.2 V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.85 mV/V + 40 $\mu$ V 0.82 mV/V + 15 $\mu$ V 0.82 mV/V + 8 $\mu$ V 0.82 mV/V + 10 $\mu$ V 0.82 mV/V + 30 $\mu$ V 0.92 mV/V + 80 $\mu$ V 1.3 mV/V + 200 $\mu$ V 1.9 mV/V + 300 $\mu$ V	
(2.2 to 22) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.86 mV/V + 400 $\mu$ V 0.82 mV/V + 150 $\mu$ V 0.82 mV/V + 50 $\mu$ V 0.82 mV/V + 100 $\mu$ V 0.82 mV/V + 200 $\mu$ V 0.86 mV/V + 600 $\mu$ V 2.3 mV/V + 2 mV 1.9 mV/V + 3.2 mV	

Parameter/Range	Frequency	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
AC Voltage <sup>3</sup> – Generate (cont)			
(22 to 220) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.86 mV/V + 4 mV 0.82 mV/V + 1.5 mV 0.82 mV/V + 0.6 mV 0.82 mV/V + 1 mV 0.84 mV/V + 2.5 mV 1.3 mV/V + 16 mV 4.5 mV/V + 40 mV 8.0 mV/V + 80 mV	Fluke 5730A
(220 to 250) V	(15 to 50) Hz	0.29 mV/V + 16 $\mu$ V	
(220 to 1100) V	50 Hz to 1 kHz	67 $\mu$ V/V + 3.5 $\mu$ V	
AC Voltage <sup>3</sup> – Measure			
Up to 199.99 mV	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.40 mV/V + 4 $\mu$ V 0.26 mV/V + 4 $\mu$ V 0.26 mV/V + 4 $\mu$ V 0.36 mV/V + 4 $\mu$ V 0.68 mV/V + 5 $\mu$ V 1.4 mV/V + 10 $\mu$ V 2.2 mV/V + 20 $\mu$ V 3.4 mV/V + 20 $\mu$ V	Fluke 5790A
200 mV to 1.9999 V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.27 mV/V + 12 $\mu$ V 0.11 mV/V + 7 $\mu$ V 0.10 mV/V + 7 $\mu$ V 0.21 mV/V + 7 $\mu$ V 0.50 mV/V + 17 $\mu$ V 0.90 mV/V + 20 $\mu$ V 1.4 mV/V + 25 $\mu$ V 2.7 mV/V + 45 $\mu$ V	
(2 to 19.999) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.44 mV/V + 40 $\mu$ V 0.38 mV/V + 150 $\mu$ V 0.37 mV/V + 8 $\mu$ V 0.38 mV/V + 10 $\mu$ V 0.39 mV/V + 30 $\mu$ V 0.56 mV/V + 80 $\mu$ V 1.0 mV/V + 200 $\mu$ V 1.7 mV/V + 300 $\mu$ V	
(20 to 199.99) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.88 mV/V + 400 $\mu$ V 0.76 mV/V + 150 $\mu$ V 0.75 mV/V + 50 $\mu$ V 0.76 mV/V + 100 $\mu$ V 0.77 mV/V + 200 $\mu$ V 0.93 mV/V + 600 $\mu$ V 2.1 mV/V + 2 mV 3.1 mV/V + 3.2 mV	

Parameter/Range	Frequency	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
AC Voltage <sup>3</sup> – Measure (cont)  (200 to 1050) V	(10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.88 mV/V + 400 $\mu$ V 0.76 mV/V + 150 $\mu$ V 0.75 mV/V + 50 $\mu$ V 0.76 mV/V + 100 $\mu$ V 0.77 mV/V + 200 $\mu$ V 0.93 mV/V + 600 $\mu$ V 2.1 mV/V + 600 $\mu$ V 2.7 mV/V + 45 $\mu$ V	Fluke 5790A
AC High Voltage – Measure (50 to 60) Hz	(1 to 10) kV (10 to 35) kV (35 to 100) kV	0.25 V + 4.6 mV/V 0.25 V + 4.8 mV/V 4.9 V + 4.9 mV/V	Vitrek 4700 with HVL100
AC Current <sup>3</sup> – Generate  Up to 220 $\mu$ A	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.36 mA/A + 16 nA 0.21 mA/A + 10 nA 0.14 mA/A + 8 nA 0.33 mA/A + 12 nA 1.3 mA/A + 65 nA	Fluke 5730A
220 $\mu$ A to 2.2 mA	(10 to 20) Hz (20 to 40) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.36 mA/A + 40 nA 0.20 mA/A + 35 nA 0.14 mA/A + 35 nA 0.23 mA/A + 0.11 $\mu$ A 1.3 mA/A + 0.65 $\mu$ A	
(2.2 to 22) mA	(10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.35 mA/A + 0.40 $\mu$ A 0.21 mA/A + 0.35 $\mu$ A 0.14 mA/A + 0.35 $\mu$ A 0.23 mA/A + 0.55 $\mu$ A 1.3 mA/A + 5.0 $\mu$ A	
(22 to 220) mA	(10 to 20) Hz (20 to 40) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.37 mA/A + 4.0 $\mu$ A 0.72 mA/A + 3.5 $\mu$ A 0.14 mA/A + 3.5 $\mu$ A 0.23 mA/A + 3.5 $\mu$ A 1.3 mA/A + 10 $\mu$ A	
220 mA to 2.2 A	20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.30 mA/A + 35 $\mu$ A 0.53 mA/A + 80 $\mu$ A 8.1 mA/A + 0.16 mA	

Parameter/Range	Frequency	CMC <sup>2, 6</sup> ( $\pm$ )	Comments
AC Current <sup>3</sup> – Generate (cont)			
Clamp-on Meters <sup>3</sup>			
Toroidal (16.5 to 149.999) A (150 to 1025) A	(45 to 65) Hz (65 to 440) Hz	0.58 A + 3 mA/A 0.57 A + 3 mA/A	Fluke 5522A w/ 5500A Coil
Non-Toroidal (16.5 to 1025) A	(45 to 440) Hz	0.58 A + 5.8 mA/A	
AC Current <sup>3</sup> – Measure			
(0 to 200) $\mu$ A	(1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz	1.0 mA/A + 20 nA 0.55 mA/A + 20 nA 0.075 mA/A + 20 nA	Fluke 8508A
200 $\mu$ A to 2 mA	(1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz	0.62 mA/A + 0.20 $\mu$ A 0.33 mA/A + 0.20 $\mu$ A 0.75 mA/A + 0.20 $\mu$ A	
(2 to 20) mA	(1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz	0.89 mA/A + 2.0 $\mu$ A 0.32 mA/A + 2.0 $\mu$ A 0.75 mA/A + 2.0 $\mu$ A	
(20 to 200) mA	10 Hz to 10 kHz	0.29 mA/A + 20 $\mu$ A	
200 mA to 2 A	10 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz	0.69 mA/A + 0.20 mA 0.82 mA/A + 0.20 mA 3.5 mA/A + 0.20 mA	
(2 to 20) A	10 Hz to 2 kHz (2 to 10) kHz	0.93 mA/A + 2.0 mA 2.9 mA/A + 2.0 mA	
(20 to 1000) A	(10 to 100) Hz (100 to 500) Hz	0.65 A + 23 mA/A 0.69 A + 29 mA/A	Clamp on ammeter

Parameter/Range	Frequency	CMC <sup>2, 5, 6</sup> (±)	Comments
AC Flatness – Measure			
(220 to 700) mV	(10 to 30) Hz (30 to 120) Hz (0.12 to 1.2) kHz (1.2 to 120) kHz (120 to 500) kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.1 % 0.03 % 0.03 % 0.04 % 0.04 % 0.06 % 0.06 % 0.14 % 0.27 % 0.35 %	Fluke 5790A
700 mV to 2.2 V	(10 to 30) Hz (30 to 120) Hz (0.12 to 1.2) kHz (1.2 to 120) kHz (120 to 500) kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.1 % 0.04 % 0.04 % 0.03 % 0.03 % 0.06 % 0.07 % 0.1 % 0.2 % 0.35 %	
(2.2 to 7) V	(10 to 30) Hz (30 to 120) Hz (0.12 to 1.2) kHz (1.2 to 120) kHz (120 to 500) kHz (0.5 to 1.2) MHz (1.2 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.1 % 0.04 % 0.04 % 0.04 % 0.04 % 0.05 % 0.06 % 0.1 % 0.2 % 0.35 %	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Calibration of Thermocouple Indicators <sup>3</sup> – Generate & Measure			
Type E	(-270 to 0) °C (0 to 600) °C (600 to 1000) °C	0.032 °C 0.035 °C 0.039 °C	Fluke 5730A. Fluke Ice Point 9101, Fluke 8508A
Type J	(-210 to -0) °C (0 to 600) °C (600 to 1200) °C	0.069 °C 0.038 °C 0.044 °C	
Type K	(-200 to 0) °C (0 to 600) °C (600 to 1372) °C	0.091 °C 0.045 °C 0.063 °C	
Type N	(-200 to 0) °C (0 to 600) °C (600 to 1300) °C	0.15 °C 0.057 °C 0.057 °C	
Type R	(0 to 400) °C (400 to 1000) °C (1000 to 1767) °C	0.27 °C 0.14 °C 0.13 °C	
Type S	(0 to 500) °C (500 to 1000) °C (1000 to 1767) °C	0.27 °C 0.18 °C 0.13 °C	
Type T	(-250 to 0) °C (0 to 120) °C (120 to 400) °C	0.21 °C 0.042 °C 0.038 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Calibration of RTD Indicators <sup>3</sup> – Generate & Measure			
Pt 385, 100 Ω	(-200 to 0) °C	0.05 °C	Fluke 5522A, Fluke 8508A
	(0 to 100) °C	0.07 °C	
	(100 to 300) °C	0.09 °C	
	(300 to 400) °C	0.10 °C	
	(400 to 630) °C	0.12 °C	
	(630 to 800) °C	0.23 °C	
Pt 3926, 100 Ω	(-200 to 0) °C	0.05 °C	
	(0 to 100) °C	0.07 °C	
	(100 to 300) °C	0.09 °C	
	(300 to 400) °C	0.10 °C	
	(400 to 630) °C	0.12 °C	
Pt 3916, 100 Ω	(-200 to -190) °C	0.25 °C	
	(-190 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.05 °C	
	(0 to 100) °C	0.06 °C	
	(100 to 260) °C	0.07 °C	
	(260 to 300) °C	0.08 °C	
	(300 to 400) °C	0.09 °C	
	(400 to 600) °C	0.10 °C	
	(600 to 630) °C	0.23 °C	
Pt 385, 200 Ω	(-200 to 100) °C	0.04 °C	
	(100 to 260) °C	0.05 °C	
	(260 to 300) °C	0.12 °C	
	(300 to 400) °C	0.13 °C	
	(400 to 600) °C	0.14 °C	
	(600 to 630) °C	0.16 °C	
Pt 385, 500 Ω	(-200 to -80) °C	0.04 °C	
	(-80 to 100) °C	0.05 °C	
	(100 to 260) °C	0.06 °C	
	(260 to 400) °C	0.08 °C	
	(400 to 600) °C	0.09 °C	
	(600 to 630) °C	0.11 °C	
Pt 385, 1000 Ω	(-200 to 0) °C	0.03 °C	
	(0 to 100) °C	0.04 °C	
	(100 to 260) °C	0.05 °C	
	(260 to 300) °C	0.06 °C	
	(300 to 600) °C	0.07 °C	
	(600 to 630) °C	0.23 °C	

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Electrical Calibration of RTD Indicators <sup>3</sup> (cont) – Generate & Measure			
PtNi 385, 120 Ω	(-80 to 100) °C (100 to 260) °C	0.08 °C 0.14 °C	Fluke 5522A, Fluke 8508A
Cu 427, 10 Ω	(-100 to 260) °C	0.3 °C	

Parameter/Range	Frequency	CMC <sup>2</sup> (±)	Comments
Distortion – Measure			
(-75 to 0) dB Dist	10 Hz to 500 kHz	0.74 dB	U8903A
(-99.99 to 0) dB Dist	9 kHz to 30 MHz	1.2 dB	E4440A
Capacitance – Generate <sup>3</sup>			
(220 to 399.9) pF	10 Hz to 10 kHz	12 pF	Fluke 5522A
(0.4 to 1.0999) nF	10 Hz to 10 kHz	0.015 nF	
(1.1 to 3.2999) nF	10 Hz to 3 kHz	0.035 nF	
(3.3 to 10.999) nF	10 Hz to 1 kHz	0.035 nF	
(11 to 32.999) nF	10 Hz to 1 kHz	0.085 nF	
(33 to 109.999) nF	10 Hz to 1 kHz	0.26 nF	
(110 to 329.999) nF	10 Hz to 1 kHz	0.78 nF	
(0.33 to 1.1) μF	(10 to 600) Hz	0.0035 μF	
(1.1 to 3.3) μF	(10 to 300) Hz	0.011 μF	
(3.3 to 11) μF	(10 to 150) Hz	0.035 μF	
(11 to 33) μF	(10 to 120) Hz	0.15 μF	
(33 to 110) μF	(10 to 80) Hz	0.55 μF	
(110 to 330) μF	DC to 50 Hz	2.4 μF	
(0.33 to 1.1) mF	DC to 20 Hz	0.0055 mF	
(1.1 to 3.3) mF	DC to 6 Hz	0.017 mF	
(3.3 to 11) mF	DC to 2 Hz	0.055 mF	
(11 to 33) mF	DC to 0.6 Hz	0.26 mF	
(33 to 110) mF	DC to 0.2 Hz	1.2 mF	

Parameter/Range	Frequency	CMC <sup>2, 5, 6</sup> ( $\pm$ )	Comments
Capacitance – Generate <sup>3</sup> (cont)  Fixed Points  1 pF 10 pF 100 pF 1000 pF 0.01 $\mu$ F 0.10 $\mu$ F 1.0 $\mu$ F	1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz 1 kHz	0.0012 pF 0.012 pF 0.12 pF 1.2 pF 0.000 74 $\mu$ F 0.000 82 $\mu$ F 0.0015 $\mu$ F	HP 16380A & HP 16380C standard capacitor sets
Capacitance – Measure  10 pF 100 pF 1000 pF 0.1 $\mu$ F 10 $\mu$ F	120 Hz to 300 kHz 120 Hz to 300 kHz 120 Hz to 300 kHz 120 Hz to 100 kHz 120 Hz to 100 kHz	0.028 pF 0.19 pF 1.2 pF 0.000 61 $\mu$ F 0.093 $\mu$ F	Agilent E4980A

### III. Electrical RF/Microwave

Parameter/Equipment	Range	CMC <sup>2, 4, 5, 7</sup> ( $\pm$ )	Comments
RF Attenuation - Tuned RF Power Measurement	100kHz to 10 MHz  10MHz to 3.05 GHz (3.05 to 6.6) GHz (6.6 to 13.2) GHz (13.2 to 19.2) GHz (19.2 to 26.5) GHz	0.89 dBm  0.2 dBm 0.2 dBm 0.2 dBm 0.2 dBm 0.2 dBm	SP 8902A with 11722A  N5531S MMR System
RF Power – Generate	(20 to 24) dBm (0.01 to 100) kHz (0.1 to 125) MHz  (14 to 20) dBm (0.01 to 100) kHz (0.1 to 125) MHz (125 to 300) MHz (0.3 to 1.4) GHz	0.06 dB 0.06 dB  0.04 dB 0.06 dB 0.09 dB 0.23 dB	Fluke 96270A Reference Source w/Leveling Head Output

Parameter/Equipment	Range	CMC <sup>3, 4, 5, 7</sup> ( $\pm$ )	Comments	
RF Power – Generate (cont)	(-17 to 14) dBm			
	(0.01 to 100) kHz	0.04 dB	Fluke 96270A Reference Source w/Leveling Head Output	
	(0.1 to 125) MHz	0.06 dB		
	(125 to 300) MHz	0.09 dB		
	(0.3 to 1.4) GHz	0.23 dB		
	(1.4 to 4.0) GHz	0.35 dB		
	(-120 to 16) dBm			Fluke 96270A Reference Source w/11667B Splitter
	(2.4 to 8.0) GHz	0.8 dB		
	(8.0 to 12.0) GHz	0.8 dB		
	(12.0 to 18.0) GHz	0.8 dB		
(18.0 to 22.0) GHz	0.8 dB			
(22.0 to 26.5) GHz	0.8 dB			

Parameter/Range	Frequency	CMC <sup>4, 4, 5, 7</sup> ( $\pm$ )	Comments	
RF Attenuation - Tuned RF Power Measurement			SP 8902A with 11722A	
	(-48 to -17) dBm		N5531S MMR System	
	(0.01 to 100) kHz	0.04 dB		
	(0.1 to 125) MHz	0.06 dB		
	(125 to 300) MHz	0.09 dB		
	(0.3 to 1.4) GHz	0.25 dB		
	(-74 to -48) dBm			N5531S MMR System
	(0.1 to 10) MHz	0.23 dB		
	(10 to 125) MHz	0.11 dB		
	(125 to 300) MHz	0.11 dB		
	(0.3 to 1.4) GHz	0.46 dB		
	(-84 to -74) dBm			N5531S MMR System
	(0.1 to 10) MHz	0.60 dB		
	(10 to 125) MHz	0.10 dB		
	(125 to 300) MHz	0.34 dB		
	(0.3 to 1.4) GHz	0.57 dB		
	(-94 to -84) dBm			N5531S MMR System
	(0.1 to 10) MHz	0.60 dB		
	(10 to 125) MHz	0.40 dB		
	(125 to 300) MHz	0.60 dB		
(0.3 to 3.0) GHz	1.2 dB			
(-94 to -124) dB			N5531S MMR System	
(10 to 125) MHz	0.80 dB			
(125 to 300) MHz	1.8 dB			
(0.3 to 1.4) GHz	1.8 dB			
(1.4 to 3.0) GHz	1.8 dB			

Parameter/Range	Frequency	CMC <sup>5, 4, 5, 7</sup> ( $\pm$ )	Comments
RF Power – Measure (+20 to -30) dBm (+20 to -10) dBm	DC – 30 MHz 30 MHz – 2.0 GHz (2 to 4.2) GHz (4.2 to 18) GHz (18 to 26.5) GHz	0.52 dBm + 1 digit 0.23 dBm 0.23 dBm 0.32 dBm 0.45 dBm	8902A with 1722A N5531S MMR System with N5532B
Power Reference Measure	1 mW, 50 MHz	1.2 %	N432A, HP 3458A, HP 478A-H55 Thermistor Mount

Parameter/Equipment	Range	CMC <sup>6, 4, 5, 7</sup> ( $\pm$ )	Comments
Amplitude Modulation – Measure 100 kHz to 10 MHz 10 MHz to 3 GHz 3 GHz to 26.5 GHz	Rate: 50 Hz to 10 kHz Depth: 5% to 99% Rate: 50 Hz to 100 kHz Depth: 20 % to 99% Depth: 5% to 20% Rate: 50 Hz to 100kHz Depth: 20 % to 99% Depth: 5% to 20%	2.3 % + 1 digit 0.58 % 2.9 % 1.7 % 5.2 %	8902A Measuring Receiver N5531S MMR System with N5532B
Frequency Modulation – Measure Rate: 0.02 to 10 kHz Rate: 0.05 to 200 kHz	$\leq 40\text{kHz}_{\text{peak}}$ (0.25 to 10) MHz $\leq 400\text{kHz}_{\text{peak}}$ (0.01 to 26.5) GHz	1.2 % Deviation 1.2 % Deviation	N5531S MMR System with N5532B
Phase Modulation – Measure 100 kHz to 6.6 GHz (6.6 to 13.2) GHz (13.2 to 26.5) GHz	$> 0.7$ Rad $> 0.3$ to $0.7$ Rad $> 2.0$ Rad $> 0.6$ to $2.0$ Rad $> 4.0$ Rad $> 1.2$ to $4.0$ Rad	1.6 % Deviation 3.7 % Deviation 1.3 % Deviation 3.5 % Deviation 1.2 % Deviation 3.5 % Deviation	N5531S MMR System with N5532B

Parameter/Equipment	Range	CMC <sup>7, 4, 5, 7</sup> (±)	Comments
Single Sideband Phase Noise (SSB) – Measure  CW Frequency: 10MHz to 26.5 GHz	Offset from CW: 100Hz to 1 MHz	1.9 dB	Keysight E4440A w/ N5532B-526

IV. Mechanical

Parameter/Equipment	Range	CMC <sup>2, 5, 7</sup> (±)	Comments
Acceleration & Vibration Amplitude Response	(5 to 9) Hz (10 to 99) Hz 100 Hz (101 to 920) Hz (921 to 5000) Hz (5 to 10) kHz (10 to 15) kHz (15 to 20) kHz	2.2 % 1.7 % 1.3 % 1.4 % 1.7 % 2.2 % 2.8 % 3.5 %	Modal Shop 9155 System using Back- to-Back Calibration Method
Pressure – Measure & Measuring Equipment			
Pneumatic	(-2.2 to 2.2) psig (0 to 15) psig	0.000 12 psi 0.0004 psi	Fluke PPC3 pressure calibrator
	(-14.7 to 60) psig Up to 200 psig Up to 500 psig Up to 1000 psig	0.0090 psi 0.024 psi 0.060 psi 0.13 psi	Mensor CPC 6000
Pneumatic	(0 to 34) inHg	0.012 inHg	Mensor CPC 6000
Hydraulic	(50 to 10 000) psi  (0 to 30 000) psi	0.0093 %  35 PSIG + 3.0 μPSIG/PSIG	Deadweight tester  Additel ADT672

Parameter/Equipment	Range	CMC <sup>8, 4, 5, 7</sup> (±)	Comments
Pressure – Field <sup>3</sup> , Measure & Measuring Equipment			
Hydraulic	(5 to 10 000) psig	0.12 %	Deadweight tester
Pneumatic	Up to 1 psig Up to 5 psig Up to 30 psig Up to 100 psig Up to 300 psig Up to 1000 psig Up to 5000 psig	0.0012 psi 0.04 psi 0.018 psi 0.12 psi 0.17 psi 0.59 psi 4.7 psi	Fluke modules
Scales & Balances <sup>3</sup> –			
Linearity	0.01 mg to 42 g (43 to 210) g 210 g to 5.1 kg (5.1 to 30) kg	0.000 21 g 0.000 25 g 0.0028 g 0.028 g	Class 00, Class 0 & Class 1 weights
Repeatability	0.01 mg to 42 g (43 to 210) g 210 g to 5.1 kg (5.1 to 30) kg	0.000 028 g 0.000 14 g 0.019 g 0.000 26 kg	
Scales & Balances <sup>3</sup>	(0 to 1500) lb	0.0058 lbf + 0.0004 lbf/lbf	Class F weights
Mass – Measure	0.01 mg to 42 g (43 to 210) g 210 g to 6 kg (6 to 30) kg	0.000 03 g + 0.36 µg/g 0.000 15 g + 0.08 µg/g 0.018 g + 1.9 µg/g 0.17 g + 0.8 mg/g	Analytical balances Top loading balances
Torque – Measure	0.5 ozf·in to 2000 lbf·ft (500 to 5000) lbf·ft	0.3 % 0.5 %	Torque transducers
Torque <sup>3</sup> – Measure	5 lbf·in to 600 lbf·ft	0.35 %	Torque transducers
Torque Transducers	(0.5 to 400) ozf·in (4 to 150) lbf·in (150 to 400) lbf·in (10 to 2000) lbf·ft	0.05 % 0.05 % 0.06 % 0.05 %	Torque arms Class F weights

Parameter/Equipment	Range	CMC <sup>2, 5, 7</sup> (±)	Comments
Force Measurement <sup>3</sup> Compression & Tension  Compression Only	(1 to 10) lbf (10 to 100) lbf (100 to 500) lbf (500 to 1000) lbf  (1000 to 5000) lbf (5000 to 10 000) lbf (10 000 to 50 000) lbf (50 000 to 500 000) lbf	0.0058 lbf + 15 µlbf/ lbf 0.0058 lbf + 34 µlbf / lbf 0.053 lbf + 58 µlbf / lbf 0.57 lbf + 20 µlbf / lbf  5.8 lbf + 33 µlbf / lbf 12 lbf + 24 µlbf / lbf 57 lbf + 74 µlbf / lbf 210 lbf + 620 µlbf / lbf	Load cells, Class F, Class 1 weights  Load cells
Force – Measure & Measuring Equipment  Compression  Tension	(1 to 1000) lbf  (65 to 6500) lbf (6500 to 50 000) lbf  (1 to 1000) lbf  (65 to 6500) lbf (500 to 50 000) lbf	0.12 %  0.98 lbf 7.3 lbf  0.12 %  0.98 lbf 7.3 lbf	Class F Weights  Proving rings  Class F Weights Proving rings
Nuclear Density Gauges, Fixed Points – Density <sup>3</sup>  Magnesium Mag/Aluminum Aluminum	110.45 lb/ft <sup>3</sup> 139.20 lb/ft <sup>3</sup> 167.70 lb/ft <sup>3</sup>	1.2 lb/ft <sup>3</sup> 1.6 lb/ft <sup>3</sup> 1.9 lb/ft <sup>3</sup>	Density blocks
Nuclear Density Gauges, Fixed Points – Moisture <sup>3</sup>	Up to 35 lb/ft <sup>3</sup>	1.1 lb/ft <sup>3</sup>	Moisture blocks

#### V. Thermodynamics

Parameter/Equipment	Range	CMC <sup>2, 5, 7</sup> (±)	Comments
Thermometers <sup>3</sup>	(-80 to -20) °C (-20 to 100) °C (100 to 400) °C (400 to 650) °C	0.012 °C 0.015 °C 0.038 °C 0.040 °C	Liquid bath Dry block/PRT

Parameter/Equipment	Range	CMC <sup>2, 5, 7</sup> (±)	Comments
Infrared Thermometers – Measuring Equipment	(-15 to 120) °C (100 to 200) °C (200 to 350) °C (350 to 500) °C	0.7 °C 0.73 °C 1.0 °C 1.7 °C	Fluke 4180/4181  Where wavelength (λ) = (8 to 14) μm & emissivity (ε) = 0.95
Furnaces, Ovens, Refrigerators & Freezers <sup>3</sup> – Temperature Only	0 °C to 900 °C	1.4 °C	Fluke 1586 RTD Thermocouple
Temp/Humidity Indicators –			
Relative Humidity	(20 to 95) % RH	0.65 %	Thunder Scientific 2500
Temperature	(0 to 50) °C	0.20 °C	
Temperature – Measure	(-80 to < 0) °C 0 °C (> 0 to 200) °C (> 200 to 400) °C (400 to 650) °C	0.084 °C 0.019 °C 0.040 °C 0.053 °C 0.067 °C	Fluke 1502A with 5628 PRT
Relative Humidity – Measure <sup>3</sup>	(13 to 73) % RH @ 25 °C  (10 to 90) % RH	0.65 % RH  0.65 % RH	RH Systems 473 chilled mirror
Temperature Uniformity Survey <sup>3</sup>			
J-Type Thermocouples	(0 to 1200) °F	2.3 °F + 0.0029 °F/°F	IAW AMS2750
K-Type Thermocouples	(0 to 1850) °F	2.2 °F + 0.0035 °F/°F	
System Accuracy Test <sup>3</sup>	(0 to 1850) °F	2.2 °F + 0.0035 °F/°F	IAW AMS2750

## V. Time & Frequency

Parameter/Equipment	Range	CMC <sup>2, 7</sup> ( $\pm$ )	Comments
Stopwatches/Timers <sup>3</sup>	Up to 30 days	0.037 s/day	Timometer
Tachometer <sup>3</sup> – Non-contact	(0 to 100 000) RPM	0.062 RPM + 0.000 005 6 RPM/RPM	33250A Function Generator & LED strobe.
Contact	(0 to 1500) RPM	0.13 RPM + 0.0014 RPM/RPM	Reference Tachometer Comparison
Frequency – Measure <sup>3</sup>	(0 to 200) MHz 50 MHz to 26.5 GHz	1.2 parts in 10 <sup>7</sup> Hz/Hz 1.2 parts in 10 <sup>7</sup> Hz/Hz	Agilent 53151A Agilent 53132A
Frequency – Measuring Equipment	10 MHz  DC to 4 GHz (4 to 26.5) GHz	3.4 parts in 10 <sup>9</sup> MHz  5.8 parts in 10 <sup>8</sup> Hz/Hz 5.8 parts in 10 <sup>8</sup> Hz/Hz	Agilent 33120A, 4410A GPS  Fluke 96270A w/Leveling Head & Microwave Output
Frequency – Measuring Equipment <sup>3</sup>	10 MHz  DC to 4 GHz (4 to 26.5) GHz	6 parts in 10 <sup>10</sup> Hz/Hz  5.8 parts in 10 <sup>8</sup> Hz/Hz 5.8 parts in 10 <sup>8</sup> Hz/Hz	4410A GPS Frequency Standard  Fluke 96270A w/Leveling Head & Microwave Output

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the Calibration and Measurement Capability Uncertainty (CMC) found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> In the statement of CMC,  $L$  is the numerical values of the nominal length of the device measured in inches,  $R$  is the numerical value of the resolution of the device in microinches, and  $D$  is the diameter of the device in inches

<sup>5</sup> In the statement of CMC, a % denotes a percent of reading unless otherwise noted.

<sup>6</sup> The measurands stated are generated using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure the measurand in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a fraction/percentage of the reading plus a fixed floor specification.

<sup>7</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter

<sup>8</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



## Accredited Laboratory

A2LA has accredited

### **NATIONAL CALIBRATION INC.**

*Scottsdale, AZ*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 28<sup>th</sup> day of August 2024.

A blue ink signature of Trace McInturff, written over a horizontal line.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1892.01  
Valid to August 31, 2026

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*