



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

ASSOCIATED ENVIRONMENTAL SYSTEMS
8 Post Office Square
Acton, MA 01720
Justin Ingram Phone: 978 772 0022

CALIBRATION

Valid To: July 31, 2022

Certificate Number: 3964.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1,4}:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2,5} (±)	Comments
DC Voltage ³ – Generate Fixed Points	110 mV 1 V 15 V	0.8 mV 0.02 V 0.08 V	Fluke 753
DC Resistance ³ – Generate	(0 to 10) Ω (0 to 100) Ω (0 to 1000) Ω	0.007 Ω 0.09 Ω 0.7 Ω	Fluke 753
DC Current ³ – Generate	(0 to 22) mA	0.12 mA	Fluke 753
Electrical Calibration of Thermocouple Indicating Systems ³ –			
Type T	(-200 to 0) °C (0 to 400) °C	0.72 °C 0.82 °C	Fluke 753
Type J	(0 to 1200) °C	1.4 °C	
Type K	(0 to 1372) °C	1.7 °C	

Parameter/Equipment	Range	CMC ² (±)	Comments
Electrical Calibration of RTD Indicating Systems ³ – Pt 385 100 Ω	(-200 to 250) °C	0.72 °C	Fluke 753

II. Thermodynamics

Parameter/Equipment	Range	CMC ^{2, 6} (±)	Comments
Relative Humidity ³ – Measure	(10 to 95) % RH	3.6 % RH	Vaisala HMP 77B
Temperature ³ – Measure	(-80 to 0) °C	1.4 °C	Fluke 753 with T, J and K thermocouples
	(0 to 200) °C	2.5 °C	
	(0 to 760) °C	2.3 °C	
	(0 to 1100) °C	1.7 °C	Fluke 2638A hydra data logger
	(-73 to 0) °C	0.35 °C	
	(12 to 45) °C	0.29 °C	
	(0 to 100) °C	0.41 °C	Class A RTD

¹ This laboratory offers commercial and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ This scope meets A2LA's P112 Flexible Scope Policy.

- ⁵ The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.
- ⁶ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.





Accredited Laboratory

A2LA has accredited

ASSOCIATED ENVIRONMENTAL SYSTEMS

Acton, MA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 11th day of November 2020.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 3964.01
Valid to July 31, 2022

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.