



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

ELLIOTT MATSUURA CANADA, INC.  
2120 Buckingham Road  
Oakville, Ontario, CANADA L6H 5X2  
Matthew Joyce Phone: 905 829 2211

CALIBRATION

Valid To: December 31, 2024

Certificate Number: 2752.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Coordinate Measuring Machines (CMM) <sup>3</sup> –			
Length Measurement Errors	Up to 1100 mm	(0.08 + 0.3L) µm	ISO 10360-2, ITP step bar
Roundness Measurement Error	Ring Nominal Diameter: 50 mm	0.13 µm	VDI/VDE 2617 sheet 2.2 (2000) with ring gauge
Scanning Error (MPE <sub>THP</sub> )	Sphere Nominal Diameter: 25 mm	0.042 µm	ISO 10360-4 with test sphere
Probing Error (MPE <sub>PFTU</sub> )	Sphere Nominal Diameter: 25 mm	0.042 µm	ISO 10360-5 with test sphere
Rotary Table as a Fourth Axis – Errors of Indication:			
(FR)	Height Difference	0.37 µm	ISO 10360-3 with two test spheres on rotary table
(FT)	and Radius: 200 mm	0.37 µm	
(FA)	(Nominal); Sphere Nominal Diameter: 30 mm	0.45 µm	

<sup>1</sup> This laboratory offers commercial and field calibration service.



<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> This laboratory performs field calibration activities for these parameters. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> In the statement of CMC,  $L$  is the numerical value of the nominal length of the device measured in meters.



# Accredited Laboratory

A2LA has accredited

**ELLIOTT MATSUURA CANADA, INC.**

*Oakville, Ontario, CANADA*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system *(refer to joint ISO-ILAC-IAF Communiqué dated April 2017)*.



Presented this 1<sup>st</sup> day of December 2022.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 2752.01  
Valid to December 31, 2024

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*