



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

PRECISION LABS BY APS DYNAMICS, INC.<sup>7</sup>  
32124 Paseo Adelanto, Suite 3  
San Juan Capistrano, CA 92675  
Bruce Swanson Phone: 949 606 4780

CALIBRATION

Valid To: February 28, 2026

Certificate Number: 2405.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1, 6</sup>:

I. Mechanical

Parameter/Equipment	Range <sup>4</sup>	CMC <sup>2, 5</sup> ( $\pm$ )	Comments <sup>3</sup>
Vibration –  Magnitude/Phase for Frequency Intervals: (0.1 to <0.2) Hz (0.2 to <0.4) Hz (0.4 to <1) Hz (1 to 63) Hz (>63 to 160) Hz  (5 to <10) Hz (10 to <20) Hz 20 Hz to 1 kHz (>1 to 5) kHz (>5 to 10) kHz (>10 to 15) kHz (>15 to 20) kHz  (Accelerometers, Velocity Sensors, Vibration Meters, Vibration Calibrators, Vibration Displacement Sensors Vibration Controllers, Portable Shakers)	(0.01 to 50) m/s <sup>2</sup>       (1.0 to 500) m/s <sup>2</sup>	1.5 %/2.0° 1.0 %/1.0° 0.7 %/0.7° 0.5 %/0.7° 1.0 %/1.0°  1.0 %/1.0° 0.7 %/0.67° 0.5 %/0.50° 0.7 %/0.7° 1.5 %/1.0° 2.5 %/2.0° 3.0 %/2.5°	Comparison method, SPEKTRA CS18 w/ Phase, APS Dynamics long stroke shaker; (0.001 to 5) g <sub>n</sub>      Comparison method, SPEKTRA CS18 w/ Phase, SE09 shaker, SE-10 or other electrodynamics shaker; (0.1 to 50) g <sub>n</sub>

Parameter/Equipment	Range <sup>3, 4</sup>	CMC <sup>2, 5</sup> ( $\pm$ )	Comments
Accelerometers (Secondary)  Charge (Charge Amplifier)  0.2 Hz to 20 kHz (>20 to 50) kHz	(0.1 to 10 000) pC	0.26 %/0.6° 1.1 %/1.1°	Internal comparison method, SPEKTRA CS18 w/ phase
Accelerometers (Secondary)  Voltage (Measuring Amplifier)  0.2 Hz to 20 kHz (>20 to 50) kHz	1.0 mV to 30 V	0.26 %/0.6° 1.1 %/1.1°	Internal comparison method, SPEKTRA CS18 w/ phase
Accelerometers –  Sensitivity/ Frequency Response  (1 to 100) Hz  (10 to <100) Hz 100 Hz to >2.5 kHz (>2.5 to 10) kHz (>10 to 15) kHz	(1 to 3) $g_n$  (0.25 to 10) $g_n$	2.3 %  1.5 % 1.2 % 2.0 % 3.4 %	Comparison method, Endevco 68357, APS dynamics long stroke shaker  Comparison method, Endevco 68357, Bouche shaker
Accelerometers <sup>3</sup> –  Shock Amplitude	(20 to 2000) $g_n$ (>2000 to 10 000) $g_n$	1.7 % 2.7 %	Comparison method, Endevco 2925 POP system

Parameter/Equipment	Range <sup>4</sup>	CMC <sup>2</sup> (±)	Comments
Acoustics –			
Sound Pressure Level Microphones			
31.5 Hz to 5 kHz (5.0 to 10) kHz (10.0 to 16) kHz	(84 to 114) dB	0.20 dB 0.25 dB 0.50 dB	Comparison method, SPEKTRA CS18AK w/ acoustic coupler
Sound Level Meters			
31.5 Hz to 5 kHz (5.0 to 10) kHz (10.0 to 16) kHz	(84 to 114) dB	0.20 dB 0.25 dB 0.50 dB	Comparison method, SPEKTRA CS18AK w/ acoustic coupler
Sound Pressure level 1000 Hz  (Microphones, Sound Level Meters, Acoustic Calibrators)	(94 to 114) dB	0.20 dB	Substitution method, SPEKTRA CS18AK w/ reference microphone and reference calibrator

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup>  $g_n$  is the standard gravitational acceleration due to gravity.

<sup>4</sup> For sinusoidal excitation and narrow-band evaluation, the amplitudes of vibration acceleration, velocity and displacement are unambiguously linked to one another by the vibration frequency; therefore, velocity and displacement sensors are calibrated using the measurand acceleration and frequency, as stated in the table.

<sup>5</sup> In the statement of CMC, percentages are to be read as percent of reading unless otherwise noted.

<sup>6</sup> This scope meets A2LA's *PI12 Flexible Scope Policy*.

<sup>7</sup> APS Dynamics, Inc is the legal entity doing business as Precision Labs.



## Accredited Laboratory

A2LA has accredited

**PRECISION LABS BY APS DYNAMICS, INC.**

*San Juan Capistrano, CA*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 31<sup>st</sup> day of January 2024.

A blue ink signature of Trace McInturff, Vice President of Accreditation Services.

Vice President, Accreditation Services  
Mr. Trace McInturff, For the Accreditation Council  
Certificate Number 2405.01  
Valid to February 28, 2026  
Revised February 9, 2024

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*