



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

ADVANCED CHAMBER SERVICES
3361 Matoaka Road
Suffolk, VA 23434-7116
Pierre Rosa Phone: 757 809 1612

CALIBRATION

Valid To: January 31, 2025

Certificate Number: 3184.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1,5}:

I. Thermodynamics

Parameter/Equipment	Range	CMC ^{2,4} (\pm)	Comments
Relative Humidity – Measure ³	(10 to 95) % RH	1.6 % RH	Vaisala M170 & HMP77B
Temperature – Measure ³	(-80 to 150) °C	1.0 °C	Fluke 1524 & 5327A

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ This laboratory performs field calibration activities for these parameters.

⁴ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

⁵ This scope meets A2LA's *P112 Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

ADVANCED CHAMBER SERVICES

Suffolk, VA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 21st day of December 2022.

A blue ink signature of a person, written over a horizontal line.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 3184.01
Valid to January 31, 2025

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.