



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

NEWAGE TESTING INSTRUMENTS, INC.  
205 Keith Valley Road  
Horsham, PA 19044  
Natercia Ball Phone: 727 538 6095

CALIBRATION

Valid To: March 31, 2023

Certificate Number: 1734.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Mechanical

Parameter/Equipment	Range	CMC <sup>2, 4, 5</sup> ( $\pm$ )	Comments
Indirect Verification of Rockwell and Rockwell Superficial Hardness Testers <sup>3</sup>	HRA: High Medium Low  HRBW: High Medium Low  HRC: High Medium Low	0.21 HRA 0.2 HRA 0.43 HRA  0.5 HRBW 0.67 HRBW 1.1 HRBW  0.32 HRC 0.33 HRC 0.39 HRC	ASTM E18, ISO 6508

Parameter/Equipment	Range	CMC <sup>2, 4, 5</sup> ( $\pm$ )	Comments
Indirect Verification of Rockwell and Rockwell Superficial Hardness Testers <sup>3</sup> (cont)	HRD: High Medium Low  HREW: High Medium Low  HRFW: High Medium Low  HRGW: High Medium Low  HRHW: High Low  HRKW: High Medium Low  HRRW: High Low  HR15N: High Medium Low  HR15TW: High Medium Low  HR30N: High Medium Low	0.12 HRD 0.28 HRD 0.15 HRD  0.50 HREW 0.19 HREW 0.22 HREW  0.47 HRFW 0.46 HRFW 0.46 HRFW  0.29 HRGW 0.25 HRGW 0.80 HRGW  0.37 HRHW 0.43 HRHW  0.37 HRKW 0.39 HRKW 0.51 HRKW  0.23 HRRW 0.34 HRRW  0.22 HR15N 0.24 HR15N 0.48 HR15N  0.43 HR15TW 0.28 HR15TW 0.28 HR15TW  0.52 HR30N 0.53 HR30N 0.37 HR30N	ASTM E18, ISO 6508

Parameter/Equipment	Range	CMC <sup>2, 4, 5</sup> ( $\pm$ )	Comments
Indirect Verification of Rockwell and Rockwell Superficial Hardness Testers <sup>3</sup> (cont)	HR30TW: High Medium Low  HR45N: High Medium Low  HR45TW: High Medium Low	0.29 HR30TW 0.32 HR30TW 0.58 HR30TW  0.18 HR45N 0.34 HR45N 0.50 HR45N  0.40 HR45TW 0.44 HR45TW 0.70 HR45TW	ASTM E18, ISO 6508
Indirect Verification of Leeb Hardness Testers <sup>3</sup>	(0 to 750) LD	8 LD	ASTM A956
Indirect Verification of Brinell Hardness Testers <sup>3</sup> –			
HBW 10/3000/10	(96 to 372) HBW (373 to 650) HBW	2.1 HBW 5.9 HBW	ASTM E10, E103; ISO 6506
HBW 10/1500/10	(48 to 186) HBW (187 to 327) HBW	1.5 HBW 2.7 HBW	
HBW 10/1000/10	(96 to 372) HBW (373 to 650) HBW	1.0 HBW 2.2 HBW	
HBW 10/500/10	(16 to 61) HBW (62 to 109) HBW	0.5 HBW 1.1 HBW	
HBW 5/750/10	(96 to 372) HBW (373 to 650) HBW	1.3 HBW 8.8 HBW	
HBW 5/250/10	(32 to 124) HBW (125 to 218) HBW	1.2 HBW 2.1 HBW	
HBW 2.5/187.5/10	(96 to 372) HBW (373 to 650) HBW	2.9 HBW 18 HBW	

Parameter/Equipment	Range	CMC <sup>2, 4, 5</sup> ( $\pm$ )	Comments
Indirect Verification of Microindentation Hardness Testers <sup>3</sup> – Knoop and Vickers	<p>Vickers (<math>\leq 1</math> kgf): (100 to 240) HV (240 to 600) HV &gt;600 HV</p> <p>Vickers (<math>\geq 1</math> kgf): (100 to 240) HV (240 to 600) HV &gt;600 HV</p> <p>Knoop: (100 to 250) HK (250 to 650) HK &gt;650 HK</p>	<p>3.9 HV 7.5 HV 11 HV</p> <p>2.9 HV 7.4 HV 14 HV</p> <p>5.5 HK 15 HK 15 HK</p>	ASTM E92, ISO 6507, ASTM E384
Direct Verification of Rockwell Hardness Testers –			
Verification of the Test Force	(3 to 150) kgf	0.06 % of full scale	ASTM E18, ISO 6508
Verification of the Depth – Measuring Device	(0 to 260) $\mu$ m	0.19 $\mu$ m	
Direct Verification of Brinell Hardness Testers <sup>3</sup> –			ASTM E10, E103; ISO 6506
Verification of the Test Force	(125 to 3000) kgf	0.19 % of full scale	Verification of the test force is by load cell per the method of ASTM E10
Verification of the Device for Measuring Indentation Diameter	(0 to 7) mm	2.1 $\mu$ m	By stage micrometer

Parameter/Equipment	Range	CMC <sup>2,4</sup> ( $\pm$ )	Comments
Direct Verification of Microindentation Hardness Testers <sup>3</sup> – Vickers and Knoop			ASTM E92, ISO 6507, ASTM E384
Verification of the Test Force	(0 to 1000) gf Vickers Knoop	0.26 % of full scale 0.48 % of full scale	Verification of the test force is by load cell per the method of ASTM E92, ASTM E384
Verification of the Device for Measuring Indentation Diagonals	(0 to 200) $\mu$ m	2.1 $\mu$ m	By stage micrometer

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> Indentors are verified by performance only. The direct verifications provided are considered partial due to the exclusion of the verification of the indenter.

<sup>5</sup> For Rockwell scales using ball indenters, both steel (S) and tungsten carbide (W) are used. CMCs are reflected with the tungsten carbide balls only.



## Accredited Laboratory

A2LA has accredited

**NEWAGE TESTING INSTRUMENTS, INC.**

Horsham, PA

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 25<sup>th</sup> day of October 2021.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1734.01  
Valid to March 31, 2023

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*