

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & ANSI/NCSL Z540-1-1994

PLATCO/OMS 33501 Lake Road, Suite I Avon Lake, OH 44012 Richard Frye Phone: 440 835 1364

CALIBRATION

Valid To: December 31, 2022

Certificate Number: 1554.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 4}:

I. Dimensional

Parameter/Equipment	Range	$CMC^{2}(\pm)$	Comments
Optical Comparators ³ –			
Horizontal & Vertical Measurements	Up to 12 in	260 µin	Master cylindrical gages, glass scales
Magnification	5X to 100X	220 µin	
Edge Detection	5	210 µin	
Surface Plates ³ –			
Flatness	(10×8) in to (36×36) in	32 µin	Planekator
9	Up to <i>D</i> = 480	23√ <i>D</i> μin	Autocollimator D is the nominal length of the diagonal in inches
Repeat Reading	0.002 in	21 µin	Repeat-o-meter

¹ This laboratory offers commercial calibration service and field calibration service.

hu

Page 1 of 2

(A2LA Cert. No. 1554.01) 05/25/2021

- ² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- ³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ This scope meets A2LA's *P112 Flexible Scope Policy*.





Accredited Laboratory

A2LA has accredited

PLATCO/OMS Avon Lake, OH

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 25th day of May 2021.

Vice President, Accreditation Services For the Accreditation Council Certificate Number 1554.01 Valid to December 31, 2022

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.