



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
& ANSI/NCSL Z540-1-1994

ALDINGER COMPANY  
8023 Interstate 30  
Little Rock, AR 72209  
Dustin R. Detten Phone: 214 638 1808

CALIBRATION

Valid To: January 31, 2023

Certificate Number: 1509.02

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1,9</sup>:

I. Dimensional

| Parameter/Equipment                     | Range                                       | CMC <sup>2,4</sup> (±)     | Comments            |
|---|---|----------------------------|---------------------|
| Calipers <sup>3</sup>                   | Up to 24 in<br>(24 to 48) in                | 550 µin<br>1100 µin        | Gage blocks (field) |
|   | Up to 48 in<br>(48 to 96) in                | 99 µin<br>1500 µin         | Gage blocks (lab)   |
| Dial and Test Indicators <sup>3,5</sup> | Up to 3 in                                  | 99 µin                     | Gage blocks, ULM    |
| Height Gages <sup>3</sup>               | Up to 48 in                                 | 290 µin                    | Gage blocks (field) |
|   |   | 100 µin                    | Gage blocks (lab)   |
| Length Standards                        | Up to 22 in                                 | (43 + 1.5L) µin            | ULM, gage blocks    |
| Micrometers <sup>3</sup>                | Up to 4 in<br>(4 to 24) in                  | 93 µin<br>550 µin          | Gage blocks (field) |
|   | Up to 4 in<br>(4 to 24) in<br>(24 to 48) in | 19 µin<br>77 µin<br>99 µin | Gage blocks (lab)   |

| Parameter/Equipment  | Range  | CMC <sup>2, 4</sup> ( $\pm$ )  | Comments  |
|--|--|--|---|
| Plain Ring Gages   | (0.125 to 12) in   | $(22 + 1.5L) \mu\text{in}$   | ULM, gage blocks  |
| Thread Plugs –<br>Pitch Diameter<br>Major Diameter<br>Pitch Diameter (Taper) | (0.040 to 4) in<br>(4 to 12) in<br>(0.040 to 4) in<br>(4 to 12) in<br>(0.040 to 12) in | $(71 + 1.5L) \mu\text{in}$<br>$(95 + 1.5L) \mu\text{in}$<br>$(19 + 1.5L) \mu\text{in}$<br>$(35 + 1.5L) \mu\text{in}$<br>$(77 + 1.5L) \mu\text{in}$ | Three-wire method,<br>direct measure<br>Two-wire method,<br>direct measure      |
| Surface Plates <sup>3, 5</sup>   | Up to 107 in DL<br><br>(-0.002 to 0.002) in  | $(48 + 0.41DL) \mu\text{in}$<br><br>31 $\mu\text{in}$  | Leveling system ( <i>DL</i> is<br>diagonal of the plate)<br>Repeat reading gage |
| Optical Comparators <sup>3, 5</sup><br>X-Axis<br>Y-Axis<br>Angle             | Up to 12 in<br>Up to 12 in<br>(0 to 90)°   | $(250 + 91L) \mu\text{in}$<br>$(250 + 91L) \mu\text{in}$<br>0.029°   | Glass scale<br><br>Angle blocks   |
| Rulers <sup>3, 5</sup>   | Up to 72 in  | 0.013 in   | Standard ruler  |
| Tape Measures <sup>3, 5</sup>  | Up to 40 in<br>Up to 100 ft  | 0.013 in<br>0.062 in   | Standard ruler<br>Standard tape   |
| PI Tapes <sup>3, 5</sup>   | Up to 72 in  | 910 $\mu\text{in}$   | Gage blocks   |

## II. Electrical/DC-Low Frequency

| Parameter/Range   | Frequency  | CMC <sup>2, 7, 8</sup> ( $\pm$ )   | Comments  |
|---|--|--|---|
| AC Current – Measure <sup>3, 5</sup><br><br>Up to 1 A<br>(1 to 3) A   | 60 Hz to 1 kHz<br>60 Hz to 1 kHz   | 0.1 % + 0.4 mA<br>0.1 % + 1.8 mA   | HP 34401  |
| AC Current – Generate <sup>3, 5</sup><br><br>(29 to 330) $\mu$ A<br>330 $\mu$ A to 3.3 mA<br>(3.3 to 33) mA<br>(33 to 330) mA<br>(0.33 to 3) A<br>(3 to 10) A<br>(10 to 20) A<br><br>(16 to 150) A<br>(150 to 1000) A | 45 Hz to 1 kHz<br><br><br><br><br><br><br><br>(45 to 440) Hz   | 0.13 % + 0.1 $\mu$ A<br>0.10 % + 0.15 $\mu$ A<br>0.07 % + 2 $\mu$ A<br>0.083 % + 20 $\mu$ A<br>0.10 % + 100 $\mu$ A<br>0.21 % + 2 mA<br>0.21 % + 5 mA<br><br>1 % + 0.25 A<br>1 % + 0.9 A | Fluke 5502A<br><br><br><br><br><br><br><br>Using 50 turn coil |
| AC Voltage – Generate <sup>3, 5</sup><br><br>(32 to 320) mV<br>320 mV to 3.2 V<br>(3.2 to 32) V<br>(32 to 105) V<br><br>(105 to 320) V<br><br>(320 to 800) V  | 60 Hz to 3 kHz<br><br><br><br><br>60 Hz to 1 kHz<br>(1 to 3) kHz<br><br>(60 to 100) Hz<br>(1 to 3) kHz | 0.041 % + 20 $\mu$ V<br>0.041 % + 200 $\mu$ V<br>0.041 % + 2.0 mV<br>0.043 % + 6.3 mV<br><br>0.066 % + 20 mV<br>0.09 % + 20 mV<br><br>0.079 % + 63 mV<br>0.1 % + 63 mV                   | Fluke 5502A   |
| AC Voltage – Measure <sup>3, 5</sup><br><br>Up to 100 mV<br>(0.1 to 1) V<br>(1 to 10) V<br>(10 to 100) V<br>(100 to 750) V<br><br>Up to 10 kV   | 60 Hz to 3 kHz   | 0.069 % + 40 $\mu$ V<br>0.064 % + 300 $\mu$ V<br>0.064 % + 3 mV<br>0.064 % + 30 mV<br>0.064 % + 230 mV<br><br>0.12 % + 100 mV  | HP 34401<br><br><br><br><br><br><br>Vitretek 4700             |

| Parameter/Equipment                    | Range   | CMC <sup>2, 6, 8</sup> ( $\pm$ )  | Comments  |
|--|---|---|---|
| Capacitance – Generate <sup>3, 5</sup> |   |   |   |
| Low                                    | (0.5 to 4) nF<br>(4 to 40) nF<br>(40 to 400) nF<br>400 nF to 4 $\mu$ F  | 0.61 % + 15 pF<br>0.6 % + 60 pF<br>0.61 % + 320 pF<br>0.8 % + 3.2 nF  | Fluke 5502A   |
| High                                   | (4 to 40) $\mu$ F<br>(40 to 400) $\mu$ F<br>400 $\mu$ F to 4 mF<br>(4 to 40) mF   | 1 % + 32 nF<br>1 % + 320 nF<br>1 % + 3.2 $\mu$ F<br>2 % + 120 $\mu$ F   |   |
| DC Current – Generate <sup>3, 5</sup>  | (0 to 320) $\mu$ A<br>320 $\mu$ A to 3.2 mA<br>(3.2 to 32) mA<br>(32 to 320) mA<br>320 mA to 3.2 A<br>(3.2 to 11) A<br><br>(10 to 150) A<br>(150 to 1000) A | 0.014 % + 11 nA<br>0.014 % + 83 nA<br>0.014 % + 900 nA<br>0.016 % + 9.6 $\mu$ A<br>0.06 % + 120 $\mu$ A<br>0.057 % + 940 $\mu$ A<br><br>0.5 % + 0.14 A<br>0.5 % + 0.5 A | Fluke 5502A<br><br><br><br><br><br>Using 50 turn coil |
| DC Current – Measure <sup>3, 5</sup>   | Up to 10 mA<br>(10 to 100) mA<br>(0.1 to 1) A<br>(1 to 3) A   | 0.076 % + 6 $\mu$ A<br>0.05 % + 5 $\mu$ A<br>0.1 % + 100 $\mu$ A<br>0.14 % + 600 $\mu$ A  | HP 34401  |
| DC Voltage – Generate <sup>3, 5</sup>  | (0 to 320) mV<br>320 mV to 3.2 V<br>(3.2 to 32) V<br>(32 to 320) V<br>(320 to 1050) V   | 0.0063 % + 4.2 $\mu$ V<br>0.0062 % + 42 $\mu$ V<br>0.0066 % + 420 $\mu$ V<br>0.0069 % + 4.5 mV<br>0.0084 % + 20 mV  | Fluke 5502A   |
| DC Voltage – Measure <sup>3, 5</sup>   | Up to 100 mV<br>(0.1 to 1) V<br>(1 to 10) V<br>(10 to 100) V<br>(100 to 1000) V<br><br>Up to 10 kV  | 0.0054 % + 3.5 $\mu$ V<br>0.0043 % + 7.0 $\mu$ V<br>0.0036 % + 50 $\mu$ V<br>0.0048 % + 600 $\mu$ V<br>0.0047 % + 10 mV<br><br>0.041 % + 30 mV                          | HP 34401<br><br><br><br><br>Vitretek 4700             |

| Parameter/Equipment   | Range  | CMC <sup>2, 6, 7, 8</sup> ( $\pm$ )  | Comments               |
|---|--|--|------------------------|
| Electrical Calibration of Temperature Controllers <sup>3, 5</sup> | (-200 to 1371) °C  | 1.0 °C   | Fluke 724 – J, K, T, E |
| Resistance – Generate <sup>3, 5</sup>                             | (0 to 40) $\Omega$<br>(40 to 400) $\Omega$<br>(0.4 to 4) k $\Omega$<br>(4 to 40) k $\Omega$<br>(40 to 400) k $\Omega$<br>(0.4 to 4) M $\Omega$<br>(4 to 40) M $\Omega$<br>(40 to 400) M $\Omega$ | 0.11 % + 50 m $\Omega$<br>0.051 % + 100 m $\Omega$<br>0.037 % + 200 m $\Omega$<br>0.052 % + 2 $\Omega$<br>0.053 % + 20 $\Omega$<br>0.049 % + 200 $\Omega$<br>0.15 % + 2 k $\Omega$<br>0.14 % + 40 k $\Omega$ | Fluke 5502A            |
| Resistance – Measure <sup>3, 5</sup>                              | Up to 100 $\Omega$<br>(0.1 to 1) k $\Omega$<br>(1 to 10) k $\Omega$<br>(10 to 100) k $\Omega$<br>(0.1 to 1) M $\Omega$<br>(1 to 10) M $\Omega$<br>(10 to 100) M $\Omega$                         | 0.01 % + 4 m $\Omega$<br>0.01 % + 10 m $\Omega$<br>0.01 % + 100 m $\Omega$<br>0.01 % + 1.0 $\Omega$<br>0.01 % + 10 $\Omega$<br>0.04 % + 100 $\Omega$<br>0.8 % + 10 k $\Omega$                                | HP 34401               |
| RTD Simulation <sup>3, 5</sup><br><br>PT385, 100 $\Omega$         | (-200 to -100) °C<br>(-100 to 100) °C<br>(100 to 630) °C<br>(630 to 850) °C  | 0.26 °C<br>0.17 °C<br>0.35 °C<br>0.53 °C   | Fluke 5502A            |
| Thermocouple Simulation <sup>3, 5</sup><br><br>Type E             | (-250 to -200) °C<br>(-200 to -100) °C<br>(-100 to 100) °C<br>(100 to 1000) °C   | 0.45 °C<br>0.23 °C<br>0.18 °C<br>0.22 °C   | Fluke 5502A            |

| Parameter/Equipment                              | Range   | CMC <sup>2,6</sup> (±)                              | Comments                         |
|--|---|---|----------------------------------|
| Thermocouple Simulation <sup>3,5</sup><br>(cont) |   |   |                                  |
| Type J   | (-210 to -100) °C<br>(-100 to 800) °C<br>(800 to 1000) °C<br>(1000 to 1200) °C                    | 0.26 °C<br>0.20 °C<br>0.22 °C<br>0.24 °C            | Fluke 5502A                      |
| Type K   | (-250 to -200) °C<br>(-200 to -100) °C<br>(-100 to 100) °C<br>(100 to 600) °C<br>(600 to 1372) °C | 0.57 °C<br>0.28 °C<br>0.20 °C<br>0.24 °C<br>0.28 °C |                                  |
| Type T   | (-250 to -200) °C<br>(-200 to -100) °C<br>(-100 to 0) °C<br>(0 to 400) °C                         | 0.59 °C<br>0.28 °C<br>0.23 °C<br>0.18 °C            |                                  |
| Welding Devices <sup>3,5</sup>                   | (0 to 350) ADC<br>(0 to 100) VDC<br>(100 to 700) Feed Rate<br>IPM                                 | 0.87 ADC<br>0.012 VDC<br>3.6 IPM                    | Load bank, current<br>shunt, DMM |

### III. Mechanical

| Parameter/Equipment     | Range  | CMC <sup>2</sup> (±)      | Comments        |
|-------------------------|--|---------------------------|-----------------|
| Balances <sup>3,5</sup> |  |                           |                 |
|                         | (0 to 20) g<br>(0 to 200) g                    | 0.11 mg<br>0.69 mg        | Class 1 weights |
|                         | (0 to 200) g<br>(0 to 1000) g<br>(0 to 5000) g | 1.3 mg<br>6.6 mg<br>34 mg | Class 2 weights |

| Parameter/Equipment  | Range   | CMC <sup>2, 10</sup> (±)  | Comments                                 |
|--|---|---|--|
| Balances <sup>3, 5</sup> (cont)  | (0 to 20) g<br>(0 to 200) g<br>(0 to 1000) g<br>(0 to 5000) g<br>(0 to 10 000) g<br>(0 to 20 000) g   | 0.81 mg<br>4.8 mg<br>24 mg<br>130 mg<br>370 mg<br>460 mg  | Class 4 weights                          |
| Force – Measuring Equipment <sup>3, 5</sup>  |   |   |  |
| Tension / Compression  | (0 to 300) lbf  | 0.019 lbf   | Class F weights                          |
| Compression  | (100 to 1000) lbf<br>(1000 to 10 000) lbf<br>(10 000 to 50 000) lbf<br>(10 000 to 100 000) lbf<br>(50 000 to 500 000) lbf                           | 0.26 % Indication<br>0.24 % Indication<br>0.33 % Indication<br>1.6 % Indication<br>0.4 % Indication         | Load cells                               |
| Tension  | (100 to 1000) lbf<br>(1000 to 10 000) lbf<br>(10 000 to 50 000) lbf   | 0.23 % Indication<br>0.23 % Indication<br>0.29 % Indication   |  |
| Optical Tachometer <sup>3, 5</sup>   | 200 FPM<br>3000 FPM<br>29 999 FPM   | 1.3 FPM<br>1.7 FPM<br>3.9 FPM   | Calibrated strobe                        |
| Pressure Measuring Equipment – Gages, Transducers and Transmitters <sup>3, 5</sup> | (0 to 10 000) psig<br>(0 to 300) psig<br><br>(0 to 1000) psig<br>(0 to 300) psig<br>(0 to 50) psig<br>(0 to 5) psig<br>(0 to 10) inH <sub>2</sub> O | 6.6 psig<br>0.25 psig<br><br>2.5 psig<br>0.090 psig<br>0.019 psig<br>0.012 psig<br>0.034 inH <sub>2</sub> O | Druck DPI 104 gages<br><br>Druck DPI 620 |

| Parameter/Equipment  | Range   | CMC <sup>2, 10</sup> (±)  | Comments   |
|--|---|---|--|
| Vacuum – Measuring Equipment <sup>3, 5</sup>                       | (0 to 15) psig  | 0.19 psig   | Druck DPI 610  |
| Scales <sup>3, 5</sup>   | (0 to 10) lb<br>(0 to 20) lb<br><br>(0 to 50) lb<br>(0 to 100) lb<br>(0 to 200) lb<br>(0 to 500) lb<br>(0 to 1000) lb<br>(0 to 5000) lb<br>(0 to 20 000) lb | 0.000 25 lb<br>0.000 25 lb<br><br>0.0063 lb<br>0.0063 lb<br>0.0064 lb<br>0.0064 lb<br>0.18 lb<br>0.19 lb<br>0.21 lb | Handbook 44 w/<br>Class 4 weights<br><br>Class F weights |
| Torque Wrenches – Measure <sup>3, 5</sup>                          | (10 to 50) in·lbf<br>(30 to 400) in·lbf<br>(80 to 1000) in·lbf<br>(240 to 3000) in·lbf  | 0.41 in·lbf<br>2.4 in·lbf<br>6.0 in·lbf<br>21 in·lbf  | CDI  |
| Rotational Speed – Measure <sup>3, 5</sup>                         | Contact:<br>200 RPM<br>3000 RPM<br>20 000 RPM<br><br>Optical:<br>3000 RPM<br>30 000 RPM   | 1.5 RPM<br>2.4 RPM<br>11 RPM<br><br>1.3 RPM<br>3.3 RPM  | Shimpo DT-207LR  |
| Indirect Verification of Rockwell Hardness Testers <sup>3, 5</sup> | HRB:<br>Low<br>Medium<br>High<br><br>HRC:<br>Low<br>Medium<br>High  | 1.0 HRB<br>0.65 HRB<br>0.49 HRB<br><br>0.45 HRC<br>0.36 HRC<br>0.37 HRC   | Indirect verification                                    |
| Indirect Verification of Brinell Hardness Testers <sup>3, 5</sup>  | HBW 10/500:<br>Low  | 4.0 HBW   | Indirect verification                                    |



| Parameter/Equipment  | Range             | CMC <sup>2</sup> (±) | Comments              |
|--|-------------------|----------------------|-----------------------|
| Indirect Verification of Brinell Hardness Testers (cont) <sup>3, 5</sup> | HBW 10/3000: High | 4.8 HBW              | Indirect verification |

#### IV. Fluid Quantities

| Parameter/Equipment      | Range   | CMC <sup>2</sup> (±)                                    | Comments           |
|--------------------------|---|---|--------------------|
| Pipettes <sup>3, 5</sup> | (1 to 50) µL<br>(50 to 100) µL<br>(100 to 500) µL<br>(500 to 1000) µL<br>(1000 to 5000) µL<br>(5000 to 10 000) µL | 0.31 µL<br>0.6 µL<br>1.3 µL<br>2.2 µL<br>12 µL<br>18 µL | Gravimetric method |

#### V. Thermodynamics

| Parameter/Equipment  | Range                                  | CMC <sup>2, 10</sup> (±) | Comments   |
|--|--|--------------------------|--|
| Temperature – Measuring Equipment <sup>3, 5</sup>          | (50 to 400) °C                         | 0.35 °C                  | Fluke 9144 in dry block  |
| Temperature – Measuring Equipment                          | (-15 to 90) °C                         | 0.35 °C                  | Fluke 9144 probe in wet bath (lab only)  |
| Temperature – Measure <sup>3, 5</sup>                      | (-100 to 1370) °C<br>(-148 to 2498) °F | 1.2 °C<br>2.2 °F         | Fluke 724 w/ thermocouple  |
|  | (-25 to 400) °C                        | 0.098 °C                 | 5609 PRT w/Fluke 9144  |
| Infrared Measuring Equipment <sup>3, 5</sup>               | (Amb + 10 to 400) °C                   | 1.8 °C                   | Ametek ETC-400R<br>$\varepsilon = 0.95$ , $\lambda = (8 \text{ to } 14) \mu\text{m}$ |
| Humidity – Measure and Measuring Equipment <sup>3, 5</sup> | (10 to 90) % RH                        | 3.4 % RH                 | Sensor Scientific B13-200  |

## VI. Time & Frequency

| Parameter/Equipment                    | Range            | CMC <sup>2, 10</sup> ( $\pm$ ) | Comments        |
|--|------------------|--------------------------------|-----------------|
| Frequency – Measure                    | 40 Hz to 300 kHz | 0.014 % Indication             | HP 34401        |
| Timers and Stopwatches <sup>3, 5</sup> | Up to 24 hr      | 0.84 sec/24 hr                 | Reference timer |

<sup>1</sup> This laboratory offers commercial calibration service and field calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution, repeatability) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

<sup>4</sup> In the statement of CMC,  $L$  is the numerical value of the nominal length of the device measured in inches.

<sup>5</sup> The CMC stated for calibrations performed in the laboratory is applicable for calibrations performed in the field.

<sup>6</sup> CMC for the Fluke 5502A is based on 1-year specifications within a temperature range of  $23\text{ }^{\circ}\text{C} \pm 5\text{ }^{\circ}\text{C}$ . Field calibrations will be performed within  $23\text{ }^{\circ}\text{C} \pm 5\text{ }^{\circ}\text{C}$ .

<sup>7</sup> CMC for the HP 34401 is based on 1-year specifications within a temperature range of  $18\text{ }^{\circ}\text{C}$  to  $28\text{ }^{\circ}\text{C}$ . Field calibrations will be performed within  $18\text{ }^{\circ}\text{C}$  to  $28\text{ }^{\circ}\text{C}$ , 30 % to 55 % humidity.

<sup>8</sup> The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

<sup>9</sup> This scope meets A2LA's P112 Flexible Scope Policy.

<sup>10</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



# Accredited Laboratory

A2LA has accredited

**ALDINGER COMPANY**

*Little Rock, AR*

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 17<sup>th</sup> day of November 2020.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1509.02  
Valid to January 31, 2023  
Revised December 22, 2022

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*