



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

NEWTECH AYALON LTD
12 Hacharoshet St
Kfar Saba 4464010, Israel
David Leizer Phone: 972 72 3229888

CALIBRATION

Valid To: September 30, 2023

Certificate Number: 6314.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 4}:

I. Mechanical

Parameter/Equipment	Range	CMC ^{2, 6} (\pm)	Comments ⁵
Balances ³	Up to 3000 kg $R=0.000\ 001\ \text{g}$ $R=0.000\ 01\ \text{g}$ $R=0.0001\ \text{g}$ $R=0.001\ \text{g}$ Up to 2000 kg (2000 to 2500) kg (2500 to 3000) kg	 13R 1.6R 1.2R 1R 1R 1.3R 1.5R	Weights class E2, F1, (OIML R111) OIML/R 76 USP 41

¹ This laboratory offers commercial calibration and field calibration services.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC

⁴ This scope meets A2LA's P112 *Flexible Scope Policy*.

⁵ Calibration can be also performed to manufacturer or specific customer requirements.

⁶ In the statement of CMC, *R* represents the resolution of the unit under test

WITHDRAWN



Accredited Laboratory

A2LA has accredited

NEWTECH AYALON LTD.

KFAR SABA, ISRAEL

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 7th of September 2021.

A blue ink signature of the Vice President, Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 6314.01
Valid to September 30, 2023
Revised August 17, 2023

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.