

### SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

## BUREAU VERITAS CONSUMER PRODUCTS SERVICES (GUANGZHOU) CO., LTD SCIENCE CITY BRANCH

Rm. 101, G5 Building, South China Advanced Materials Innovation Park NO. 31 Kefeng Rd, Guangzhou Science City Guangdong, 510663, China

Ms. Yujin Tong Phone: 86-20-87148529 Email: yujin.tong@bureauveritas.com

#### **MECHANICAL**

Valid to: March 31, 2024 Certificate Number: 3971.02

In recognition of the successful completion of the A2LA evaluation, accreditation is granted to this laboratory to perform the following <u>Environmental Simulation</u> tests:

<b>Test Technology:</b>	<b>Test Method(s):</b>
Cold Test (Temperature range: ambient to -70°C)	IEC 60068-2-1
Dry Heat Test (Temperature range: ambient to +150°C)	IEC 60068-2-2
Damp Heat, Steady State (Temperature $\leq 100^{\circ}C$ ) (Humidity (25 to 98) %RH)	IEC 60068-2-78
Damp Heat, Cyclic (Temperature range $\leq 100^{\circ}$ C) (Humidity (25 to 98) %RH)	IEC 60068-2-30
Composite Temperature / Humidity Cyclic Test (Temperature range ≤ 100°C) (Humidity (25 to 98) %RH)	IEC 60068-2-38
Damp Heat, Steady State (Temperature range $\leq 100^{\circ}$ C) (Humidity (25 to 98) %RH)	IEC 60068-2-67

Page 1 c



# **Accredited Laboratory**

A2LA has accredited

# BUREAU VERITAS CONSUMER PRODUCTS SERVICES (GUANGZHOU) CO., LTD, SCIENCE CITY BRANCH

Guangdong, China

for technical competence in the field of

## Mechanical Testing

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SEAL 1978 VIOLED TO FOOD OF COLUMN AZLA

Presented this 13th day of May 2022.

Vice President, Accreditation Services For the Accreditation Council Certificate Number 3971.02 Valid to March 31, 2024

For the tests to which this accreditation applies, please refer to the laboratory's Mechanical Scope of Accreditation.