



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

VIATRAN
199 Fire Tower Dr.
Tonawanda, NY 14150
Joel Neri Phone: 716 629 3800

CALIBRATION

Valid To: February 29, 2024

Certificate Number: 5655.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 4}:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2, 5} (±)	Comments
DC Voltage – Generate	(0 to 90) mV	3.4 µV	EDC CR103

II. Mechanical

Parameter/Equipment	Range	CMC ^{2, 6} (±)	Comments
Pressure – Measuring Equipment	(0 to 1500) psi (0 to 5000) psi (0 to 15 000) psi (0 to 30 000) psi (0 to 100 000) psi	1.2 psi 1.8 psi 2.8 psi 12 psi 53 psi	Transducer with meter: TS-1500 TS-5000 TS-15 000 TS-30k TS-100k
	(-150 to 150) in H ₂ O (-15 to 50) psi (-15 to 500) psi	0.030 in H ₂ O 0.007 psi 0.052 psi	DPG

Parameter/Equipment	Range	CMC ^{2, 3, 6} (\pm)	Comments
Pressure – Measuring Equipment (cont)	(-15 to 7.5) psi	0.002 psi	Mensor sensor
	(-15 to 15) psi	0.003 psi	
	(-15 to 25) psi	0.004 psi	
	(-15 to 50) psi	0.007 psi	
	(-15 to 65) psi	0.008 psi	
	(-15 to 200) psi	0.022 psi	
	(-15 to 500) psi	0.052 psi	
	(-15 to 1000) psi	0.10 psi	
	(65 to 145) psi	0.01 %	
	(500 to 60 000) psi	0.009 %	Pressurements deadweight tester M3820
	(0 to 10/30) in H ₂ O	0.009 %	Fluke 7250LP

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ In the statement of CMC, percentages are percentage of reading, unless otherwise indicated.

⁴ This scope meets A2LA's *PI12 Flexible Scope Policy*.

⁵ The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

⁶ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



Accredited Laboratory

A2LA has accredited

VIATRAN

Tonawanda, NY

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 29th day of November 2021.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 5655.01
Valid to February 29, 2024

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.