

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

BURTON PRECISION INC. 5737 Vinton NW Comstock Park, MI 49321 James Krug Phone: 616 784 1756

CALIBRATION

Valid To: December 31, 2023

Certificate Number: 5497.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 5}:

I. Dimensional

Parameter/Equipment	Range	CMC ^{2, 4} (±)	Comments
Portable Arm Calibration ³ – All Makes	Up to 4500 mm	(23 + 0.013 <i>L</i>) μm	Ball bar, 35 mm master sphere

II. Dimensional Testing⁶

Parameter/Equipment	Range	CMC ^{2, 4, 7} (±)	Comments
Geometric & Dimensional Length – Measure	Up to (1500 x 2500 x 1000) mm	(11 + 0.013 <i>L</i>) μm	СММ

¹ This laboratory offers commercial calibration service and field calibration service.

(A2LA Cert. No. 5497.01) 01/07/2022

Page 1 of 2

5202 Presidents Court, Suite 220 | Frederick, MD 21703-8515 | Phone: 301 644 3248 | Fax: 240 454 9449 | www.A2LA.org

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMC's represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

- ³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- ⁴ In the statement of CMC, L is the numerical value of the nominal length of the device measured in mm.
- ⁵ This scope meets A2LA's *P112 Flexible Scope Policy*.
- ⁶ This test is not equivalent to that of a calibration.
- ⁷ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

Page 2 of 2





Accredited Laboratory

A2LA has accredited

BURTON PRECISION INC.

Comstock Park, MI

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 7th day of January 2022.

Vice President, Accreditation Services For the Accreditation Council Certificate Number 5497.01 Valid to December 31, 2023

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.