



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

MICRO PRECISION CALIBRATION DE MEXICO
Alvaro Obregon No. 252, int. 1
Zona Centro
Silao, Guanajuato, Mexico C.P. 36100
Victor Alberto Gonzalez Phone: +52 (664) 623-3383

CALIBRATION

Valid To: January 31, 2023

Certificate Number: 0935.20

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations and dimensional testing^{1, 10}:

I. Acoustics & Vibration

| Parameter/Equipment | Range | CMC ² (±) | Comments |
|--|--------------------------|----------------------|--------------------------------------|
| Sound Pressure Level/Sound Level Meter | (94 & 114) dB @ 1 kHz | 0.24 dB | Comparison to sound level calibrator |

II. Chemical Quantities

| Parameter/Equipment | Range | CMC ² (±) | Comments |
|-------------------------------------|---|---|----------------------------------|
| Conductivity Meters | 10 µS/cm 100 µS/cm 1000 µS/cm 1408 µS/cm | 0.55 µS/cm 2.2 µS/cm 5.9 µS/cm 7.0 µS/cm | Comparison to standard solutions |
| pH Meters | (4, 7, 10) pH unit | 0.02 pH unit | Comparison to standard solutions |
| Mass Concentration – Refractometers | (0 to 70) % Brix | 0.12 % Brix | Balance |

III. Dimensional

| Parameter/Equipment | Range | CMC ^{2, 5} (\pm) | Comments |
|--|-----------------------------|--|--|
| Calipers | Up to 40 in | $(310 + 4.1L) \mu\text{in}$ | Gage blocks and check master |
| Indicators | Up to 4.0 in | $(3.3 + 5.4L) \mu\text{in}$ | Indicator calibrator, master gage blocks |
| Micrometers (Internal, External, Depth, Bore) | Up to 8 in (8 to 24) in | $(33 + 1.4L) \mu\text{in}$ $(19 + 5.2L) \mu\text{in}$ | Gage blocks and gage rods |
| Height Gages | Up to 40 in | $(23 + 3.2L) \mu\text{in}$ | Gage blocks and check master |
| Rules and Tapes | Up to 40 in Up to 164 ft | $(75 + 15L) \mu\text{in}$ $(140 + 19L) \mu\text{in}$ | Comparison to glass scales |
| Gauge Blocks – Length | Up to 12 in | $(2.3 + 1.3L) \mu\text{in}$ | P&W Labmaster™, master gage blocks |
| Cylindrical Plain Plug Gages – Outside Diameter | Up to 12 in | $(2.5 + 1.3L) \mu\text{in}$ | P&W Labmaster™, master gage blocks |
| Thread Plug Gage – Major Diameter Pitch Diameter | Up to 12 in Up to 12 in | $(2.5 + 1.4L) \mu\text{in}$ $(47 + 0.29L) \mu\text{in}$ | P&W Labmaster™ & thread measuring wires |
| Thread Ring Gage – Minor Diameter Pitch Diameter | Up to 12 in Up to 12 in | $(10 + 0.87L) \mu\text{in}$ $(40 + 0.58L) \mu\text{in}$ | P&W Labmaster™, master ring gage, master gage blocks |

| Parameter/Equipment | Range | CMC ^{2, 5, 6} (\pm) | Comments |
|--|--|--|--|
| Ring Gage – Plain Cylindrical | Up to 12 in | $(10 + 0.87L) \mu\text{in}$ | P&W Labmaster™, master rings and master gage blocks |
| Micrometers/Length Standards | Up to 12 in | $(2.2 + 1.6L) \mu\text{in}$ | P&W Labmaster™, master gage blocks |
| Feeler/Taper Gage | Up to 0.5 in | $(18 + 0.30L) \mu\text{in}$ | Supermicrometer and master gage blocks |
| Angle Meter (Protractor) | (0 to 90)° | 0.0085° | Sine bar and gage blocks |
| Supermicrometer – Length | Up to 1 in | $(4.8 + 3.4L) \mu\text{in}$ | Gage blocks and optical parallel |
| Surface Plates – Local Area Flatness Overall Flatness | [(12 x 12) to (70 x 140)] in [(12 x 12) to (70 x 140)] in | 16 μin 0.23 % + 0.25 arcsec | Repeat-o-meter Wyler surface plate measuring system |
| Radius Gage | Up to 6 in | $(110 + 1.0L) \mu\text{in}$ | Optical comparator |
| CMM ³ – 3D Length – Error of Indication Probing Error | Up to 40 in 0.785 985 in | $(2.5 + 24L) \mu\text{in}$ 8.2 μin | ISO 10360-2: 2001: Checker master Master sphere |

| Parameter/Equipment | Range | CMC ^{2, 5} (\pm) | Comments |
|---|--|--|---|
| Optical Comparators and Visual Systems ³ – X & Y Axis Length Error of Indication Angle | Up to 0.5 in Up to 24 in 0, 30°, 45°, 60°, 90° | (50 + 0.19L) μ in (79 + 6.4L) μ in 0.0017° | Glass scale and gage blocks Angle blocks |
| Coating Thickness Gauge – Ferrous and Non-Ferrous | Up to 60 mils | 0.010 mils | Comparison to standard shims |

IV. Dimensional Testing⁷

| Parameter/Equipment | Range | CMC ² (\pm) | Comments |
|--|---------------|-----------------------------|---------------------|
| Length (3D) ³ – Fixtures and Workpieces | Up to 1500 mm | 110 μ m/m + 6.2 μ m | CMM |
| Length (2D) – Fixtures and Workpieces | Up to 200 mm | 250 μ m/m + 1.1 μ m | Optical comparators |

V. Electrical – DC/Low Frequency

| Parameter/Equipment | Range | CMC ^{2, 4} (\pm) | Comments |
|------------------------------------|---|--|---|
| DC Voltage – Generate ³ | (0 to 220) mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1100) V | 12 μ V/V + 1.0 μ V 12 μ V/V + 1.0 μ V 10 μ V/V + 2.2 μ V 15 μ V/V 14 μ V/V + 43 μ V 14 μ V/V + 43 μ V | Direct comparison to Fluke multi-calibrator |

| Parameter/Equipment | Range | CMC ^{2, 4, 6} (\pm) | Comments |
|------------------------------------|---|---|---|
| DC Voltage – Measure ³ | (0 to 100) mV 100 mV to 1V (1 to 10) V (10 to 100) V (100 to 1000) V | 1.8 μ V/V + 0.90 μ V 0.46 μ V/V + 1.0 μ V 1.1 μ V/V + 0.42 μ V 1.2 μ V/V + 0.48 μ V 6.1 μ V/V + 0.50 mV | Direct comparison to DMM |
| High Voltage | (100 to 1000) V (1000 to 20 000) V | 0.71 mV/V + 0.53 V 3.3 mV/V + 8.3 V | Comparison to high voltage meter |
| DC Current - Generate ³ | (0 to 22) mA (22 to 220) mA 220 mA to 2.2 A (2.2 to 11) A (11 to 20.5) A (0.1 to 1000) A | 66 μ A/A + 0.24 μ A 0.010 % 0.023 % 0.056 % 0.17 % 0.21 % | Direct comparison to Fluke multi-calibrator Comparison to Fluke multi-calibrator with current coil |
| DC Current – Measure ³ | (0 to 1) μ A (1 to 10) μ A (10 to 100) μ A (0.1 to 1) mA (1 to 10) mA (10 to 100) mA (0.1 to 1) A (0.1 to 15) A (15 to 100) A (100 to 300) A (300 to 1000) A (0.1 to 2000) A | 4.3 μ A/A + 0.059 nA 1.8 μ A/A + 0.061 nA 3.9 μ A/A + 0.041 nA 6.7 μ A/A + 0.24 nA 5.7 μ A/A + 0.73 nA 10 μ A/A + 0.040 μ A 0.14 mA/A + 0.013 mA 0.0042 % 0.0048 % 0.0058 % 0.038 % 0.84 % | Direct comparison to DMM Shunt monitored with DMM Comparison to digital clamp meter |

| Parameter/Equipment | Range | CMC ^{2, 4, 6} (\pm) | Comments |
|------------------------------------|--|--|---|
| Resistance – Generate ³ | (0 to 1.9) Ω (1.9 to 19) Ω (19 to 190) Ω (0.19 to 1.9) k Ω (1.9 to 19) k Ω (19 to 190) k Ω (0.19 to 1.9) M Ω (1.9 to 19) M Ω (19 to 100) M Ω | 96 $\mu\Omega/\Omega$ + 51 $\mu\Omega$ 38 $\mu\Omega/\Omega$ + 0.12 m Ω 21 $\mu\Omega/\Omega$ + 0.39 m Ω 18 $\mu\Omega/\Omega$ + 0.65 m Ω 17 $\mu\Omega/\Omega$ + 0.32 m Ω 22 $\mu\Omega/\Omega$ + 0.69 m Ω 38 $\mu\Omega/\Omega$ + 3.8 Ω 34 $\mu\Omega/\Omega$ + 72 Ω 11 $\mu\Omega/\Omega$ + 0.45 k Ω | Comparison to Fluke multi-calibrator |
| | (0.1 to 100) m Ω (0.1 to 1) Ω (1 to 10) Ω (10 to 100) Ω (0.1 to 1) k Ω (1 to 10) k Ω (10 to 100) k Ω (0.1 to 1) M Ω (1 to 10) M Ω (10 to 100) M Ω (0.1 to 1) G Ω (1 to 10) G Ω (10 to 100) G Ω 1 T Ω 10 T Ω | 0.60 % 0.064 % 0.0087 % 0.0035 % 0.0031 % 0.0028 % 0.0027 % 0.0060 % 0.0062 % 0.0092 % 0.016 % 0.016 % 0.025 % 0.35 % 1.5 % | Comparison to resistor standards |
| Resistance ³ – Measure | (0 to 1) Ω (1 to 10) Ω (10 to 100) Ω (0.1 to 1) k Ω (1 to 10) k Ω (10 to 100) k Ω (0.1 to 1) M Ω (1 to 10) M Ω (10 to 100) M Ω (0.1 to 1) G Ω (1 to 10) G Ω | 53 $\mu\Omega/\Omega$ + 26 $\mu\Omega$ 2.2 $\mu\Omega/\Omega$ + 76 $\mu\Omega$ 6.7 $\mu\Omega/\Omega$ + 32 $\mu\Omega$ 2.9 $\mu\Omega/\Omega$ + 0.41 m Ω 2.9 $\mu\Omega/\Omega$ + 0.42 m Ω 6.6 $\mu\Omega/\Omega$ + 37 m Ω 17 $\mu\Omega/\Omega$ + 1 Ω 78 $\mu\Omega/\Omega$ + 63 Ω 0.30 m Ω/Ω + 2.3 k Ω 2.3 m Ω/Ω + 0.20 M Ω 4.5 m Ω/Ω + 0.20 M Ω | Direct comparison to DMM LCR meter |

| Parameter/Range | Frequency | CMC ^{2, 4, 6} (\pm) | Comments |
|---|---|---|---|
| Capacitance ³ – Generate (220 to 399.9) pF (0.4 to 1.0999) nF (1.1 to 3.2999) nF (3.3 to 10.999) nF (11 to 32.999) nF (33 to 109.99) nF (110 to 329.99) nF (0.33 to 1.0999) μ F (1.1 to 3.2999) μ F (3.3 to 10.999) μ F (11 to 32.999) μ F (33 to 109.99) μ F (110 to 329.99) μ F (0.33 to 1.0999) mF (1.1 to 3.2999) mF (3.3 to 10.999) mF (11 to 32.999) mF (33 to 110) mF | 10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 3 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz (10 to 600) Hz (10 to 300) Hz (10 to 150) Hz (10 to 120) Hz (10 to 80) Hz (0.1 to 50) Hz (0.1 to 20) Hz (0.1 to 6) Hz (0.1 to 2) Hz (0.1 to 0.6) Hz (0.1 to 0.2) Hz | 4.4 mF/F + 12 pF 4.3 mF/F + 12 pF 10 mF/F 2.3 mF/F + 11 pF 8.0 mF/F 2.3 mF/F + 0.11 nF 2.3 mF/F + 0.34 nF 2.3 mF/F + 1.2 nF 2.2 mF/F + 3.6 nF 2.3 mF/F + 11 nF 3.6 mF/F + 3.4 nF 4.2 mF/F + 0.11 μ F 3.9 mF/F + 0.35 μ F 3.9 mF/F + 1.2 μ F 4.0 mF/F + 3.5 μ F 3.9 mF/F + 12 μ F 8.1 mF/F + 35 μ F 12 mF/F + 0.12 mF | Comparison to Fluke multi-calibrator |
| Capacitance ⁸ – Generate 1 pF to 10 μ F | (60 to 1000) Hz | 0.052 % | Direct comparison to capacitance decade |
| Inductance ⁸ – Generate 1.0 mH 10 mH 100 mH 1 H | Fixed Points (0.1 to 10) kHz | 0.029 % 0.029 % 0.032 % 0.028 % | Direct comparison to standard inductor |

| Parameter/Equipment | Range | CMC ² (±) | Comments |
|---|--|--|--------------------------------------|
| Electrical Simulation of Thermocouple – Generate / Measure | | | |
| Type B | (600 to 800) °C (800 to 1550) °C (1550 to 1820) °C | 0.57 °C 0.47 °C 0.46 °C | Comparison to Fluke multi-calibrator |
| Type C | (0 to 1000) °C (1000 to 1800) °C (1800 to 2316) °C | 0.37 °C 0.59 °C 0.98 °C | |
| Type E | (-250 to -100) °C (-100 to 650) °C (650 to 1000) °C | 0.58 °C 0.19 °C 0.25 °C | |
| Type J | (-210 to -100) °C (-100 to 760) °C (760 to 1200) °C | 0.32 °C 0.20 °C 0.27 °C | |
| Type K | (-200 to -100) °C (-100 to 120) °C (120 to 1000) °C (1000 to 1372) °C | 0.39 °C 0.22 °C 0.31 °C 0.47 °C | |
| Type L | (-200 to -100) °C (-100 to 800) °C (800 to 900) °C | 0.43 °C 0.31 °C 0.21 °C | |
| Type N | (-200 to -100) °C (-100 to 410) °C (410 to 1300) °C | 0.47 °C 0.26 °C 0.32 °C | |
| Type R | (0 to 250) °C (250 to 1000) °C (1000 to 1767) °C | 0.68 °C 0.44 °C 0.49 °C | |
| Type S | (0 to 250) °C (250 to 1400) °C (1400 to 1767) °C | 0.58 °C 0.47 °C 0.57 °C | |
| Type T | (-250 to -150) °C (-150 to 0) °C (0 to 400) °C | 0.73 °C 0.28 °C 0.20 °C | |
| Type U | (-200 to 0) °C (0 to 600) °C | 0.65 °C 0.32 °C | |

| Parameter/Equipment | Range | CMC ^{2, 4} (\pm) | Comments |
|---|------------------|-------------------------------|--------------------------------------|
| Electrical Simulation of RTD Indicating Systems ³ – | | | |
| Pt 385, 100 Ω | (-200 to 800) °C | 0.0064 % + 0.053 °C | Comparison to Fluke multi-calibrator |
| Pt 3926, 100 Ω | (-200 to 630) °C | 0.0050 % + 0.055 °C | |
| Pt 3916, 100 Ω | (-200 to 630) °C | 0.0059 % + 0.054 °C | |
| Pt 385, 200 Ω | (-200 to 630) °C | 0.0060 % + 0.052 °C | |
| Pt 385, 500 Ω | (-200 to 630) °C | 0.0060 % + 0.058 °C | |
| Pt 385, 1000 Ω | (-200 to 630) °C | 0.0051 % + 0.054 °C | |
| PtNi 385, 120 Ω | (-80 to 260) °C | 0.0023 % + 0.044 °C | |
| Cu 427, 10 Ω | (-100 to 260) °C | 0.0065 % + 0.071 °C | |
| Electrical Simulation of RTD Calibrators Systems ³ – | | | |
| Pt 385, 100 Ω | (-200 to 800) °C | 0.039 °C | Comparison to DMM |
| Pt 3926, 100 Ω | (-200 to 630) °C | 0.044 °C | |
| Pt 3916, 100 Ω | (-200 to 630) °C | 0.041 °C | |
| Pt 385, 200 Ω | (-200 to 630) °C | 0.039 °C | |
| Pt 385, 500 Ω | (-200 to 630) °C | 0.053 °C | |
| Pt 385, 1000 Ω | (-200 to 630) °C | 0.043 °C | |
| PtNi 385, 120 Ω | (-80 to 260) °C | 0.042 °C | |
| Cu 427, 10 Ω | (-100 to 260) °C | 0.04 °C | |

| Parameter/Range | Frequency | CMC ^{2, 4} (\pm) | Comments |
|------------------------------------|---|---|--|
| AC Voltage ³ – Generate | | | |
| (0 to 220) mV | (10 to 40) Hz (40 to 20) kHz (20 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz | 0.33 mV/V + 6.3 μ V 0.17 mV/V + 6.8 μ V 1.3 mV/V + 6.1 μ V 1.5 mV/V + 12 μ V 2.4 mV/V + 22 μ V 4.9 mV/V + 35 μ V | Direct comparison to Fluke multi- calibrator |
| 220 mV to 2.2 V | (10 to 40) Hz 40 Hz to 20 kHz (20 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz | 0.44 mV/V + 15 μ V 0.10 mV/V + 20 μ V 0.31 mV/V + 0.20 mV 0.63 mV/V + 0.19 mV 1.6 mV/V + 0.17 mV 3.7 mV/V + 0.29 mV | |
| (2.2 to 22) V | (10 to 40) Hz 40 Hz to 20 kHz (20 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz | 0.25 mV/V + 0.36 mV 0.19 mV/V 0.87 mV/V 2.1 mV/V 2.1 mV/V 5.2 mV/V | |
| (22 to 220) V | (10 to 40) Hz 40 Hz to 20 kHz (20 to 100) kHz | 0.26 mV/V + 0.19 mV 0.12 mV/V + 1.2 mV 0.80 mV/V + 0.32 mV | |
| (220 to 1000) V | 50 Hz to 1 kHz | 0.14 mV/V + 2.1 mV | |
| AC Voltage ³ – Measure | | | |
| (0 to 100) mV | (0.01 to 1) kHz (1 to 20) kHz (20 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz | 56 μ V + 6.1 μ V 0.16 mV/V + 4.3 μ V 0.40 mV/V + 55 μ V 1.2 mV/V + 0.48 mV 11 mV/V + 32 μ V | Direct comparison to DMM |
| 100 mV to 1 V | (0.01 to 1) kHz (1 to 20) kHz (20 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz | 0.11 mV/V + 0.14 μ V 0.19 mV/V + 1.2 μ V 0.95 mV/V + 0.16 μ V 3.6 mV/V + 0.20 μ V 12 mV/V + 0.12 μ V | |

| Parameter/Range | Frequency | CMC ^{2, 4, 6} (\pm) | Comments |
|---|---|---|--|
| AC Voltage ³ – Measure (cont) | | | |
| (1 to 10) V | (0.01 to 1) kHz (1 to 20) kHz (20 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 4) MHz | 0.11 mV/V + 4.3 μ V 0.19 mV/V + 0.66 μ V 0.95 mV/V + 0.61 μ V 3.6 mV/V + 0.17 μ V 12 mV/V + 3.5 μ V 6.0 mV/V + 57 mV | Direct comparison to DMM |
| (10 to 100) V | (0.01 to 1) kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz | 0.27 mV/V + 1.6 mV 0.26 mV/V + 0.75 mV 0.44 mV/V + 0.73 mV 1.5 mV/V + 1.3 mV | |
| (100 to 1000) V | (0.01 to 1) kHz (1 to 20) kHz | 0.52 mV/V + 27 mV 0.79 mV/V + 27 mV | |
| High Voltage (0.1 to 1) kV (1 to 20) kV | (50 to 100) Hz | 0.89 mV/V + 1.4 V 3.3 mV/V + 6.7 V | Comparison to high voltage meter |
| AC Current ³ – Generate | | | |
| (0 to 22) mA | 10 Hz to 1 kHz (1 to 10) kHz | 0.23 mA/A + 15 nA 0.28 % | Direct comparison to Fluke multi-calibrator |
| (22 to 220) mA | 10 Hz to 1 kHz (1 to 10) kHz | 1.4 mA/A 0.17 % + 21 μ A | |
| 220 mA to 2.2 A | 40 Hz to 1 kHz (1 to 10) kHz | 0.38 mA/A + 0.19 mA 0.65 % | |
| (2.2 to 11) A | 45 Hz to 1 kHz (1 to 5) kHz | 2.6 mA/A 3.9 % | |
| (11 to 20.5) A | 45 Hz to 1 kHz (1 to 5) kHz | 1.6 mA/A + 4.8 mA 3.1 % | |
| (20 to 1000) A | (50 to 400) Hz | 0.29 % | Comparison to Fluke multi-calibrator with current coil |

| Parameter/Range | Frequency | CMC ^{2, 4, 6} (\pm) | Comments |
|-----------------------------------|-----------------|----------------------------------|-----------------------------------|
| AC Current ³ – Measure | | | |
| (0 to 100) μ A | (0.01 to 1) kHz | 16 nA | Direct comparison to DMM |
| (0.1 to 1) mA | (0.01 to 1) kHz | 74 μ A/A + 3.9 nA | |
| (1 to 10) mA | (0.01 to 1) kHz | 73 μ A/A + 5.5 nA | |
| (10 to 100) mA | (0.01 to 1) kHz | 73 μ A/A + 0.013 nA | |
| (0.1 to 1) A | (0.01 to 1) kHz | 0.13 mA/A | |
| (0.1 to 15) A | (0.01 to 1) kHz | 0.027 % | Shunt monitored with DMM |
| (15 to 100) A | (0.01 to 1) kHz | 0.11 % | |
| (100 to 300) A | (0.01 to 1) kHz | 0.10 % | |
| (300 to 1000) A | (0.01 to 1) kHz | 0.081 % | Comparison to digital clamp meter |
| (0 to 2000) A | (0.01 to 1) kHz | 1.6 % | |

VI. Fluid Quantities

| Parameter/Equipment | Range | CMC ^{2, 6, 9} (\pm) | Comments |
|--|---|--|--|
| Dynamic Viscosity ³ – Measurement Equipment | (0.1 to 140) mPa·s (0.1 to 3000) mPa·s (0.1 to 25 000) mPa·s (0.1 to 70 000) mPa·s (0.1 to 210 000) mPa·s | 0.54 % 0.65 % 0.48 % 0.65 % 0.48 % | Comparison to Cannon viscosity standards |
| Flow ³ (Air) – Measurement Equipment | (0.1 to 500) ml/min (0.1 to 10) l/min (0.1 to 50) l/min | 0.42 % 0.41 % 0.41 % | Comparison to flow calibrator system |
| Velocity (Air) – Measurement Equipment | (0 to 20) m/s | 4.2 % + 0.098 m/s | Comparison to standard anemometer with wind tunnel |

VII. Mechanical

| Parameter/Equipment | Range | CMC ^{2, 6, 9} (\pm) | Comments |
|---|---|--|---|
| Indirect Verification of Rockwell Hardness Testers ³ | <p>HRC: < 35 HRC $(\geq 35 \text{ and } < 60)$ HRC ≥ 60 HRC</p> <p>HRBW: < 60 HRBW $(\geq 60 \text{ and } < 80)$ HRBW ≥ 80 HRBW</p> <p>HR15N < 78 HR15N $(\geq 78 \text{ and } < 90)$ HR15N ≥ 90 HR15N</p> <p>HR45N < 37 HR45N $(\geq 37 \text{ and } < 66)$ HR45N ≥ 66 HR45N</p> <p>HR30TW < 57 HR30TW $(\geq 57 \text{ and } < 70)$ HR30TW ≥ 70 HR30TW</p> <p>HR15TW < 81 HR15TW $(\geq 81 \text{ and } < 87)$ HR15TW ≥ 87 HR15TW</p> | <p>0.40 HRC 0.36 HRC 0.33 HRC</p> <p>0.65 HRBW 0.65 HRBW 0.48 HRBW</p> <p>0.43 HR15N 0.33 HR15N 0.25 HR15N</p> <p>0.47 HR45N 0.48 HR45N 0.21 HR45N</p> <p>0.56 HR30TW 0.36 HR30TW 0.34 HR30TW</p> <p>0.44 HR15TW 0.37 HR15TW 0.33 HR15TW</p> | Indirect method - ASTM E18 |
| Shore Hardness | (20 to 90) points | 0.35 points | Direct comparison to gage blocks and balance |
| Gauge Pressure ³ – Measurement Equipment | (-15 to +30) psi (0 to 1000) psi (0 to 10 000) psi | 0.030 % + 0.000 64 psi 0.016 % + 0.080 psi 0.015 % + 0.71 psi | Comparison to Fluke pressure calibrator systems |

| Parameter/Equipment | Range | CMC ^{2, 6, 9} (±) | Comments |
|---|--|--|---|
| Torque – Measurement Equipment | (0.1 to 400) in·ozf (0.1 to 50) in·lbf (0.1 to 150) in·lbf (0.1 to 400) in·lbf (0.1 to 600) ft·lbf | 0.56 % 0.55 % 0.55 % 0.58 % 0.56 % | CDI torque system |
| Torque Analyzers, Transducers | (0.1 to 200) in·lbf (0.1 to 370) ft·lbf | 0.080 % 0.14 % | Direct comparison to torque wheel & dead weights |
| Force – Tension and Compression | (0.1 to 1000) kgf (0.1 to 200) kgf (0.1 to 4500) kgf (0.1 to 45 000) kgf | 0.018 % 0.071 % 0.088 % 0.33 % | Direct comparison to dead weights Comparison to force transducer |
| Scales and Balances ³ | (0 to 300) g (0 to 100) kg (0 to 1000) kg | 1.5 µg/g+.015 mg 3.4 µg/g 0.20 mg/g | Comparison to weights ASTM Class 0 ASTM Class 1 OIML Class M2 |
| Mass | (1 to 100) g (1 to 34) kg | 0.16 mg 54 µg/g + 69 mg | Substitution method using reference weight & balance |
| Piston/Plunger Operated Volumetric Apparatus (POVA) | (10 to 20) µl (20 to 50) µl (50 to 100) µl 100 µl to 20 ml | 0.77 % of reading 0.40 % of reading 0.21 % of reading 0.14 % of reading | Scale, Weight Set Class 0, Barometer DP97 |
| Other Volumetric Devices | (0.1 to 100) ml (0.1 to 34) l | 0.13 ml/l 0.20 ml/l + 0.25 ml | Gravimetric method using reference weight & balance |

VIII. Thermodynamics

| Parameter/Equipment | Range | CMC ^{2, 6, 9} (\pm) | Comments |
|---|--|---|--|
| Relative Humidity – Measuring Equipment | 11 % RH 33 % RH 75 % RH 97 % RH | 0.70 % RH 0.61 % RH 0.90 % RH 1.5 % RH | Saturated salt solutions |
| Relative Humidity – Measure ³ | (0 to 100) % RH | 1.6 % | Comparison to humidity meter |
| Temperature – Infrared Thermometers | (35 to 500) °C (500 to 1000) °C | 0.73 % 0.77 % | Comparison to infrared calibrator |
| Temperature – Measuring Equipment | (-40 to 300) °C (300 to 1100) °C | 0.21 °C 0.46 % | Comparison to temperature blocks calibrators |
| Temperature ³ – Measure | (-80 to 300) °C | 0.084 °C | Comparison to RTD thermometer |

IX. Time & Frequency

| Parameter/Equipment | Range | CMC ^{2, 9} (\pm) | Comments |
|----------------------|------------------|-------------------------------|--|
| Frequency – Measure | 1 Hz to 10 MHz | 0.12 nHz/Hz + 0.39 nHz | Direct comparison using DMM |
| | 1 MHz to 3 GHz | 4.6 nHz/Hz + 0.41 mHz | Using universal counter |
| Frequency – Generate | 0.01 Hz to 2 MHz | 30 μ Hz/Hz + 0.88 mHz | Direct comparison to Fluke multi-calibrator |

| Parameter/Equipment | Range | CMC ^{2, 6, 9} (\pm) | Comments |
|---------------------------------------|----------------------|----------------------------------|---|
| Timers/Stop Watches | (0 to 24) hr | 55 ms | Method totalize, using universal counter |
| Angular Frequency – Photo Tachometers | (1 to 1 000 000) rpm | 0.000 43 % | Comparison to Fluke multi-calibrator with LED |

¹ This laboratory offers commercial dimensional testing/calibration service and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration and this laboratory meets A2LA R104 – *General Requirements: Accreditation of Field Testing and Field Calibration Laboratories* for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.

⁵ In the statement of CMC, L is the numerical value of the nominal length of the device measured in inches. Pitch diameter is measured by the three-wire method.

⁶ In the statement of CMC, the value is defined as the percentage of reading.

⁷ This laboratory meets R205 – *Specific Requirements: Calibration Laboratory Accreditation Program* for the types of dimensional tests listed above and is considered equivalent to that of a calibration.

⁸ The CMC stated for calibrations performed in the laboratory is not applicable for calibrations performed in the field.

⁹ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

¹⁰ This scope meets A2LA's P112 Flexible Scope Policy.



Accredited Laboratory

A2LA has accredited

MICRO PRECISION CALIBRATION DE MEXICO

Silao, Guanajuato, MEXICO

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 5th day of January 2021.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 935.20
Valid to January 31, 2023

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.