



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

PIPETTE MASTER LLC
18855 Lansdowne Stream Path
Katy, TX 77449
Aziz Rahim Phone: 214 642 6524

CALIBRATION

Valid To: June 30, 2023

Certificate Number: 5387.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter/Equipment	Range	CMC ² (±)	Comments
Piston Operated Volumetric Apparatus (POVA) ^{3,4} – Pipettes, Syringes, & Bottle-Top Dispensers	(1 to 10) µL (10 to 100) µL (100 to 1000) µL (1 to 10) mL (10 to 50) mL	0.073 µL 0.21 µL 1.9 µL 0.024 mL 0.12 mL	Gravimetric using A&D 4212B-PT analytical balance; method & tolerances to ISO 8655 (2002)

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the Calibration and Measurement Capability Uncertainty (CMC) found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ For work conducted in the field, due to the environmental conditions at the customer's location, the environmental specifications outlined in ISO 8655 may not be able to be achieved.

SUSPENDED



Accredited Laboratory

A2LA has accredited

PIPETTE MASTER LLC

Katy, TX

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 26th day of August 2021.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 5387.01
Valid to June 30, 2023

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.